



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB127948

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.4	10/24/2023	12:35
2	O4991-01		1	12	100.00	>212.0	>212.0	10/24/2023	13:05
3	O4991-01DUP		1	12	100.00	>212.0	>212.0	10/24/2023	13:35
4	O4991-03		1	14	100.00	>212.0	>212.0	10/24/2023	14:05
5	O4991-04		1	12	100.00	>212.0	>212.0	10/24/2023	14:35
6	O4991-05		1	13	100.00	>212.0	>212.0	10/24/2023	15:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

16127948

WorkList Name : fp-10-23

WorkList ID : 174825

Department : Wet-Chemistry

Date : 10-23-2023 08:10:30

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4991-01 F	1021-1024	Water	Flash Point	Cool 4 deg C	PSEG03	E31	10/20/2023	1010B
O4991-03 C	1025	Water	Flash Point	Cool 4 deg C	PSEG03	I31	10/20/2023	1010B
O4991-04 C	1026	Water	Flash Point	Cool 4 deg C	PSEG03	I31	10/20/2023	1010B
O4991-05 C	1027	Water	Flash Point	Cool 4 deg C	PSEG03	I31	10/20/2023	1010B

Date/Time 10/24/2023 12:30
 Raw Sample Received by: RMA
 Raw Sample Relinquished by: SGA

Date/Time 10/24/2023 15:20
 Raw Sample Received by: 7d (acc)
 Raw Sample Relinquished by: RMA