

LB127951


SFAM01.1

INSTRUMENT ID: CVJ

Linear

 μ Abs.:

54719



A= 0.0000e+000

B= 1.8376e-004 slope

C= -1.3817e-002 $\rightarrow$ intercept

Rho= 0.9995266

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.007	-0.007	35	0.000	35					0.00
0.05	0.050					419	0				1
0.20	0.200	0.207	0.007	1200	0.0 %	1200					4
2.50	2.500	2.437	-0.063	13339	0.0 %	13339					-3
5.00	5.000	5.200	0.200	28374	0.0 %	28374					4
7.50	7.500	7.322	-0.178	39923	0.0 %	39923					-2
10.0	10.000	10.041	0.041	54719	0.0 %	54719					0

LB127951

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	35	-		0 SFAM01.1	PPB	10/24/2023 14:30	Std
	0.2 S01	1200	-		0.2 SFAM01.1	PPB	10/24/2023 14:32	Std
	2.5 S02	13339	-		2.5 SFAM01.1	PPB	10/24/2023 14:34	Std
	5 S03	28374	-		5 SFAM01.1	PPB	10/24/2023 14:38	Std
	7.5 S04	39923	-		7.5 SFAM01.1	PPB	10/24/2023 14:43	Std
	10 S05	54719	-		10 SFAM01.1	PPB	10/24/2023 14:45	Std
ICV034	ICV034	20440	3.7421 -		SFAM01.1	PPB	10/24/2023 14:48	SMPL
ICB034	ICB034	-247	-0.0592 -		SFAM01.1	PPB	10/24/2023 14:50	SMPL
CCV073	CCV073	29249	5.3608 -		SFAM01.1	PPB	10/24/2023 14:52	SMPL
CCB073	CCB073	-265	-0.0625 -		SFAM01.1	PPB	10/24/2023 14:55	SMPL
PB156617BL	PBW617	-31	-0.0195 -		SFAM01.1	PPB	10/24/2023 14:57	SMPL
O4847-17	YC652	99	0.0044 -		SFAM01.1	PPB	10/24/2023 14:59	SMPL
O4848-01	YC657	171	0.0176 -		SFAM01.1	PPB	10/24/2023 15:01	SMPL
O4848-02	YC657D	200	0.0229 -		SFAM01.1	PPB	10/24/2023 15:04	SMPL
O4848-03	YC657S	5723	1.0378 -		SFAM01.1	PPB	10/24/2023 15:09	SMPL
O4848-04	YC658	-130	-0.0377 -		SFAM01.1	PPB	10/24/2023 15:11	SMPL
O4857-01	YBRR9	28414	5.2074 -		SFAM01.1	PPB	10/24/2023 15:13	SMPL
O4857-02	YBRR9D	25887	4.7431 -		SFAM01.1	PPB	10/24/2023 15:16	SMPL
O4857-03	YBRR9S	29551	5.4163 -		SFAM01.1	PPB	10/24/2023 15:18	SMPL
O4857-04	YBR50	7	-0.0125 -		SFAM01.1	PPB	10/24/2023 15:20	SMPL
PB156618BL	PBW617	437	0.0665 -		SFAM01.1	PPB	10/24/2023 15:22	SMPL
O4866-01	YCCZ1	330	0.0468 -		SFAM01.1	PPB	10/24/2023 15:25	SMPL
O4866-02	YCD28	-95	-0.0313 -		SFAM01.1	PPB	10/24/2023 15:27	SMPL
O4866-03	YCD30	8	-0.0123 -		SFAM01.1	PPB	10/24/2023 15:29	SMPL
O4866-04	YCD05	294	0.0402 -		SFAM01.1	PPB	10/24/2023 15:31	SMPL
O4866-05	YCD16	-23	-0.018 -		SFAM01.1	PPB	10/24/2023 15:34	SMPL
O4866-06	YCD17	4	-0.0131 -		SFAM01.1	PPB	10/24/2023 15:36	SMPL
O4866-07	YCD20	20	-0.0101 -		SFAM01.1	PPB	10/24/2023 15:38	SMPL
O4866-08	YCD20D	18	-0.0105 -		SFAM01.1	PPB	10/24/2023 15:41	SMPL
O4866-09	YCD20S	5986	1.0861 -		SFAM01.1	PPB	10/24/2023 15:43	SMPL
CCV074	CCV074	28110	5.1515 -		SFAM01.1	PPB	10/24/2023 15:45	SMPL
CCB074	CCB074	-41	-0.0214 -		SFAM01.1	PPB	10/24/2023 15:47	SMPL
O4866-10	YCD09	-29	-0.0191 -		SFAM01.1	PPB	10/24/2023 15:50	SMPL
O4866-11	YCD10	4	-0.0131 -		SFAM01.1	PPB	10/24/2023 15:52	SMPL
O4866-12	YCD26	30	-0.0083 -		SFAM01.1	PPB	10/24/2023 15:54	SMPL
O4866-13	YCD38	472	0.0729 -		SFAM01.1	PPB	10/24/2023 15:56	SMPL
O4866-14	YCD45	-38	-0.0208 -		SFAM01.1	PPB	10/24/2023 15:59	SMPL
O4866-15	YCD22	6	-0.0127 -		SFAM01.1	PPB	10/24/2023 16:01	SMPL
O4866-16	YCD23	12	-0.0116 -		SFAM01.1	PPB	10/24/2023 16:03	SMPL
O4866-17	YCD34	23	-0.0096 -		SFAM01.1	PPB	10/24/2023 16:05	SMPL
O4866-18	YCD36	744	0.1229 -		SFAM01.1	PPB	10/24/2023 16:08	SMPL
O4866-19	YCD13	-41	-0.0214 -		SFAM01.1	PPB	10/24/2023 16:10	SMPL
O4866-20	YCD32	-1	-0.014 -		SFAM01.1	PPB	10/24/2023 16:12	SMPL
O4866-21	YCD42	2659	0.4748 -		SFAM01.1	PPB	10/24/2023 16:15	SMPL
CCV075	CCV075	28827	5.2833 -		SFAM01.1	PPB	10/24/2023 16:17	SMPL
CCB075	CCB075	-31	-0.0195 -		SFAM01.1	PPB	10/24/2023 16:19	SMPL