

Clear table

Instrument ID: IC-2											
Analyst : IZ											
Method: 300.0 / 9056A											
ident	concentra tion F-	concentratio n CL-	concentratio on NO2	concentratio on BR-	concentratio on NO3	concentratio on HPO4	concentratio on SO4	file name	date time	Initial wt/ Final Vol	Analyst
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2023-10-13_	10/13/23 11:55		IZ/MA
STD2	0.4760	0.6790	0.6900	2.2840	0.5760	1.1830	3.5590	2023-10-13_	10/13/23 12:26		IZ/MA
STD3	0.8320	1.1990	1.1960	3.9990	1.0010	1.9570	6.0170	2023-10-13_	10/13/23 12:58		IZ/MA
STD4	1.0190	1.5130	1.5130	5.0250	1.2570	2.5160	7.5280	2023-10-13_	10/13/23 13:29		IZ/MA
STD5	1.8540	2.9090	2.9050	9.6870	2.4170	4.9030	14.3650	2023-10-13_	10/13/23 14:01		IZ/MA
STD6	3.9210	5.8890	5.8710	19.6300	4.8970	9.6600	29.3630	2023-10-13_	10/13/23 14:33		IZ/MA
STD7	5.0970	7.6110	7.6260	25.3750	6.3520	12.7810	38.1690	2023-10-13_	10/13/23 15:04		IZ/MA
ICV	2.0270	2.9970	3.0100	9.9240	2.5070	5.2390	14.8000	2023-10-13_	10/13/23 16:39		IZ/MA
ICB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2023-10-13_	10/13/23 17:11		IZ/MA
CCV	1.9400	3.1170	2.8310	10.3130	2.5080	4.8420	13.6300	2023-10-23_	10/23/23 13:08		IZ/MA
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2023-10-23_	10/23/23 13:44		IZ/MA
LB127952BLW	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2023-10-23_	10/23/23 14:18		IZ/MA
LB127952BSW	1.9970	3.1410	2.8880	10.4360	2.5640	5.0380	13.9070	2023-10-23_	10/23/23 14:53		IZ/MA
O5020-01	0.0000	0.0000	0.2560	0.0000	0.5590	0.0000	0.0000	2023-10-23_	10/23/23 15:59		IZ/MA
O5020-02	0.0000	0.0000	0.0000	0.0000	0.4300	0.0000	0.0000	2023-10-23_	10/23/23 16:35		IZ/MA
O5020-03	0.0000	0.3430	0.0000	0.0000	0.4230	0.0000	0.0000	2023-10-23_	10/23/23 17:10		IZ/MA
O5020-03MS	1.9370	3.1060	2.6650	9.6120	2.7000	4.7450	12.8760	2023-10-23_	10/23/23 17:46		IZ/MA
O5020-03MSD	1.9920	3.0890	2.6390	9.5210	2.6880	4.9990	12.6710	2023-10-23_	10/23/23 18:22		IZ/MA
CCV	2.0410	3.1710	2.9260	10.5480	2.6110	5.4150	13.9750	2023-10-23_	10/23/23 18:57		IZ/MA
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2023-10-23_	10/23/23 19:33		IZ/MA

Report date: 10/24/2023 2:32:54 PM
Printed by: wet


Ident: CCV
Analysis from: 10/23/2023 1:08:50 PM
File: _2023-10-23_

Last save: 10/23/2023 1:42:21 PM

Method: AnionsIC2-101323.mtw
Run operator: wet
Analysis number: 36272

Last save: 10/19/2023 12:0

SAMPLE:
: IZ/MA
Vial number: 11
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	7.34	1.585	4.13	2.81	419.267	6.20
2	10.53	0.325	18.56	12.66	466.615	6.90
3	12.83	0.415	31.63	21.56	985.077	14.57
4	16.87	0.487	18.08	12.32	690.008	10.20
5	19.50	0.604	20.49	13.97	964.084	14.26
6	23.85	0.900	22.72	15.49	1653.304	24.45
7	28.11	0.679	31.08	21.18	1584.736	23.43
7	33.00	0.714	146.68	100.00	6763.092	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/24/2023 2:33:01 PM
Printed by: wet


Ident: CCB
Analysis from: 10/23/2023 1:44:36 PM
File: _2023-10-23_

Last save: 10/23/2023 2:15:36 PM

Method: AnionsIC2-101323.mtw
Run operator: wet
Analysis number: 36273

Last save: 10/19/2023 12:0

SAMPLE:
: IZ/MA
Vial number: 12
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 10/24/2023 2:33:07 PM
Printed by: wet


Ident: LB127952BLW
Analysis from: 10/23/2023 2:18:14 PM
File: _2023-10-23_

Last save: 10/24/2023 2:32:45 PM

Method: AnionsIC2-101323.mtw
Run operator: wet
Analysis number: 36274

Last save: 10/23/2023 1:42

SAMPLE:
: IZ/MA
Vial number: 13
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 10/24/2023 2:33:13 PM
Printed by: wet


Ident: LB127952BSW
Analysis from: 10/23/2023 2:53:51 PM
File: _2023-10-23_

Last save: 10/24/2023 2:32:45 PM

Method: AnionsIC2-101323.mtw
Run operator: wet
Analysis number: 36275

Last save: 10/23/2023 1:42

SAMPLE:
: IZ/MA
Vial number: 14
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	7.24	1.530	4.41	2.93	433.077	6.24
2	10.33	0.318	18.87	12.51	470.412	6.78
3	12.55	0.403	32.42	21.49	1006.200	14.49
4	16.45	0.474	18.77	12.44	698.747	10.06
5	18.97	0.588	21.31	14.12	987.060	14.22
6	23.63	0.905	23.68	15.69	1728.228	24.89
7	27.85	0.677	31.43	20.83	1618.932	23.32
7	33.00	0.699	150.90	100.00	6942.656	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/24/2023 2:33:45 PM
Printed by: wet


Ident: 05020-01
Analysis from: 10/23/2023 3:59:33 PM
File: _2023-10-23_

Last save: 10/23/2023 4:32:33 PM

Method: AnionsIC2-101323.mtw
Run operator: wet
Analysis number: 36276

Last save: 10/23/2023 1:42

SAMPLE:
: IZ/MA
Vial number: 30
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	0.00	0.000	0.00	0.00	0.000	0.00
3	12.38	0.357	1.16	20.80	30.346	15.26
4	0.00	0.000	0.00	0.00	0.000	0.00
5	18.68	0.544	4.40	79.11	168.564	84.74
6	0.00	0.000	0.00	0.00	0.000	0.00
7	0.00	0.000	0.00	0.00	0.000	0.00
7	33.00	0.129	5.55	99.91	198.910	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/24/2023 2:33:55 PM
Printed by: wet


Ident: O5020-02
Analysis from: 10/23/2023 4:35:11 PM
File: _2023-10-23_

Last save: 10/23/2023 5:08:11 PM

Method: AnionsIC2-101323.mtw
Run operator: wet
Analysis number: 36277

Last save: 10/23/2023 1:42

SAMPLE:
: I2/MA
Vial number: 31
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	0.00	0.000	0.00	0.00	0.000	0.00
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	18.75	0.573	2.56	99.97	115.900	100.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	0.00	0.000	0.00	0.00	0.000	0.00
7	33.00	0.082	2.56	99.97	115.900	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/24/2023 2:34:01 PM
Printed by: wet


Ident: O5020-03
Analysis from: 10/23/2023 5:10:49 PM
File: _2023-10-23_

Last save: 10/23/2023 5:43:49 PM

Method: AnionsIC2-101323.mtw
Run operator: wet
Analysis number: 36278

Last save: 10/23/2023 1:42

SAMPLE:
: IZ/MA
Vial number: 32
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	10.23	0.302	1.20	32.52	27.232	19.41
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	18.87	0.574	2.49	67.37	113.061	80.59
6	0.00	0.000	0.00	0.00	0.000	0.00
7	0.00	0.000	0.00	0.00	0.000	0.00
7	33.00	0.125	3.70	99.90	140.294	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/24/2023 2:34:15 PM
Printed by: wet


Ident: O5020-03MS
Analysis from: 10/23/2023 5:46:26 PM
File: _2023-10-23_

Last save: 10/23/2023 6:19:26 PM

Method: AnionsIC2-101323.mtw
Run operator: wet
Analysis number: 36279

Last save: 10/23/2023 1:42

SAMPLE:
: IZ/MA
Vial number: 33
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	7.01	1.446	4.47	3.05	418.444	6.34
2	10.07	0.308	19.19	13.09	464.965	7.05
3	12.23	0.391	30.57	20.85	923.359	14.00
4	16.02	0.457	17.90	12.21	640.064	9.70
5	18.45	0.566	23.40	15.96	1042.628	15.80
6	23.17	0.918	21.92	14.95	1616.213	24.50
7	27.31	0.670	29.16	19.89	1491.516	22.61
7	33.00	0.679	146.62	100.00	6597.189	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/24/2023 2:34:22 PM
Printed by: wet


Ident: O5020-03MSD
Analysis from: 10/23/2023 6:22:04 PM
File: _2023-10-23_

Last save: 10/24/2023 9:42:50 AM

Method: AnionsIC2-101323.mtw
Run operator: wet
Analysis number: 36280

Last save: 10/23/2023 1:42

SAMPLE:
: IZ/MA
Vial number: 34
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	7.05	1.595	4.28	2.96	432.007	6.49
2	10.17	0.312	18.81	12.98	462.317	6.94
3	12.35	0.399	29.84	20.59	913.899	13.72
4	16.18	0.461	17.54	12.10	633.619	9.52
5	18.64	0.573	22.91	15.81	1037.573	15.58
6	23.38	0.929	22.95	15.83	1713.257	25.73
7	27.56	0.675	28.59	19.73	1466.167	22.02
7	33.00	0.706	144.92	99.99	6658.839	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/24/2023 2:34:38 PM
Printed by: wet


Ident: CCV
Analysis from: 10/23/2023 6:57:51 PM
File: _2023-10-23_

Last save: 10/23/2023 7:30:51 PM

Method: AnionsIC2-101323.mtw
Run operator: wet
Analysis number: 36281

Last save: 10/23/2023 1:42

SAMPLE:
: IZ/MA
Vial number: 11
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	7.21	1.504	4.59	2.96	443.905	6.21
2	10.28	0.313	19.33	12.47	475.222	6.64
3	12.49	0.399	33.08	21.35	1020.206	14.26
4	16.37	0.467	19.26	12.43	706.732	9.88
5	18.87	0.579	21.98	14.19	1006.222	14.07
6	23.60	0.917	25.25	16.29	1872.253	26.18
7	27.84	0.680	31.47	20.31	1627.302	22.75
7	33.00	0.694	154.96	100.00	7151.842	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/24/2023 2:34:45 PM
Printed by: wet


Ident: CCB
Analysis from: 10/23/2023 7:33:29 PM
File: _2023-10-23_

Last save: 10/23/2023 8:06:29 PM

Method: AnionsIC2-101323.mtw
Run operator: wet
Analysis number: 36282

Last save: 10/23/2023 1:42

SAMPLE:
: IZ/MA
Vial number: 12
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD