


LB127967

SFAM01.1

INSTRUMENT 15: CV1

Linear



A= 0.0000e+000

B= 1.9196e-004 slope

C= 1.2793e-002 y-intercept

Rho= 0.9995896

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	of 10
0.00	0.000	0.039	0.039	138	0.000	138					1
0.05	0.050					419	0				20
0.20	0.200	0.240	0.040	1185	0.0 %	1185					5
2.50	2.500	2.382	-0.118	12344	0.0 %	12344					1
5.00	5.000	4.936	-0.064	25645	0.0 %	25645					1
7.50	7.500	7.694	0.194	40014	0.0 %	40014					3
10.0	10.000	9.909	-0.091	51553	0.0 %	51553					1

LB127967

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	138	-		0 SFAM01.1	PPB	10/25/2023 11:33	Std
	0.2 S01	1185	-		0.2 SFAM01.1	PPB	10/25/2023 11:35	Std
	2.5 S02	12344	-		2.5 SFAM01.1	PPB	10/25/2023 11:37	Std
	5 S03	25645	-		5 SFAM01.1	PPB	10/25/2023 11:42	Std
	7.5 S04	40014	-		7.5 SFAM01.1	PPB	10/25/2023 11:44	Std
	10 S05	51553	-		10 SFAM01.1	PPB	10/25/2023 11:47	Std
ICV035	ICV035	21288	4.0992 -		SFAM01.1	PPB	10/25/2023 11:50	SMPL
ICB035	ICB035	-134	-0.0129 -		SFAM01.1	PPB	10/25/2023 11:52	SMPL
CCV076	CCV076	26486	5.097 -		SFAM01.1	PPB	10/25/2023 11:54	SMPL
CCB076	CCB076	-5	0.0118 -		SFAM01.1	PPB	10/25/2023 11:59	SMPL
PB156625BL	PBS925	31	0.0187 -		SFAM01.1	PPB	10/25/2023 12:01	SMPL
O4847-01	YC659	5427	1.0545 -		SFAM01.1	PPB	10/25/2023 12:03	SMPL
O4847-02	YC664	5335	1.0369 -		SFAM01.1	PPB	10/25/2023 12:05	SMPL
O4847-03	YC665	4008	0.7822 -		SFAM01.1	PPB	10/25/2023 12:08	SMPL
O4847-04	YC666	5367	1.043 -		SFAM01.1	PPB	10/25/2023 12:10	SMPL
O4847-05	YC667	8740	1.6905 -		SFAM01.1	PPB	10/25/2023 12:12	SMPL
O4847-06	YC668	8376	1.6206 -		SFAM01.1	PPB	10/25/2023 12:14	SMPL
O4847-07	YC669	4846	0.943 -		SFAM01.1	PPB	10/25/2023 12:17	SMPL
O4847-08	YC669D	5758	1.1181 -		SFAM01.1	PPB	10/25/2023 12:21	SMPL
O4847-09	YC669S	19941	3.8406 -		SFAM01.1	PPB	10/25/2023 12:24	SMPL
O4847-10	YC670	7466	1.4459 -		SFAM01.1	PPB	10/25/2023 12:26	SMPL
O4847-11	YC671	4965	0.9659 -		SFAM01.1	PPB	10/25/2023 12:28	SMPL
O4847-12	YC672	10482	2.0249 -		SFAM01.1	PPB	10/25/2023 12:30	SMPL
O4847-13	YC622	3364	0.6585 -		SFAM01.1	PPB	10/25/2023 12:33	SMPL
O4847-14	YC623	4741	0.9229 -		SFAM01.1	PPB	10/25/2023 12:35	SMPL
O4847-15	YC638	50106	9.631 -		SFAM01.1	PPB	10/25/2023 12:37	SMPL
O4847-16	YC639	78708	15.1214 -		SFAM01.1	PPB	10/25/2023 12:39	SMPL
O4847-18	YC655	93683	17.996 -		SFAM01.1	PPB	10/25/2023 12:42	SMPL
O4847-19	YC656	30878	5.9401 -		SFAM01.1	PPB	10/25/2023 12:44	SMPL
O4847-20	YC682	53867	10.353 -		SFAM01.1	PPB	10/25/2023 12:46	SMPL
CCV077	CCV077	25617	4.9302 -		SFAM01.1	PPB	10/25/2023 12:51	SMPL
CCB077	CCB077	-33	0.0065 -		SFAM01.1	PPB	10/25/2023 12:54	SMPL
O4847-21	YC683	38529	7.4087 -		SFAM01.1	PPB	10/25/2023 12:56	SMPL
O4847-22	YC684	50183	9.6458 -		SFAM01.1	PPB	10/25/2023 12:58	SMPL
PB156627BL	PBS627	-81	-0.0028 -		SFAM01.1	PPB	10/25/2023 13:00	SMPL
O4849-01	YC610	10315	1.9928 -		SFAM01.1	PPB	10/25/2023 13:03	SMPL
O4849-02	YC611	6813	1.3206 -		SFAM01.1	PPB	10/25/2023 13:05	SMPL
O4849-03	YC624	9389	1.8151 -		SFAM01.1	PPB	10/25/2023 13:07	SMPL
O4849-04	YC625	8603	1.6642 -		SFAM01.1	PPB	10/25/2023 13:09	SMPL
O4849-05	YC626	26251	5.0519 -		SFAM01.1	PPB	10/25/2023 13:12	SMPL
O4849-06	YC627	6589	1.2776 -		SFAM01.1	PPB	10/25/2023 13:14	SMPL
O4849-07	YC648	5996	1.1638 -		SFAM01.1	PPB	10/25/2023 13:16	SMPL
O4849-08	YC649	29927	5.7575 -		SFAM01.1	PPB	10/25/2023 13:19	SMPL
O4849-09	YC650	5026	0.9776 -		SFAM01.1	PPB	10/25/2023 13:21	SMPL
O4849-10	YC651	8879	1.7172 -		SFAM01.1	PPB	10/25/2023 13:23	SMPL
O4849-11	YC620	3947	0.7705 -		SFAM01.1	PPB	10/25/2023 13:25	SMPL

INSTRUMENT ID : CV1

O4849-12	YC620D	3702	0.7234 -	SFAM01.1	PPB	10/25/2023 13:28 SMPL
O4849-13	YC620S	20073	3.866 -	SFAM01.1	PPB	10/25/2023 13:30 SMPL
O4849-14	YC621	16196	3.1217 -	SFAM01.1	PPB	10/25/2023 13:32 SMPL
O4849-15	YC628	1784	0.3552 -	SFAM01.1	PPB	10/25/2023 13:34 SMPL
O4849-16	YC629	2847	0.5593 -	SFAM01.1	PPB	10/25/2023 13:37 SMPL
O4849-17	YC632	3085	0.605 -	SFAM01.1	PPB	10/25/2023 13:39 SMPL
CCV078	CCV078	25730	4.9519 -	SFAM01.1	PPB	10/25/2023 13:41 SMPL
CCB078	CCB078	-1	0.0126 -	SFAM01.1	PPB	10/25/2023 13:43 SMPL
O4849-18	YC633	2825	0.5551 -	SFAM01.1	PPB	10/25/2023 13:46 SMPL
O4849-19	YC634	4127	0.805 -	SFAM01.1	PPB	10/25/2023 13:48 SMPL
O4849-20	YC635	1712	0.3414 -	SFAM01.1	PPB	10/25/2023 13:50 SMPL
O4849-21	YC636	4961	0.9651 -	SFAM01.1	PPB	10/25/2023 13:52 SMPL
O4849-22	YC637	2845	0.5589 -	SFAM01.1	PPB	10/25/2023 13:55 SMPL
PB156628BL	PBS628	7	0.0141 -	SFAM01.1	PPB	10/25/2023 13:57 SMPL
O4850-01	YC630	12055	2.3268 -	SFAM01.1	PPB	10/25/2023 13:59 SMPL
O4850-02	YC631	26504	5.1004 -	SFAM01.1	PPB	10/25/2023 14:02 SMPL
O4850-03	YC640	2824	0.5549 -	SFAM01.1	PPB	10/25/2023 14:04 SMPL
O4850-04	YC641	4041	0.7885 -	SFAM01.1	PPB	10/25/2023 14:06 SMPL
O4850-05	YC642	2715	0.534 -	SFAM01.1	PPB	10/25/2023 14:08 SMPL
O4850-06	YC642D	2846	0.5591 -	SFAM01.1	PPB	10/25/2023 14:11 SMPL
O4850-07	YC642S	18430	3.5506 -	SFAM01.1	PPB	10/25/2023 14:13 SMPL
O4850-08	YC654	25955	4.9951 -	SFAM01.1	PPB	10/25/2023 14:15 SMPL
O4850-09	YC660	4138	0.8071 -	SFAM01.1	PPB	10/25/2023 14:17 SMPL
O4850-10	YC678	5150	1.0014 -	SFAM01.1	PPB	10/25/2023 14:20 SMPL
O4850-11	YC680	25719	4.9498 -	SFAM01.1	PPB	10/25/2023 14:22 SMPL
O4850-12	YC681	13602	2.6238 -	SFAM01.1	PPB	10/25/2023 14:24 SMPL
O4850-13	YC5T5	50571	9.7203 -	SFAM01.1	PPB	10/25/2023 14:27 SMPL
O4850-14	YC643	34959	6.7234 -	SFAM01.1	PPB	10/25/2023 14:29 SMPL
CCV079	CCV079	27048	5.2049 -	SFAM01.1	PPB	10/25/2023 14:31 SMPL
CCB079	CCB079	-24	0.0082 -	SFAM01.1	PPB	10/25/2023 14:33 SMPL
O4850-15	YC644	34431	6.6221 -	SFAM01.1	PPB	10/25/2023 14:36 SMPL
O4850-16	YC645	31470	6.0537 -	SFAM01.1	PPB	10/25/2023 14:38 SMPL
O4850-17	YC673	6951	1.3471 -	SFAM01.1	PPB	10/25/2023 14:40 SMPL
O4850-18	YC674	31216	6.0049 -	SFAM01.1	PPB	10/25/2023 14:42 SMPL
O4850-19	YC675	3867	0.7551 -	SFAM01.1	PPB	10/25/2023 14:45 SMPL
O4850-20	YC676	14984	2.8891 -	SFAM01.1	PPB	10/25/2023 14:47 SMPL
O4850-21	YC677	14569	2.8094 -	SFAM01.1	PPB	10/25/2023 14:49 SMPL
O4850-22	YC679	6025	1.1693 -	SFAM01.1	PPB	10/25/2023 14:52 SMPL
PB156629BL	PBS629	55	0.0234 -	SFAM01.1	PPB	10/25/2023 14:54 SMPL
O4856-01	YC602	9123	1.764 -	SFAM01.1	PPB	10/25/2023 14:56 SMPL
O4856-02	YC603	25585	4.924 -	SFAM01.1	PPB	10/25/2023 14:58 SMPL
O4856-03	YC604	26081	5.0192 -	SFAM01.1	PPB	10/25/2023 15:01 SMPL
O4856-04	YC605	7628	1.477 -	SFAM01.1	PPB	10/25/2023 15:03 SMPL
O4856-05	YC606	8851	1.7118 -	SFAM01.1	PPB	10/25/2023 15:05 SMPL
O4856-06	YC607	6274	1.2171 -	SFAM01.1	PPB	10/25/2023 15:07 SMPL
O4856-07	YC608	5516	1.0716 -	SFAM01.1	PPB	10/25/2023 15:10 SMPL
O4856-08	YC609	1076	0.2193 -	SFAM01.1	PPB	10/25/2023 15:17 SMPL

LB127967

INSTRUMENT ID : CV1

CCV080	CCV080	28265	5.4385 -	SFAM01.1	PPB	10/25/2023 15:30 SMPL
CCB080	CCB080	0	0.0128 -	SFAM01.1	PPB	10/25/2023 15:32 SMPL
O4847-18DLX5	YC655	18750	3.612 -	SFAM01.1	PPB	10/25/2023 15:37 SMPL
O4847-20DLX2	YC682	29056	5.5903 -	SFAM01.1	PPB	10/25/2023 15:39 SMPL
O4856-09		1116	0.227 -	SFAM01.1	PPB	10/25/2023 15:45 SMPL
O4856-10		13876	2.6764 -	SFAM01.1	PPB	10/25/2023 15:47 SMPL
O4847-16DLX5		15868	3.0588 -	SFAM01.1	PPB	10/25/2023 15:49 SMPL
CCV081	CCV081	25110	4.8329 -	SFAM01.1	PPB	10/25/2023 15:51 SMPL
CCB081	CCB081	-43	0.0045 -	SFAM01.1	PPB	10/25/2023 15:54 SMPL