



PERCENT SOLID

Supervisor: Iwona
Analyst: JIGNESH
Date: 10/27/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 17:15
In Date: 10/25/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 10/26/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB127969

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O5027-01	214	1	1.00	1.00	2.00	2.00	100.0	STONE SAMPLE,100% SOLIDS
O5027-02	214-EPH	2	1.00	1.00	2.00	2.00	100.0	STONE SAMPLE,100% SOLIDS
O5027-04	H-DRUMS-SOLID	3	1.14	8.59	9.73	8.33	83.7	
O5027-05	H-DRUMS-SOLID-EPH	4	1.15	8.81	9.96	8.16	79.6	
O5036-01	BUR-1277	5	1.00	1.00	2.00	2.00	100.0	debris
O5038-01	AUD-1059	6	1.14	8.38	9.52	9.33	97.7	
O5038-03	AUD-1068	7	1.00	1.00	2.00	2.00	100.0	debris
O5039-01	MMFB-1	8	1.15	8.40	9.55	8.63	89.0	
O5039-02	MMFB-2	9	1.15	8.40	9.55	8.6	88.7	
O5039-03	MMFB-3	10	1.14	8.80	9.94	9.18	91.4	
O5039-04	MMFB-4	11	1.15	8.82	9.97	9.03	89.3	
O5040-01	M-4 (50FT) (0-5)	12	1.18	8.45	9.63	8.58	87.6	
O5046-01	001	13	1.15	8.78	9.93	6.15	56.9	
O5046-02	002	14	1.18	8.59	9.77	6.00	56.1	
O5047-01	PT-AN-SOIL	15	1.00	1.00	2.00	2.00	100.0	
O5047-02	PT-CORR-SOIL	16	1.00	1.00	2.00	2.00	100.0	
O5047-03	PT-CN-SOIL	17	1.00	1.00	2.00	2.00	100.0	
O5047-04	PT-CN-SOIL	18	1.00	1.00	2.00	2.00	100.0	
O5047-05	PT-FP-SOIL	19	1.00	1.00	2.00	2.00	100.0	
O5047-06	PT-CR6-SOIL	20	1.00	1.00	2.00	2.00	100.0	
O5047-07	PT-NUT-SOIL	21	1.00	1.00	2.00	2.00	100.0	
O5047-08	PT-NUT-SOIL	22	1.00	1.00	2.00	2.00	100.0	
O5047-09	PT-OGR-SOIL	23	1.00	1.00	2.00	2.00	100.0	
O5047-10	PT-MET-SOIL	24	1.00	1.00	2.00	2.00	100.0	
O5047-11	PT-BNA-SOIL	25	1.00	1.00	2.00	2.00	100.0	
O5047-12	PT-TRIAZINE-SOIL	26	1.00	1.00	2.00	2.00	100.0	
O5047-13	PT-PAH-SOIL	27	1.00	1.00	2.00	2.00	100.0	
O5047-14	PT-DIES-SOIL	28	1.00	1.00	2.00	2.00	100.0	



PERCENT SOLID

Supervisor: Iwona
Analyst: JIGNESH
Date: 10/27/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 17:15
In Date: 10/25/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 10/26/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB127969

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O5047-15	PT-GAS-SOIL	29	1.00	1.00	2.00	2.00	100.0	
O5047-16	PT-NJEPH-SOIL	30	1.00	1.00	2.00	2.00	100.0	
O5047-17	PT-HERB-SOIL	31	1.00	1.00	2.00	2.00	100.0	
O5047-18	PT-PCB-SOIL	32	1.00	1.00	2.00	2.00	100.0	
O5047-19	PT-PCBO-SOIL	33	1.00	1.00	2.00	2.00	100.0	
O5047-20	PT-PEST-SOIL	34	1.00	1.00	2.00	2.00	100.0	
O5047-21	PT-CHLR-SOIL	35	1.00	1.00	2.00	2.00	100.0	
O5047-22	PT-TXP-SOIL	36	1.00	1.00	2.00	2.00	100.0	
O5047-23	PT-VOA-SOIL	37	1.00	1.00	2.00	2.00	100.0	
O5047-24	PT-SOL-SOIL	38	1.14	8.48	9.62	7.99	80.8	
O5050-06	B1	39	1.2	8.38	9.58	8.52	87.4	
O5050-07	B2	40	1.18	8.56	9.74	8.75	88.4	
O5050-08	B3	41	1.16	8.74	9.9	9.47	95.1	
O5050-09	B4	42	1.18	8.81	9.99	8.81	86.6	
O5050-10	B5	43	1.17	8.58	9.75	9.3	94.8	
O5051-01	DUCT-BANK-1A	44	1.15	8.81	9.96	9.22	91.6	
O5051-02	DUCT-BANK-1A-E2	45	1.13	8.54	9.67	8.37	84.8	
O5051-03	DUCT-BANK-2A	46	1.12	8.67	9.79	8.59	86.2	
O5051-04	DUCT-BANK-2A-E2	47	1.13	8.84	9.97	8.95	88.5	
O5073-01	OR-02-102523	48	1.18	8.56	9.74	9.06	92.1	
O5073-02	OR-02-102523-E2	49	1.16	8.69	9.85	8.92	89.3	
O5075-01	SWC-1	50	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
O5075-02	SWG-1	51	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
O5075-03	ADC-1	52	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
O5075-04	AWC-1	53	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
O5075-05	AWG-1	54	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
O5075-06	AWG-2	55	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
O5075-07	RWC-1	56	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE



PERCENT SOLID

Supervisor: Iwona
Analyst: JIGNESH
Date: 10/27/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 17:15
In Date: 10/25/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 10/26/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB127969

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O5075-08	RDCR-1	57	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
O5075-09	RWG-1	58	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
O5075-10	JWG-1	59	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
O5075-11	WC-1	60	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
O5075-12	WC-2	61	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
O5077-01	LP-1	70	1.00	1.00	2.00	2.00	100.0	PAINT CHIPS
O5077-03	LP-2	71	1.00	1.00	2.00	2.00	100.0	PAINT CHIPS
O5077-05	LP-3	72	1.00	1.00	2.00	2.00	100.0	PAINT CHIPS
O5077-07	SOL-1	73	1.16	8.48	9.64	8.48	86.3	
O5077-09	SOL-2	74	1.11	8.72	9.83	9.41	95.2	
O5080-01	HEX-1	62	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O5080-03	HEX-2	63	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O5080-05	HEX-3	64	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O5080-07	HEX-4	65	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O5080-09	HEX-5	66	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O5080-11	HEX-6	67	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O5081-01	CONTAINMENT	68	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
O5081-03	BALLARDS	69	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$