



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB127973

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	10/25/2023	11:35
2	CAL2	1	Water	NA	NA	20.3	7.00	10/25/2023	11:36
3	CAL3	1	Water	NA	NA	20.3	10.02	10/25/2023	11:37
4	ICV	1	Water	NA	NA	20.6	6.99	10/25/2023	11:40
5	CCV1	1	Water	NA	NA	20.6	2.01	10/25/2023	11:44
6	O5027-07	1	Water	NA	NA	20.8	6.65	10/25/2023	11:50
7	O5027-07DUP	1	Water	NA	NA	20.7	6.66	10/25/2023	11:51
8	O5027-09	1	Water	NA	NA	20.3	6.69	10/25/2023	11:54
9	CCV2	1	Water	NA	NA	20.6	12.02	10/25/2023	12:00

VB 127413

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORRSOVITY O5028

WorkList ID : 174929

Department : Wet-Chemistry

Date : 10-25-2023 11:26:49

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5027-07	H-DRUMS-WATER	Water	Corrosivity	Cool 4 deg C	PSEG03	I31	10/24/2023	9040C
O5027-09	H-TOTE-WATER	Water	Corrosivity	Cool 4 deg C	PSEG03	I31	10/24/2023	9040C

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

10/25/23 11:30

JD (wcl)

JDCSM

Date/Time 10/25/23

Raw Sample Received by:

Raw Sample Relinquished by:

13/1.0c

JDCSM

JDCSM