



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB127975

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	10/25/2023	11:35
2	CAL2	1	Water	NA	NA	20.3	7.00	10/25/2023	11:36
3	CAL3	1	Water	NA	NA	20.3	10.02	10/25/2023	11:37
4	ICV	1	Water	NA	NA	20.6	6.99	10/25/2023	11:40
5	CCV1	1	Water	NA	NA	20.7	2.02	10/25/2023	12:30
6	O5030-01	1	Water	NA	NA	20.5	7.12	10/25/2023	12:35
7	O5030-01DUP	1	Water	NA	NA	20.6	7.13	10/25/2023	12:36
8	O5038-05	1	Water	NA	NA	20.6	7.22	10/25/2023	12:40
9	O5038-07	1	Water	NA	NA	20.2	6.06	10/25/2023	12:45
10	CCV2	1	Water	NA	NA	20.7	12.02	10/25/2023	12:46

10/25/23

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH O5030 WorkList ID : 174910 Department : Wet-Chemistry Date : 10-25-2023 10:38:19

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5030-01	TRE-1723	Water	pH	Cool 4 deg C	PSEG03	I41	10/24/2023	9040C
O5038-05	AUD-1067	Water	pH	Cool 4 deg C	PSEG03	I41	10/24/2023	9040C
O5038-07	AUD-1069	Water	pH	Cool 4 deg C	PSEG03	I41	10/24/2023	9040C

Date/Time 10/25/23 10:45
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 10/25/23 13:30
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]