



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB127980

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	81.7	10/25/2023	11:10
2	O5027-07		1	12	100.00	>212.0	>212.0	10/25/2023	11:40
3	O5027-09		1	14	100.00	>212.0	>212.0	10/25/2023	12:10
4	O5030-01		1	14	100.00	>212.0	>212.0	10/25/2023	12:40
5	O5036-02		1	12	100.00	>212.0	>212.0	10/25/2023	13:10
6	O5036-02DUP		1	12	100.00	>212.0	>212.0	10/25/2023	13:40
7	O5038-05		1	14	100.00	>212.0	>212.0	10/25/2023	14:10
8	O5038-07		1	12	100.00	>212.0	>212.0	10/25/2023	14:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

10/27/2023

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp-10-25 WorkList ID : 174896 Department : Wet-Chemistry Date : 10-25-2023 08:22:32

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5027-07 C	H-DRUMS-WATER	Water	Flash Point	Cool 4 deg C	PSEG03	I31	10/24/2023	1010B
O5027-09 E	H-TOTE-WATER	Water	Flash Point	Cool 4 deg C	PSEG03	I31	10/24/2023	1010B
O5030-01 E	TRE-1723	Water	Flash Point	Cool 4 deg C	PSEG03	I41	10/24/2023	1010B
O5036-02	BUR-1282	Water	Flash Point	Cool 4 deg C	PSEG03	I31	10/24/2023	1010B
O5038-05 E	AUD-1067	Water	Flash Point	Cool 4 deg C	PSEG03	I41	10/24/2023	1010B
O5038-07 E	AUD-1069	Water	Flash Point	Cool 4 deg C	PSEG03	I41	10/24/2023	1010B

Date/Time 10/25/2023 11:00
 Raw Sample Received by: RPA cwc
 Raw Sample Relinquished by: JAL cwc

Date/Time 10/25/2023 15:45
 Raw Sample Received by: JAL cwc
 Raw Sample Relinquished by: RPA cwc