


LB127994

SFAM01.1

INSTRUMENT ID: (V2)



Linear

A= 0.0000e+000

B= 1.5710e-004 slope

C= -4.5300e-002 y-intercept

Rho= 0.9998265

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.009	-0.009	234	0.000	234					0.0
0.05	0.050					419	0				-14
0.20	0.200	0.172	-0.028	1386	0.0 %	1386					0
2.50	2.500	2.507	0.007	16247	0.0 %	16247					0
5.00	5.000	5.122	0.122	32888	0.0 %	32888					-2
7.50	7.500	7.389	-0.111	47320	0.0 %	47320					-1
10.0	10.000	10.018	0.018	64058	0.0 %	64058					0

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	234	-		0 SFAM01.1	PPB	10/26/2023 14:51	Std
	0.2 S01	1386	-		0.2 SFAM01.1	PPB	10/26/2023 14:53	Std
	2.5 S02	16247	-		2.5 SFAM01.1	PPB	10/26/2023 14:56	Std
	5 S03	32888	-		5 SFAM01.1	PPB	10/26/2023 14:58	Std
	7.5 S04	47320	-		7.5 SFAM01.1	PPB	10/26/2023 15:03	Std
	10 S05	64058	-		10 SFAM01.1	PPB	10/26/2023 15:08	Std
ICV037	ICV037	23985	3.7228 -		SFAM01.1	PPB	10/26/2023 15:11	SMPL
ICB037	ICB037	-189	-0.075 -		SFAM01.1	PPB	10/26/2023 15:13	SMPL
CCV084	CCV084	31206	4.8573 -		SFAM01.1	PPB	10/26/2023 15:15	SMPL
CCB084	CCB084	-213	-0.0788 -		SFAM01.1	PPB	10/26/2023 15:17	SMPL
PB156677BL	PBS677	-86	-0.0588 -		SFAM01.1	PPB	10/26/2023 15:20	SMPL
O4877-01	MEZZ60	14316	2.2038 -		SFAM01.1	PPB	10/26/2023 15:22	SMPL
O4877-02	MEZZ80	98050	15.3588 -		SFAM01.1	PPB	10/26/2023 15:24	SMPL
O4877-03	MEZZ81	120069	18.8181 -		SFAM01.1	PPB	10/26/2023 15:26	SMPL
O4877-04	MEZZ82	35578	5.5442 -		SFAM01.1	PPB	10/26/2023 15:29	SMPL
O4877-05	MEZZ50	8168	1.2379 -		SFAM01.1	PPB	10/26/2023 15:31	SMPL
O4877-06	MEZZ51	132	-0.0246 -		SFAM01.1	PPB	10/26/2023 15:33	SMPL
O4877-07	MEZZ55	2276	0.3123 -		SFAM01.1	PPB	10/26/2023 15:36	SMPL
O4877-08	MEZZ56	548	0.0408 -		SFAM01.1	PPB	10/26/2023 15:38	SMPL
O4877-09	MEZZ57	589	0.0472 -		SFAM01.1	PPB	10/26/2023 15:40	SMPL
O4877-10	MEZZ95	6829	1.0276 -		SFAM01.1	PPB	10/26/2023 15:42	SMPL
O4877-11	MEZZ96	14751	2.2721 -		SFAM01.1	PPB	10/26/2023 15:45	SMPL
O4877-12	MEZZ97	13482	2.0728 -		SFAM01.1	PPB	10/26/2023 15:47	SMPL
O4877-13	MEZZ98	2878	0.4068 -		SFAM01.1	PPB	10/26/2023 15:49	SMPL
O4877-14	MEZZ99	265	-0.0037 -		SFAM01.1	PPB	10/26/2023 15:51	SMPL
O4877-15	MEZZ30	87363	13.6798 -		SFAM01.1	PPB	10/26/2023 15:54	SMPL
O4877-16	MEZZ31	111123	17.4126 -		SFAM01.1	PPB	10/26/2023 15:56	SMPL
O4877-17	MEZZ32	5472	0.8144 -		SFAM01.1	PPB	10/26/2023 15:58	SMPL
O4877-18	MEZZ35	21969	3.4061 -		SFAM01.1	PPB	10/26/2023 16:00	SMPL
O4877-19	MEZZ45	7248	1.0934 -		SFAM01.1	PPB	10/26/2023 16:03	SMPL
CCV085	CCV085	30253	4.7076 -		SFAM01.1	PPB	10/26/2023 16:05	SMPL
CCB085	CCB085	-49	-0.053 -		SFAM01.1	PPB	10/26/2023 16:07	SMPL
O4877-20	MEZZ45D	7577	1.1451 -		SFAM01.1	PPB	10/26/2023 16:10	SMPL
O4877-21	MEZZ45S	23022	3.5716 -		SFAM01.1	PPB	10/26/2023 16:14	SMPL
PB156678BL	PBS678	-118	-0.0638 -		SFAM01.1	PPB	10/26/2023 16:16	SMPL
O4921-02	MEZZ85	73083	11.4364 -		SFAM01.1	PPB	10/26/2023 16:19	SMPL
O4921-03	MEZZ86	6285	0.9421 -		SFAM01.1	PPB	10/26/2023 16:21	SMPL
O4921-04	MEZZ87	1746	0.229 -		SFAM01.1	PPB	10/26/2023 16:23	SMPL
O4921-05	MEZZA0	3279	0.4698 -		SFAM01.1	PPB	10/26/2023 16:25	SMPL
O4921-06	MEZZA1	2595	0.3624 -		SFAM01.1	PPB	10/26/2023 16:28	SMPL
O4921-07	MEZZA2	391	0.0161 -		SFAM01.1	PPB	10/26/2023 16:30	SMPL
O4921-08	MEZZA3	258	-0.0048 -		SFAM01.1	PPB	10/26/2023 16:32	SMPL
O4921-09	MEZZA4	183	-0.0165 -		SFAM01.1	PPB	10/26/2023 16:34	SMPL
O4921-13	MEZZ20	18572	2.8724 -		SFAM01.1	PPB	10/26/2023 16:42	SMPL
O4921-14	MEZZ20D	18051	2.7906 -		SFAM01.1	PPB	10/26/2023 16:44	SMPL
O4921-15	MEZZ20S	32815	5.1101 -		SFAM01.1	PPB	10/26/2023 16:46	SMPL

LB127994

INSTRUMENT ID : CV1

O4921-16	MEZZ25	2092	0.2834 -	SFAM01.1	PPB	10/26/2023 16:49	SMPL
O4921-17	MEZZ26	8913	1.355 -	SFAM01.1	PPB	10/26/2023 16:51	SMPL
O4921-18	MEZZ27	259	-0.0046 -	SFAM01.1	PPB	10/26/2023 16:53	SMPL
O4921-19	MEZZC0	2682	0.3761 -	SFAM01.1	PPB	10/26/2023 16:55	SMPL
PB156679BL	PBS679	-11	-0.047 -	SFAM01.1	PPB	10/26/2023 16:58	SMPL
O4949-01	MCOAG3	1080	0.1244 -	SFAM01.1	PPB	10/26/2023 17:00	SMPL
CCV086	CCV086	29362	4.5676 -	SFAM01.1	PPB	10/26/2023 17:02	SMPL
CCB086	CCB086	-28	-0.0497 -	SFAM01.1	PPB	10/26/2023 17:04	SMPL
O4949-02	MCOAG4	78	-0.033 -	SFAM01.1	PPB	10/26/2023 17:07	SMPL
O4949-03	MCOAH5	1758	0.2309 -	SFAM01.1	PPB	10/26/2023 17:09	SMPL
O4949-04	MCOAH6	65	-0.0351 -	SFAM01.1	PPB	10/26/2023 17:11	SMPL
O4949-05	MCOAJ4	8035	1.217 -	SFAM01.1	PPB	10/26/2023 17:14	SMPL
O4949-06	MCOAJ6	58	-0.0362 -	SFAM01.1	PPB	10/26/2023 17:16	SMPL
O4949-07	MCOAK1	1496	0.1897 -	SFAM01.1	PPB	10/26/2023 17:18	SMPL
O4949-08	MCOAK1D	1389	0.1729 -	SFAM01.1	PPB	10/26/2023 17:20	SMPL
O4949-09	MCOAK1S	15974	2.4643 -	SFAM01.1	PPB	10/26/2023 17:23	SMPL
O4949-10	MCOAK2	25	-0.0414 -	SFAM01.1	PPB	10/26/2023 17:25	SMPL
O4949-11	MCOAK4	353	0.0102 -	SFAM01.1	PPB	10/26/2023 17:27	SMPL
O4949-12	MCOAL4	2087	0.2826 -	SFAM01.1	PPB	10/26/2023 17:29	SMPL
O4949-13	MCOAL5	447	0.0249 -	SFAM01.1	PPB	10/26/2023 17:32	SMPL
CCV087	CCV087	30384	4.7282 -	SFAM01.1	PPB	10/26/2023 17:34	SMPL
CCB087	CCB087	-83	-0.0583 -	SFAM01.1	PPB	10/26/2023 17:36	SMPL