


LB128002

7470A

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.2564e-004

C= 1.4759e-002

Rho= 0.9994663

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.032	0.032	76	0.000	76	0				1
0.2	0.200	0.216	0.016	892	0.0 %	892					8
2.5	2.500	2.493	-0.007	10985	0.0 %	10985					0
5.0	5.000	5.055	0.055	22337	0.0 %	22337					1
7.5	7.500	7.254	-0.246	32083	0.0 %	32083					3
10.0	10.000	10.150	0.150	44915	0.0 %	44915					2

LB128002 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 27 Oct 2023 10:31:14

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	76	-	0.0000	7470A	PPB	27 Oct 2023 10:41:46	S	Std
0.2 - 1	50.2	892	-	0.2000	7470A	PPB	27 Oct 2023 10:44:03	S	Std
2.5 - 1	52.5	10985	-	2.5000	7470A	PPB	27 Oct 2023 10:46:20	S	Std
5.0 - 1	55	22337	-	5.0000	7470A	PPB	27 Oct 2023 10:48:37	S	Std
7.5 - 1	57.5	32083	-	7.5000	7470A	PPB	27 Oct 2023 10:50:53	S	Std
10.0 - 1	510	44915	-	10.0000	7470A	PPB	27 Oct 2023 10:53:10	S	Std
ICV39 - 1	ICV39	16713	3.7859	-	7470A	PPB	27 Oct 2023 10:55:56	U	SMPL
ICB39 - 1	ICB39	-329	-0.0595	-	7470A	PPB	27 Oct 2023 10:58:12	U	SMPL
CCV91 - 1	CCV91	21688	4.9085	-	7470A	PPB	27 Oct 2023 11:00:29	U	SMPL
CCB91 - 1	CCB91	-325	-0.0586	-	7470A	PPB	27 Oct 2023 11:02:44	U	SMPL
CRA - 1	CRA	673	0.1666	-	7470A	PPB	27 Oct 2023 11:05:01	U	SMPL
HighStd - 1	HighStd	42774	9.6664	-	7470A	PPB	27 Oct 2023 11:07:16	U	SMPL
ChkStd - 1	ChkStd	29053	6.5704	-	7470A	PPB	27 Oct 2023 11:09:32	U	SMPL
PB156668BL - 1	PBW	-474	-0.0922	-	7470A	PPB	27 Oct 2023 11:11:52	U	SMPL
O5027-03 - 1	214	-241	-0.0396	-	7470A	PPB	27 Oct 2023 11:16:28	U	SMPL
O5027-03DUP - 1	214DUP	-66	-0.0001	-	7470A	PPB	27 Oct 2023 11:18:44	U	SMPL
O5027-03MS - 1	214MS	16812	3.8083	-	7470A	PPB	27 Oct 2023 11:27:47	U	SMPL
O5027-03MSD - 1	214MSD	17615	3.9895	-	7470A	PPB	27 Oct 2023 11:30:03	U	SMPL
O5027-06 - 1	H-DRUMS-SOLID	-116	-0.0114	-	7470A	PPB	27 Oct 2023 11:32:20	U	SMPL
CCV92 - 1	CCV92	21760	4.9248	-	7470A	PPB	27 Oct 2023 11:34:36	U	SMPL
CCB92 - 1	CCB92	-334	-0.0606	-	7470A	PPB	27 Oct 2023 11:36:51	U	SMPL
O5038-02 - 1	AUD-1059	-41	0.0055	-	7470A	PPB	27 Oct 2023 11:41:23	U	SMPL
O5038-04 - 1	AUD-1068	-24	0.0093	-	7470A	PPB	27 Oct 2023 11:43:38	U	SMPL
PB156669BL - 1	PBW	-46	0.0044	-	7470A	PPB	27 Oct 2023 11:45:54	U	SMPL
PB156669BS - 1	LCSW	17440	3.9500	-	7470A	PPB	27 Oct 2023 11:48:10	U	SMPL
O5030-02 - 1	TRE-1723	97	0.0366	-	7470A	PPB	27 Oct 2023 11:50:26	U	SMPL
O5030-02DUP - 1	TRE-1723DUP	50	0.0260	-	7470A	PPB	27 Oct 2023 11:52:43	U	SMPL
O5030-02MS - 1	TRE-1723MS	18149	4.1100	-	7470A	PPB	27 Oct 2023 11:57:41	U	SMPL
O5030-02MSD - 1	TRE-1723MSD	18227	4.1276	-	7470A	PPB	27 Oct 2023 11:59:58	U	SMPL
O5038-06 - 1	AUD-1067	-461	-0.0893	-	7470A	PPB	27 Oct 2023 12:02:16	U	SMPL
CCV93 - 1	CCV93	22234	5.0317	-	7470A	PPB	27 Oct 2023 12:04:33	U	SMPL
CCB93 - 1	CCB93	-399	-0.0753	-	7470A	PPB	27 Oct 2023 12:06:49	U	SMPL
O5038-08 - 1	AUD-1069	14	0.0179	-	7470A	PPB	27 Oct 2023 12:09:06	U	SMPL
PB156668BS - 1	LCSW	17720	4.0132	-	7470A	PPB	27 Oct 2023 12:11:22	U	SMPL
PB156634TB - 1	PB156634TB	-328	-0.0593	-	7470A	PPB	27 Oct 2023 12:13:38	U	SMPL
PB156643TB - 1	PB156643TB	-119	-0.0121	-	7470A	PPB	27 Oct 2023 12:15:56	U	SMPL
O5027-03LX5 - 1		-41	0.0055	-	7470A	PPB	27 Oct 2023 12:18:12	U	SMPL
O5027-03A - 1		18829	4.2634	-	7470A	PPB	27 Oct 2023 12:23:17	U	SMPL
O5030-02LX5 - 1		-212	-0.0331	-	7470A	PPB	27 Oct 2023 12:25:33	U	SMPL
O5030-02A - 1		20547	4.6511	-	7470A	PPB	27 Oct 2023 12:27:52	U	SMPL
CCV94 - 1	CCV94	20551	4.6520	-	7470A	PPB	27 Oct 2023 12:30:09	U	SMPL
CCB94 - 1	CCB94	-356	-0.0656	-	7470A	PPB	27 Oct 2023 12:32:25	U	SMPL