

LB128012


7471B

INSTRUMENT ID: CVJ

Linear

 μ Abs.:

7268E



A= 0.0000e+000

B= 1.3882e-004

C= 4.8049e-002

Rho= 0.9996070

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0/0
0.0	0.000	0.054	0.054	41	0.000	41					13
0.2	0.200	0.226	0.026	1280	0.0 %	1280					-2
2.5	2.500	2.450	-0.050	17305	0.0 %	17305					1
5.0	5.000	5.027	0.027	35868	0.0 %	35868					-3
7.5	7.500	7.304	-0.196	52272	0.0 %	52272					1
10.0	10.000	10.139	0.139	72688	0.0 %	72688					

LB128012 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 27 Oct 2023 16:10:17

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	41	-	0.0000	7471B	PPB	27 Oct 2023 16:14:25	S	Std
0.2 - 1	50.2	1280	-	0.2000	7471B	PPB	27 Oct 2023 16:16:42	S	Std
2.5 - 1	2.5	17305	-	2.5000	7471B	PPB	27 Oct 2023 16:18:59	S	Std
5.0 - 1	55	35868	-	5.0000	7471B	PPB	27 Oct 2023 16:21:15	S	Std
7.5 - 1	57.5	52272	-	7.5000	7471B	PPB	27 Oct 2023 16:23:31	S	Std
10.0 - 1	510	72688	-	10.0000	7471B	PPB	27 Oct 2023 16:25:48	S	Std
ICV40 - 1	ICV40	28610	4.0197	-	7471B	PPB	27 Oct 2023 16:28:37	U	SMPL
ICB40 - 1	ICB40	-383	-0.0051	-	7471B	PPB	27 Oct 2023 16:30:54	U	SMPL
CCV97 - 1	CCV97	36047	5.0521	-	7471B	PPB	27 Oct 2023 16:33:11	U	SMPL
CCB97 - 1	CCB97	-373	-0.0037	-	7471B	PPB	27 Oct 2023 16:35:26	U	SMPL
CRA - 1	CRA	1241	0.2203	-	7471B	PPB	27 Oct 2023 16:37:42	U	SMPL
HighStd - 1	HighStd	73375	10.2339	-	7471B	PPB	27 Oct 2023 16:39:57	U	SMPL
ChkStd - 1	ChkStd	50323	7.0339	-	7471B	PPB	27 Oct 2023 16:42:12	U	SMPL
PB156702BL - 1	PBS	-406	-0.0083	-	7471B	PPB	27 Oct 2023 16:44:31	U	SMPL
PB156702BS - 1	LCSS	26810	3.7698	-	7471B	PPB	27 Oct 2023 16:46:49	U	SMPL
O5027-01 - 1	214	-225	0.0168	-	7471B	PPB	27 Oct 2023 16:49:05	U	SMPL
O5027-04 - 1	H-DRUMS-SOLID	3532	0.5384	-	7471B	PPB	27 Oct 2023 16:51:23	U	SMPL
O5040-01 - 1	M-4(50FT)(0-5)	35360	4.9567	-	7471B	PPB	27 Oct 2023 16:53:41	U	SMPL
O5048-01 - 1	OILY-STONE	888	0.1713	-	7471B	PPB	27 Oct 2023 16:55:58	U	SMPL
O5051-01 - 1	DUCT-BANK-1A	1427	0.2461	-	7471B	PPB	27 Oct 2023 16:58:15	U	SMPL
CCV98 - 1	CCV98	35375	4.9588	-	7471B	PPB	27 Oct 2023 17:00:30	U	SMPL
CCB98 - 1	CCB98	-103	0.0338	-	7471B	PPB	27 Oct 2023 17:02:45	U	SMPL
O5051-03 - 1	DUCT-BANK-2A	6779	0.9891	-	7471B	PPB	27 Oct 2023 17:05:01	U	SMPL
O5062-01 - 1	NB-07-102523	1225	0.2181	-	7471B	PPB	27 Oct 2023 17:07:17	U	SMPL
O5073-01 - 1	OR-02-102523	2198	0.3532	-	7471B	PPB	27 Oct 2023 17:09:32	U	SMPL
O5078-01 - 1	EO-03-102323	4997	0.7417	-	7471B	PPB	27 Oct 2023 17:11:47	U	SMPL
O5083-01 - 1	SU-04-10252023	2800	0.4367	-	7471B	PPB	27 Oct 2023 17:14:03	U	SMPL
O5083-01DUP - 1	SU-04-10252023DUP	1783	0.2956	-	7471B	PPB	27 Oct 2023 17:16:20	U	SMPL
O5083-01MS - 1	SU-04-10252023MS	27355	3.8455	-	7471B	PPB	27 Oct 2023 17:18:36	U	SMPL
O5083-01MSD - 1	SU-04-10252023MSD	28619	4.0209	-	7471B	PPB	27 Oct 2023 17:20:53	U	SMPL
O5090-01 - 1	WC-1	5161	0.7645	-	7471B	PPB	27 Oct 2023 17:23:10	U	SMPL
CCV99 - 1	CCV99	36709	5.1440	-	7471B	PPB	27 Oct 2023 17:25:25	U	SMPL
CCB99 - 1	CCB99	-53	0.0407	-	7471B	PPB	27 Oct 2023 17:27:41	U	SMPL
O5083-01LX5 - 1		473	0.1137	-	7471B	PPB	27 Oct 2023 17:29:56	U	SMPL
O5083-01A - 1		27788	3.9056	-	7471B	PPB	27 Oct 2023 17:32:11	U	SMPL
CCV01 - 1	CCV01	35154	4.9281	-	7471B	PPB	27 Oct 2023 17:34:27	U	SMPL
CCB01 - 1	CCB01	22	0.0511	-	7471B	PPB	27 Oct 2023 17:36:42	U	SMPL