



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB128013

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.1	10/27/2023	11:15
2	O5107-16		1	12	100.00	>212.0	>212.0	10/27/2023	11:45
3	O5107-16DUP		1	12	100.00	>212.0	>212.0	10/27/2023	12:15
4	O5107-17		1	14	100.00	>212.0	>212.0	10/27/2023	12:45
5	O5107-18		1	14	100.00	>212.0	>212.0	10/27/2023	13:15
6	O5107-19		1	12	100.00	>212.0	>212.0	10/27/2023	13:45
7	O5107-20		1	14	100.00	>212.0	>212.0	10/27/2023	14:15
8	O5107-21		1	13	100.00	>212.0	>212.0	10/27/2023	14:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6128013

WorkList Name : fp 10-27

WorkList ID : 174992

Department : Wet-Chemistry

Date : 10-27-2023 08:32:00

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5107-16	AQ-COMP	Water	Flash Point	Cool 4 deg C	PSEG03	I41	10/26/2023	1010B
O5107-17	1072	Water	Flash Point	Cool 4 deg C	PSEG03	I41	10/26/2023	1010B
O5107-18	1071	Water	Flash Point	Cool 4 deg C	PSEG03	I41	10/26/2023	1010B
O5107-19	1051	Water	Flash Point	Cool 4 deg C	PSEG03	I41	10/26/2023	1010B
O5107-20	1050	Water	Flash Point	Cool 4 deg C	PSEG03	I41	10/26/2023	1010B
O5107-21	1082	Water	Flash Point	Cool 4 deg C	PSEG03	I41	10/26/2023	1010B

Date/Time 10/27/2023 11:00

Raw Sample Received by: RM

Raw Sample Relinquished by: JG

Date/Time 10/27/2023 15:00

Raw Sample Received by: JG

Raw Sample Relinquished by: RM