



Analytical Summary Report

Analysis Method: 9040C
Parameter: Corrosivity
Run Number: LB128015

Analyst By : JIGNESH
Supervisor Review By : Iwona
Slope : 98.2
pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	10/27/2023	15:55
2	CAL2	1	Water	NA	NA	20.6	7.01	10/27/2023	15:56
3	CAL3	1	Water	NA	NA	20.5	10.02	10/27/2023	15:59
4	ICV	1	Water	NA	NA	20.5	6.98	10/27/2023	16:00
5	CCV1	1	Water	NA	NA	20.5	2.01	10/27/2023	16:05
6	O5094-01	1	Water	NA	NA	22.3	6.85	10/27/2023	16:35
7	O5094-01DUP	1	Water	NA	NA	22.4	6.86	10/27/2023	16:36
8	O5106-02	1	Water	NA	NA	23.3	7.27	10/27/2023	16:39
9	O5106-03	1	Water	NA	NA	23.4	8.30	10/27/2023	16:40
10	CCV2	1	Water	NA	NA	20.6	12.02	10/27/2023	16:41

WB-128015

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORRSOVITY O5110 WorkList ID : 175014 Department : Wet-Chemistry Date : 10-27-2023 15:38:24

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5094-01	A 375	Water	Corrosivity	Cool 4 deg C	PSEG03	I21	10/25/2023	9040C
O5106-02	A T3212	Water	Corrosivity	Cool 4 deg C	PSEG03	I41	10/26/2023	9040C
O5106-03	A D3265	Water	Corrosivity	Cool 4 deg C	PSEG03	I41	10/26/2023	9040C

Date/Time

10/27/23

15:40

Raw Sample Received by:

JP (WJC)

JP (WJC)

Raw Sample Relinquished by:

JP (WJC)

JP (WJC)

Date/Time

10/27/23

18:30

Raw Sample Received by:

JP (WJC)

JP (WJC)

Raw Sample Relinquished by:

JP (WJC)

JP (WJC)