



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB128016

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	10/27/2023	15:55
2	CAL2	1	Water	NA	NA	20.6	7.01	10/27/2023	15:56
3	CAL3	1	Water	NA	NA	20.5	10.02	10/27/2023	15:59
4	ICV	1	Water	NA	NA	20.5	6.98	10/27/2023	16:00
5	CCV1	1	Water	NA	NA	20.5	2.01	10/27/2023	16:05
6	O5108-01	1	Solid	20.02	20	26.3	6.23	10/27/2023	16:50
7	O5108-01DUP	1	Solid	20.03	20	26.4	6.24	10/27/2023	16:51
8	O5108-02	1	Solid	20.05	20	26.6	6.13	10/27/2023	17:00
9	O5108-03	1	Solid	20.04	20	26.6	6.00	10/27/2023	17:01
10	CCV2	1	Water	NA	NA	20.6	12.02	10/27/2023	17:05

VB 128016

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH S O5098

WorkList ID : 175013

Department : Wet-Chemistry

Date : 10-27-2023 15:37:20

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5108-01	1023-A	Solid	pH	Cool 4 deg C	PSEG03	I41	10/26/2023	9045D
O5108-02	1023-B	Solid	pH	Cool 4 deg C	PSEG03	I41	10/26/2023	9045D
O5108-03	1023-C	Solid	pH	Cool 4 deg C	PSEG03	I41	10/26/2023	9045D

Date/Time 10/27/23 15:40

Raw Sample Received by: JD (WOC)

Raw Sample Relinquished by: JT (SM)

Date/Time

Raw Sample Received by: JT (SM)

Raw Sample Relinquished by: JD (WOC)