



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB128017

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	10/27/2023	15:55
2	CAL2	1	Water	NA	NA	20.6	7.01	10/27/2023	15:56
3	CAL3	1	Water	NA	NA	20.5	10.02	10/27/2023	15:59
4	ICV	1	Water	NA	NA	20.5	6.98	10/27/2023	16:00
5	CCV1	1	Water	NA	NA	20.5	2.01	10/27/2023	16:05
6	O5107-12	1	Water	NA	NA	25.3	6.18	10/27/2023	17:10
7	O5107-12DUP	1	Water	NA	NA	25.4	6.19	10/27/2023	17:11
8	O5107-13	1	Water	NA	NA	25.2	6.31	10/27/2023	17:15
9	O5107-14	1	Water	NA	NA	25.1	6.45	10/27/2023	17:20
10	O5107-15	1	Water	NA	NA	25.2	6.21	10/27/2023	17:25
11	O5107-17	1	Water	NA	NA	22.7	5.54	10/27/2023	17:30
12	O5107-18	1	Water	NA	NA	23.2	6.07	10/27/2023	17:35
13	O5107-19	1	Water	NA	NA	23.1	6.97	10/27/2023	17:40
14	O5107-20	1	Water	NA	NA	23.6	6.09	10/27/2023	17:45
15	O5107-21	1	Water	NA	NA	25.0	9.55	10/27/2023	17:50
16	CCV2	1	Water	NA	NA	20.6	12.02	10/27/2023	18:00
17	O5108-04	1	Water	NA	NA	25.5	6.60	10/27/2023	18:02
18	O5109-01	1	Water	NA	NA	25.1	7.24	10/27/2023	18:10
19	O5109-02	1	Water	NA	NA	24.9	6.96	10/27/2023	18:15
20	CCV3	1	Water	NA	NA	20.8	2.01	10/27/2023	18:20

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH O5098

WorkList ID : 175012

Department : Wet-Chemistry

Date : 10-27-2023 15:36:22

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5107-12	928	Water	pH	Cool 4 deg C	PSEG03	I41	10/26/2023	9040C
O5107-13	929	Water	pH	Cool 4 deg C	PSEG03	I41	10/26/2023	9040C
O5107-14	930	Water	pH	Cool 4 deg C	PSEG03	I41	10/26/2023	9040C
O5107-15	931	Water	pH	Cool 4 deg C	PSEG03	I41	10/26/2023	9040C
O5107-17	1072	Water	pH	Cool 4 deg C	PSEG03	I41	10/26/2023	9040C
O5107-18	1071	Water	pH	Cool 4 deg C	PSEG03	I41	10/26/2023	9040C
O5107-19	1051	Water	pH	Cool 4 deg C	PSEG03	I41	10/26/2023	9040C
O5107-20	1050	Water	pH	Cool 4 deg C	PSEG03	I41	10/26/2023	9040C
O5107-21	1082	Water	pH	Cool 4 deg C	PSEG03	I41	10/26/2023	9040C
O5108-04	NWB-1965	Water	pH	Cool 4 deg C	PSEG03	I41	10/26/2023	9040C
O5109-01	0818	Water	pH	Cool 4 deg C	PSEG03	I11	10/26/2023	9040C
O5109-02	1025	Water	pH	Cool 4 deg C	PSEG03	I11	10/26/2023	9040C

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

10/27/23 15:40

20 (WOC)

27 (SM)

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

10/27/23

27 (SM)

20 (WOC)

19:00

10/28/23