



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB128024

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	10/28/2023	09:15
2	CAL2	1	Water	NA	NA	20.6	7.00	10/28/2023	09:16
3	CAL3	1	Water	NA	NA	20.6	10.02	10/28/2023	09:20
4	ICV	1	Water	NA	NA	20.6	7.02	10/28/2023	09:25
5	CCV1	1	Water	NA	NA	20.6	2.01	10/28/2023	09:30
6	O5143-03	1	Water	NA	NA	24.4	6.30	10/28/2023	09:35
7	O5143-03DUP	1	Water	NA	NA	24.4	6.30	10/28/2023	09:36
8	O5143-04	1	Water	NA	NA	24.4	6.30	10/28/2023	09:37
9	CCV2	1	Water	NA	NA	20.8	12.02	10/28/2023	09:40

VP 128024

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH 05143

WorkList ID : 175028

Department : Wet-Chemistry

Date : 10-28-2023 09:01:43

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5143-03	COMPOSITE-1	Water	pH	Cool 4 deg C	PSEG03	I41	10/27/2023	9040C
O5143-04	COMPOSITE-1	Water	pH	Cool 4 deg C	PSEG03	I41	10/27/2023	9040C

Date/Time 10/28/23 09:10

Raw Sample Received by: 20 (WC)

Raw Sample Relinquished by: PM CW

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

10/28/23 12:00

PM (WC)

20 (WC)