



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB128025

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	10/28/2023	09:15
2	CAL2	1	Water	NA	NA	20.6	7.00	10/28/2023	09:16
3	CAL3	1	Water	NA	NA	20.6	10.02	10/28/2023	09:20
4	ICV	1	Water	NA	NA	20.6	7.02	10/28/2023	09:25
5	CCV1	1	Water	NA	NA	20.6	2.01	10/28/2023	09:30
6	O5134-01	1	Solid	20.02	20	24.1	5.02	10/28/2023	09:37
7	O5134-03	1	Solid	20.05	20	24.6	5.59	10/28/2023	09:40
8	O5136-01	1	Solid	20.02	20	24.1	4.30	10/28/2023	09:45
9	O5136-01DUP	1	Solid	20.03	20	24.2	4.31	10/28/2023	09:46
10	CCV2	1	Water	NA	NA	20.5	12.02	10/28/2023	09:50

WB128026

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH 05134

WorkList ID : 175029

Department : Wet-Chemistry

Date : 10-28-2023 09:03:23

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5134-01	60194	Solid	pH	Cool 4 deg C	PSEG03	I11	10/27/2023	9045D
O5134-03	60195	Solid	pH	Cool 4 deg C	PSEG03	I11	10/27/2023	9045D
O5136-01	SLUDGE-COMP	Solid	pH	Cool 4 deg C	PSEG03	I41	10/27/2023	9045D

Date/Time 10/28/23 09:10

Raw Sample Received by: JOL (WC1)

Raw Sample Relinquished by: PHA C.W.2

Date/Time 10/28/23

Raw Sample Received by: PHA C.W.2

Raw Sample Relinquished by: PHA C.W.2