



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB128028

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.1	10/28/2023	09:10
2	O5056-01		1	14	100.00	>212.0	>212.0	10/28/2023	09:40
3	O5056-03		1	12	100.00	>212.0	>212.0	10/28/2023	10:10
4	O5098-04		1	8	57.00	134.6	134.3	10/28/2023	10:40
5	O5098-04DUP		1	8	57.00	134.6	134.3	10/28/2023	11:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6128028

WorkList Name : fp -10-26

WorkList ID : 174957

Department : Wet-Chemistry

Date : 10-26-2023 08:17:23

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5056-01	TOTE	Solid	Flash Point	Cool 4 deg C	PSEG03	I51	10/25/2023	1010B
O5056-03	BUCKET-COMP	Solid	Flash Point	Cool 4 deg C	PSEG03	I51	10/25/2023	1010B
O5098-04	00092	Solid	Flash Point	Cool 4 deg C	PSEG03	I41	10/26/2023	1010B

Date/Time 10/28/2023 09:00
 Raw Sample Received by: RM (ave)
 Raw Sample Relinquished by: J D Ave

Date/Time 10/28/2023 11:15
 Raw Sample Received by: J D (ave)
 Raw Sample Relinquished by: RM (ave)