



PERCENT SOLID

Supervisor: Iwona
Analyst: JIGNESH
Date: 10/30/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 14:15
In Date: 10/29/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 10/30/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB128038

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O5122-01	EZMN7	1	1.18	8.43	9.61	8.88	91.3	
O5122-02	EZMN8	2	1.18	8.50	9.68	8.72	88.7	
O5122-03	EZMN9	3	1.15	8.40	9.55	8.59	88.6	
O5122-04	EZMP0	4	1.15	8.39	9.54	8.4	86.4	
O5122-05	EZMP1	5	1.18	8.58	9.76	8.81	88.9	
O5122-06	EZMP2	6	1.12	8.76	9.88	8.97	89.6	
O5122-07	EZMP3	7	1.18	8.56	9.74	8.53	85.9	
O5122-08	EZMP4	8	1.15	8.41	9.56	8.62	88.8	
O5122-09	EZMP5	9	1.19	8.40	9.59	8.8	90.6	
O5122-10	EZMP6	10	1.18	8.54	9.72	8.94	90.9	
O5122-11	EZMS3	11	1.12	8.77	9.89	8.39	82.9	
O5122-12	EZMS4	12	1.18	8.59	9.77	9.16	92.9	
O5122-13	EZMS5	13	1.16	8.61	9.77	8.91	90.0	
O5122-14	EZMS6	14	1.19	8.41	9.6	8.81	90.6	
O5122-15	EZMS7	15	1.14	8.59	9.73	8.97	91.2	
O5122-16	EZMS7D	16	1.14	8.59	9.73	8.97	91.2	
O5122-17	EZMS7S	17	1.14	8.59	9.73	8.97	91.2	
O5122-18	EZMR0	18	1.18	8.34	9.52	8.72	90.4	
O5122-19	EZMR1	19	1.1	8.78	9.88	8.77	87.4	
O5122-20	EZMR2	20	1.19	8.52	9.71	8.65	87.6	
O5122-21	EZMR3	21	1.15	8.82	9.97	8.78	86.5	
O5122-22	EZMR4	22	1.19	8.41	9.6	8.66	88.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$