



PERCENT SOLID

Supervisor: Iwona  
Analyst: JIGNESH  
Date: 10/30/2023

OVENTEMP IN Celsius(°C): 107  
Time IN: 13:50  
In Date: 10/29/2023  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:15  
Out Date: 10/30/2023  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB128040

| Lab ID   | Client SampleID | Dish # | Dish Wt(g) (A) | Sample Wt(g) | Dish + Sample Wt(g) (B) | Dish+Dry Sample Wt(g) (C) | % Solid | Comments   |
|----------|-----------------|--------|----------------|--------------|-------------------------|---------------------------|---------|------------|
| O5060-01 | DCKL0           | 1      | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | P.E.SAMPLE |
| O5060-02 | DCKG0           | 2      | 1.15           | 8.54         | 9.69                    | 9.31                      | 95.6    |            |
| O5060-03 | DCKG1           | 3      | 1.18           | 8.56         | 9.74                    | 9.00                      | 91.4    |            |
| O5060-04 | DCKG2           | 4      | 1.19           | 8.48         | 9.67                    | 8.58                      | 87.1    |            |
| O5060-05 | DCKG3           | 5      | 1.16           | 8.42         | 9.58                    | 8.68                      | 89.3    |            |
| O5060-06 | DCKG4           | 6      | 1.14           | 8.69         | 9.83                    | 9.09                      | 91.5    |            |
| O5060-07 | DCKG5           | 7      | 1.19           | 8.47         | 9.66                    | 8.46                      | 85.8    |            |
| O5060-08 | DCKG6           | 8      | 1.15           | 8.50         | 9.65                    | 8.69                      | 88.7    |            |
| O5060-09 | DCKG7           | 9      | 1.16           | 8.51         | 9.67                    | 7.93                      | 79.6    |            |
| O5060-10 | DCKG8           | 10     | 1.12           | 8.77         | 9.89                    | 8.5                       | 84.2    |            |
| O5060-11 | DCKG9           | 11     | 1.15           | 8.40         | 9.55                    | 9.18                      | 95.6    |            |
| O5060-12 | DCKH0           | 12     | 1.15           | 8.40         | 9.55                    | 8.57                      | 88.3    |            |
| O5060-13 | DCKH1           | 13     | 1.12           | 8.76         | 9.88                    | 8.59                      | 85.3    |            |
| O5060-14 | DCKH2           | 14     | 1.18           | 8.42         | 9.6                     | 8.68                      | 89.1    |            |
| O5060-15 | DCKH3           | 15     | 1.19           | 8.61         | 9.8                     | 9.56                      | 97.2    |            |
| O5060-16 | DCKH3MS         | 16     | 1.19           | 8.61         | 9.8                     | 9.56                      | 97.2    |            |
| O5060-17 | DCKH3MSD        | 17     | 1.19           | 8.61         | 9.8                     | 9.56                      | 97.2    |            |
| O5060-18 | DCKH4           | 18     | 1.19           | 8.33         | 9.52                    | 8.4                       | 86.6    |            |
| O5060-19 | DCKH5           | 19     | 1.13           | 8.69         | 9.82                    | 8.55                      | 85.4    |            |
| O5060-20 | DCKH8           | 20     | 1.16           | 8.63         | 9.79                    | 9.6                       | 97.8    |            |
| O5060-21 | DCKH9           | 21     | 1.15           | 8.40         | 9.55                    | 8.33                      | 85.5    |            |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$