



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB128053

Ambient Barometric Pressure (mmHg): 750.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	83.0	10/30/2023	09:10
2	O5094-01		1	12	100.00	>212.0	>212.0	10/30/2023	09:40
3	O5094-01DUP		1	12	100.00	>212.0	>212.0	10/30/2023	10:10
4	O5106-02		1	14	100.00	>212.0	>212.0	10/30/2023	10:40
5	O5106-03		1	13	100.00	>212.0	>212.0	10/30/2023	11:10
6	O5108-04		1	12	100.00	>212.0	>212.0	10/30/2023	11:40
7	O5109-01		1	14	100.00	>212.0	>212.0	10/30/2023	12:10
8	O5109-02		1	12	100.00	>212.0	>212.0	10/30/2023	12:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

16128053

WorkList Name : fp-10-30

WorkList ID : 175062

Department : Wet-Chemistry

Date : 10-30-2023 08:19:29

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5094-01	375	Water	Flash Point	Cool 4 deg C	PSEG03	I21	10/25/2023	1010B
O5106-02	T3212	Water	Flash Point	Cool 4 deg C	PSEG03	I41	10/26/2023	1010B
O5106-03	D3265	Water	Flash Point	Cool 4 deg C	PSEG03	I41	10/26/2023	1010B
O5108-04	NWB-1965	Water	Flash Point	Cool 4 deg C	PSEG03	I41	10/26/2023	1010B
O5109-01	0818	Water	Flash Point	Cool 4 deg C	PSEG03	I11	10/26/2023	1010B
O5109-02	1025	Water	Flash Point	Cool 4 deg C	PSEG03	I11	10/26/2023	1010B

Date/Time 10/30/2023 08:45
 Raw Sample Received by: Rrr cwc
 Raw Sample Relinquished by: 10/10C

Date/Time 10/30/2023 12:45
 Raw Sample Received by: 10/10C
 Raw Sample Relinquished by: Rrr cwc