

LB128067


245.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

42854



A= 0.0000e+000

B= 2.3056e-004

C= -1.2825e-002

Rho= 0.9993376

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.008	-0.008	19	0.000	19					1
0.05	0.050	0.045	-0.005	252	0.0 %	252	0				10
0.2	0.200	0.200	-0.000	923	0.0 %	923					0
2.5	2.500	2.335	-0.165	10185	0.0 %	10185					17
5.0	5.000	5.287	0.287	22988	0.0 %	22988					6
7.5	7.500	7.523	0.023	32683	0.0 %	32683					0
10.0	10.000	9.868	-0.132	42854	0.0 %	42854					10

LB128067 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 31 Oct 2023 10:13:17

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	19	-	0.0000	245.1	PPB	31 Oct 2023 10:23:15	S	Std
0.05 - 1	50.05	252	-	0.0500	245.1	PPB	31 Oct 2023 10:25:32	S	Std
0.2 - 1	50.2	923	-	0.2000	245.1	PPB	31 Oct 2023 10:27:50	S	Std
2.5 - 1	52.5	10185	-	2.5000	245.1	PPB	31 Oct 2023 10:30:07	S	Std
5.0 - 1	55	22988	-	5.0000	245.1	PPB	31 Oct 2023 10:32:23	S	Std
7.5 - 1	57.5	32683	-	7.5000	245.1	PPB	31 Oct 2023 10:37:14	S	Std
10.0 - 1	510	42854	-	10.0000	245.1	PPB	31 Oct 2023 10:42:09	S	Std
ICV45 - 1	ICV45	18132	4.1677	-245.1		PPB	31 Oct 2023 10:45:03	U	SMPL
ICB45 - 1	ICB45	-136	-0.0442	-245.1		PPB	31 Oct 2023 10:49:57	U	SMPL
CCV14 - 1	CCV14	22771	5.2373	-245.1		PPB	31 Oct 2023 10:52:12	U	SMPL
CCB14 - 1	CCB14	-33	-0.0204	-245.1		PPB	31 Oct 2023 10:57:03	U	SMPL
CRA - 1	CRA	219	0.0377	-245.1		PPB	31 Oct 2023 11:04:28	U	SMPL
HighStd - 1	HighStd	41582	9.5744	-245.1		PPB	31 Oct 2023 11:06:42	U	SMPL
ChkStd - 1	ChkStd	29774	6.8520	-245.1		PPB	31 Oct 2023 11:08:58	U	SMPL
PB156786BL - 1	PBW	-80	-0.0313	-245.1		PPB	31 Oct 2023 11:13:53	U	SMPL
PB156786BS - 1	LCSW	17508	4.0239	-245.1		PPB	31 Oct 2023 11:20:25	U	SMPL
O4934-07 - 1	14B-(1-4)-COMP	3711	0.8428	-245.1		PPB	31 Oct 2023 11:28:38	U	SMPL
O4934-07DUP - 1	14B-(1-4)-COMPDUP	3252	0.7370	-245.1		PPB	31 Oct 2023 11:30:54	U	SMPL
O4934-07MS - 1	14B-(1-4)-COMPMS	22813	5.2470	-245.1		PPB	31 Oct 2023 11:33:11	U	SMPL
O4934-07MSD - 1	14B-(1-4)-COMPMSD	20949	4.8172	-245.1		PPB	31 Oct 2023 11:35:28	U	SMPL
O4943-07 - 1	14B-(1-4)-COMP	1956	0.4382	-245.1		PPB	31 Oct 2023 11:37:44	U	SMPL
CCV15 - 1	CCV15	21274	4.8922	-245.1		PPB	31 Oct 2023 11:42:48	U	SMPL
CCB15 - 1	CCB15	-88	-0.0331	-245.1		PPB	31 Oct 2023 11:45:04	U	SMPL
O5033-07 - 1	14B-(1-4)-COMP	3450	0.7826	-245.1		PPB	31 Oct 2023 11:47:19	U	SMPL
O5054-07 - 1	14B-(1-4)-COMP	2750	0.6212	-245.1		PPB	31 Oct 2023 11:49:34	U	SMPL
O4934-07LX5 - 1		710	0.1509	-245.1		PPB	31 Oct 2023 11:51:50	U	SMPL
O4934-07A - 1		22147	5.0934	-245.1		PPB	31 Oct 2023 11:54:06	U	SMPL
CCV16 - 1	CCV16	22955	5.2797	-245.1		PPB	31 Oct 2023 11:59:05	U	SMPL
CCB16 - 1	CCB16	-112	-0.0386	-245.1		PPB	31 Oct 2023 12:01:21	U	SMPL