

LB128071


7470A

INSTRUMENT MD. CV2

Linear

 μ Abs.:

61219



A= 0.0000e+000

B= 1.6160e-004

C= 9.5160e-002

Rho= 0.9998187

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%RSD
0.0	0.000	0.096	0.096	8	0.000	8	0				1
0.2	0.200	0.189	-0.011	583	0.0 %	583					6
2.5	2.500	2.397	-0.103	14243	0.0 %	14243					4
5.0	5.000	4.948	-0.052	30030	0.0 %	30030					1
7.5	7.500	7.582	0.082	46329	0.0 %	46329					1
10.0	10.000	9.988	-0.012	61219	0.0 %	61219					0

LB128071 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 31 Oct 2023 12:48:01

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	583	-	0.000	7470A	PPB	31 Oct 2023 13:45:30	S	Std
0.2 - 1	51.2	14243	-	0.200	7470A	PPB	31 Oct 2023 13:47:47	S	Std
2.5 - 1	52.5	30030	-	2.500	7470A	PPB	31 Oct 2023 13:50:05	S	Std
5.0 - 1	55	46329	-	5.000	7470A	PPB	31 Oct 2023 13:52:22	S	Std
7.5 - 1	57.5	61219	-	7.500	7470A	PPB	31 Oct 2023 13:54:38	S	Std
10.0 - 1	510	22407	3.7160	10.000	7470A	PPB	31 Oct 2023 13:59:43	S	Std
ICV46 - 1	ICV46	-581	0.0013	-	7470A	PPB	31 Oct 2023 14:06:04	U	SMPL
ICB46 - 1	ICB46	32603	5.3636	-	7470A	PPB	31 Oct 2023 14:10:20	U	SMPL
CCV17 - 1	CCV17	-503	0.0139	-	7470A	PPB	31 Oct 2023 14:12:37	U	SMPL
CCB17 - 1	CCB17	739	0.2146	-	7470A	PPB	31 Oct 2023 14:14:52	U	SMPL
CRA - 1	CRA	60877	9.9326	-	7470A	PPB	31 Oct 2023 14:19:42	U	SMPL
HighStd - 1	HighStd	41822	6.8534	-	7470A	PPB	31 Oct 2023 14:21:57	U	SMPL
ChkStd - 1	ChkStd	-731	-0.0230	-	7470A	PPB	31 Oct 2023 14:24:12	U	SMPL
PB156772BL - 1	PBW	20010	3.3287	-	7470A	PPB	31 Oct 2023 14:26:31	U	SMPL
PB156772BS - 1	LCSW	9456	1.6230	-	7470A	PPB	31 Oct 2023 14:28:48	U	SMPL
O5077-02 - 1	LP-1	-302	0.0464	-	7470A	PPB	31 Oct 2023 14:31:05	U	SMPL
O5077-04 - 1	LP-2	-166	0.0683	-	7470A	PPB	31 Oct 2023 14:33:23	U	SMPL
O5077-06 - 1	LP-3	-118	0.0761	-	7470A	PPB	31 Oct 2023 14:35:40	U	SMPL
O5077-08 - 1	SOL-1	-107	0.0779	-	7470A	PPB	31 Oct 2023 14:37:57	U	SMPL
O5077-10 - 1	SOL-2	32435	5.3365	-	7470A	PPB	31 Oct 2023 14:40:12	U	SMPL
CCV18 - 1	CCV18	-427	0.0262	-	7470A	PPB	31 Oct 2023 14:42:27	U	SMPL
CCB18 - 1	CCB18	-69	0.0840	-	7470A	PPB	31 Oct 2023 14:44:43	U	SMPL
O5098-01 - 1	COMP-1	-260	0.0531	-	7470A	PPB	31 Oct 2023 14:47:00	U	SMPL
O5098-02 - 1	000126	-93	0.0801	-	7470A	PPB	31 Oct 2023 14:49:14	U	SMPL
O5098-03 - 1	00094	143	0.1183	-	7470A	PPB	31 Oct 2023 14:51:29	U	SMPL
O5098-04 - 1	00092	85	0.1089	-	7470A	PPB	31 Oct 2023 14:53:45	U	SMPL
O5106-01 - 1	D3291	2859	0.5572	-	7470A	PPB	31 Oct 2023 14:56:01	U	SMPL
O5110-01 - 1	2943	-155	0.0701	-	7470A	PPB	31 Oct 2023 14:58:17	U	SMPL
O5124-01 - 1	LP-1	-8	0.0939	-	7470A	PPB	31 Oct 2023 15:00:34	U	SMPL
O5124-02 - 1	LP-2	11	0.0969	-	7470A	PPB	31 Oct 2023 15:02:52	U	SMPL
O5124-03 - 1	LP-3	-2	0.0948	-	7470A	PPB	31 Oct 2023 15:05:09	U	SMPL
O5124-04 - 1	LP-4	31470	5.1806	-	7470A	PPB	31 Oct 2023 15:07:26	U	SMPL
CCV19 - 1	CCV19	-248	0.0551	-	7470A	PPB	31 Oct 2023 15:09:42	U	SMPL
CCB19 - 1	CCB19	2012	0.4203	-	7470A	PPB	31 Oct 2023 15:11:57	U	SMPL
O5129-01 - 1	BLDG-39-2ND-FLOOR-WINDOWS-102623JE14-TCLP	2096	0.4342	-	7470A	PPB	31 Oct 2023 15:14:13	U	SMPL
O5129-01DUP - 1	BLDG-39-2ND-FLOOR-WINDOWS-102623JE14-TCLPDUP	25026	4.1392	-	7470A	PPB	31 Oct 2023 15:16:28	U	SMPL
O5129-01MS - 1	BLDG-39-2ND-FLOOR-WINDOWS-102623JE14-TCLPMS	26577	4.3899	-	7470A	PPB	31 Oct 2023 15:18:44	U	SMPL
O5129-01MSD - 1	BLDG-39-2ND-FLOOR-WINDOWS-102623JE14-TCLPMSD	-358	0.0373	-	7470A	PPB	31 Oct 2023 15:21:00	U	SMPL
PB156773BL - 1	PBW	22845	3.7868	-	7470A	PPB	31 Oct 2023 15:23:17	U	SMPL
PB156773BS - 1	LCSW	-313	0.0446	-	7470A	PPB	31 Oct 2023 15:25:34	U	SMPL
O5094-01 - 1	375	-129	0.0743	-	7470A	PPB	31 Oct 2023 15:27:50	U	SMPL
O5094-01DUP - 1	375DUP	22206	3.6835	-	7470A	PPB	31 Oct 2023 15:30:07	U	SMPL
O5094-01MS - 1	375MS	25311	4.1853	-	7470A	PPB	31 Oct 2023 15:32:25	U	SMPL
O5094-01MSD - 1	375MSD	29157	4.8068	-	7470A	PPB	31 Oct 2023 15:34:41	U	SMPL
CCV20 - 1	CCV20	-94	0.0800	-	7470A	PPB	31 Oct 2023 15:36:58	U	SMPL
CCB20 - 1	CCB20	-168	0.0680	-	7470A	PPB	31 Oct 2023 15:39:14	U	SMPL
O5106-03 - 1	D3265	-92	0.0803	-	7470A	PPB	31 Oct 2023 15:41:31	U	SMPL
PB156774BL - 1	PBW	24932	4.1241	-	7470A	PPB	31 Oct 2023 15:43:47	U	SMPL
PB156774BS - 1	LCSW	-293	0.0478	-	7470A	PPB	31 Oct 2023 15:46:02	U	SMPL
O5136-02 - 1	SAND-BLAST-COMP	-109	0.0775	-	7470A	PPB	31 Oct 2023 15:48:18	U	SMPL
O5138-03 - 1	348	-76	0.0829	-	7470A	PPB	31 Oct 2023 15:50:35	U	SMPL
O5142-02 - 1	PSC-124037	-41	0.0885	-	7470A	PPB	31 Oct 2023 15:52:52	U	SMPL
O5142-02DUP - 1	PSC-124037DUP	20216	3.3620	-	7470A	PPB	31 Oct 2023 15:55:09	U	SMPL
O5142-02MS - 1	PSC-124037MS	24230	4.0106	-	7470A	PPB	31 Oct 2023 15:57:26	U	SMPL
O5142-02MSD - 1	PSC-124037MSD	-309	0.0452	-	7470A	PPB	31 Oct 2023 15:59:43	U	SMPL
PB156799BL - 1	PBW	29704	4.8952	-	7470A	PPB	31 Oct 2023 16:01:59	U	SMPL
CCV21 - 1	CCV21	-76	0.0829	-	7470A	PPB	31 Oct 2023 16:04:16	U	SMPL
CCB21 - 1	CCB21	21315	3.5396	-	7470A	PPB	31 Oct 2023 16:06:31	U	SMPL
PB156799BS - 1	LCSW	-283	0.0494	-	7470A	PPB	31 Oct 2023 16:08:48	U	SMPL
O5040-03 - 1	EB-2	-126	0.0748	-	7470A	PPB	31 Oct 2023 16:11:04	U	SMPL
O5040-04 - 1	FB-3	41870	6.8611	-	7470A	PPB	31 Oct 2023 16:13:19	U	SMPL
O5107-16 - 1	AQ-COMP	35442	5.8224	-	7470A	PPB	31 Oct 2023 16:15:35	U	SMPL
O5107-16DUP - 1	AQ-COMPDUP	51957	8.4912	-	7470A	PPB	31 Oct 2023 16:17:51	U	SMPL
O5107-16MS - 1	AQ-COMPMS	56263	9.1870	-	7470A	PPB	31 Oct 2023 16:20:13	U	SMPL
O5107-16MSD - 1	AQ-COMPMSD	10685	1.8216	-	7470A	PPB	31 Oct 2023 16:22:31	U	SMPL
O5143-03 - 1	COMPOSITE-1	7425	1.2957	-	7470A	PPB	31 Oct 2023 16:24:50	U	SMPL
O5143-04 - 1	COMPOSITE-1	-124	0.0751	-	7470A	PPB	31 Oct 2023 16:27:08	U	SMPL
PB156675TB - 1	PB156675TB	29886	4.9246	-	7470A	PPB	31 Oct 2023 16:29:26	U	SMPL
CCV22 - 1	CCV22	-33	0.0898	-	7470A	PPB	31 Oct 2023 16:31:43	U	SMPL
CCB22 - 1	CCB22	-34	0.0897	-	7470A	PPB	31 Oct 2023 16:33:59	U	SMPL
PB156714TB - 1	PB156714TB	-16	0.0926	-	7470A	PPB	31 Oct 2023 16:36:15	U	SMPL
PB156737TB - 1	PB156737TB	375	0.1558	-	7470A	PPB	31 Oct 2023 16:38:31	U	SMPL
O5129-01LX5 - 1	CCV23	23950	3.9654	-	7470A	PPB	31 Oct 2023 16:40:48	U	SMPL
O5129-01A - 1	CCB23	-65	0.0847	-	7470A	PPB	31 Oct 2023 16:43:03	U	SMPL
O5094-01LX5 - 1		21898	3.6338	-	7470A	PPB	31 Oct 2023 16:45:19	U	SMPL
O5094-01A - 1		-82	0.0819	-	7470A	PPB	31 Oct 2023 16:47:35	U	SMPL
O5142-02LX5 - 1		22853	3.7881	-	7470A	PPB	31 Oct 2023 16:49:51	U	SMPL
O5142-02A - 1		9655	1.6554	-	7470A	PPB	31 Oct 2023 16:52:08	U	SMPL
O5107-16LX5 - 1		54612	8.9202	-	7470A	PPB	31 Oct 2023 16:54:25	U	SMPL
O5107-16A - 1		31920	5.2533	-	7470A	PPB	31 Oct 2023 16:56:41	U	SMPL
CCV23 - 1	CCV23	-113	0.0768	-	7470A	PPB	31 Oct 2023 16:58:58	U	SMPL
CCB23 - 1	CCB23				7470A	PPB	31 Oct 2023 17:01:17	U	SMPL