



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB128073

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	10/31/2023	11:15
2	O5134-01		1	12	100.00	>212.0	>212.0	10/31/2023	11:45
3	O5134-01DUP		1	12	100.00	>212.0	>212.0	10/31/2023	12:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6128073

WorkList Name : fp s-10-31

WorkList ID : 175102

Department : Wet-Chemistry

Date : 10-31-2023 08:15:37

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5134-013	60194	Solid	Flash Point	Cool 4 deg C	PSEG03	I11	10/27/2023	1010B

Date/Time 10/31/2023 11:00
 Raw Sample Received by: Dr
 Raw Sample Relinquished by: gpc

Date/Time 10/31/2023 12:20
 Raw Sample Received by: gpc
 Raw Sample Relinquished by: gpc