



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB128074

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	10/31/2023	12:50
2	O5143-03		1	14	100.00	>212.0	>212.0	10/31/2023	13:20
3	O5143-03DUP		1	14	100.00	>212.0	>212.0	10/31/2023	13:50
4	O5143-04		1	12	100.00	>212.0	>212.0	10/31/2023	14:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

16128074

WorkList Name : fp w 10-31 WorkList ID : 175101 Department : Wet-Chemistry Date : 10-31-2023 08:15:26

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5143-03 U	COMPOSITE-1	Water	Flash Point	Cool 4 deg C	PSEG03	I41	10/27/2023	1010B
O5143-04 D	COMPOSITE-1	Water	Flash Point	Cool 4 deg C	PSEG03	I41	10/27/2023	1010B

Date/Time 10/31/2023 12:45
 Raw Sample Received by: RM awc
 Raw Sample Relinquished by: 7076001

Date/Time 10/31/2023 14:50
 Raw Sample Received by: 7076001
 Raw Sample Relinquished by: 7076001