



PERCENT SOLID

Supervisor: Iwona
Analyst: JIGNESH
Date: 11/1/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 14:55
In Date: 10/31/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:44
Out Date: 11/01/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB128076

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O5151-01	EZMT1	1	1.18	8.69	9.87	9.22	92.5	
O5151-02	EZMT2	2	1.16	8.50	9.66	9.03	92.6	
O5151-03	EZMT3	3	1.19	8.39	9.58	8.91	92.0	
O5151-04	EZMT4	4	1.15	8.38	9.53	8.95	93.1	
O5151-05	EZMT5	5	1.18	8.54	9.72	9.1	92.7	
O5151-06	EZMT6	6	1.18	8.40	9.58	8.85	91.3	
O5151-07	EZMT7	7	1.18	8.40	9.58	8.97	92.7	
O5151-08	EZMT8	8	1.13	8.46	9.59	9.00	93.0	
O5151-09	EZMT9	9	1.16	8.40	9.56	9.06	94.0	
O5151-10	EZMW0	10	1.16	8.77	9.93	8.92	88.5	
O5151-11	EZMW1	11	1.19	8.45	9.64	8.63	88.0	
O5151-12	EZMW2	12	1.18	8.72	9.9	8.93	88.9	
O5151-13	EZMW3	13	1.17	8.46	9.63	8.76	89.7	
O5151-14	EZMW4	14	1.17	8.50	9.67	8.74	89.1	
O5151-15	EZMW5	15	1.19	8.62	9.81	8.82	88.5	
O5151-16	EZMW6	16	1.19	8.40	9.59	8.52	87.3	
O5151-17	EZMW7	17	1.19	8.52	9.71	8.89	90.4	
O5151-18	EZMW8	18	1.19	8.61	9.8	8.85	89.0	
O5151-19	EZMW9	19	1.18	8.54	9.72	8.75	88.6	
O5151-20	EZMX0	20	1.19	8.52	9.71	8.64	87.4	
O5151-21	EZMX0D	21	1.19	8.52	9.71	8.64	87.4	
O5151-22	EZMX0S	22	1.19	8.52	9.71	8.64	87.4	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$