



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB128090

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	10/31/2023	09:10
2	CAL2	1	Water	NA	NA	20.3	7.00	10/31/2023	09:11
3	CAL3	1	Water	NA	NA	20.1	10.02	10/31/2023	09:15
4	ICV	1	Water	NA	NA	20.2	7.02	10/31/2023	09:20
5	CCV1	1	Water	NA	NA	20.3	2.01	10/31/2023	09:25
6	O5047-02	1	Solid	20.02	20	22.6	8.76	10/31/2023	09:35
7	O5164-02	1	Solid	20.01	20	22.1	7.00	10/31/2023	09:40
8	O5164-02DUP	1	Solid	20.02	20	22.2	7.01	10/31/2023	09:41
9	CCV2	1	Water	NA	NA	20.3	12.02	10/31/2023	09:45

128090

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORRROVITY O5156

WorkList ID : 175115

Department : Wet-Chemistry

Date : 10-31-2023 08:36:38

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5047-02	PT-CORR-SOIL	Solid	Corrosivity	Cool 4 deg C	CHEM02	QA OI	10/23/2023	9045D
O5164-02	ETGI-352	Solid	Corrosivity	Cool 4 deg C	PSEG03	I31	10/30/2023	9045D

Date/Time 10/31/23 09:00
 Raw Sample Received by: J. C. Carr
 Raw Sample Relinquished by: J. C. Carr

Date/Time 10/31/23 12:00
 Raw Sample Received by: J. C. Carr
 Raw Sample Relinquished by: J. C. Carr