

LB128096


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

46586



A= 0.0000e+000

B= 2.1243e-004 slope

C= 3.9771e-003 $\rightarrow$ intercept

Rho= 0.9997135

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0)
0.00	0.000	0.007	0.007	16	0.000	16					=
0.05	0.050					419	0				18
0.20	0.200	0.235	0.035	1088	0.0 %	1088					-5
2.50	2.500	2.385	-0.115	11210	0.0 %	11210					0
5.00	5.000	5.024	0.024	23630	0.0 %	23630					2
7.50	7.500	7.648	0.148	35983	0.0 %	35983					-1
10.0	10.000	9.900	-0.100	46586	0.0 %	46586					

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	16	-		0 SFAM01.1	PPB	11/1/2023 13:37	Std
	0.2 S01	1088	-		0.2 SFAM01.1	PPB	11/1/2023 13:40	Std
	2.5 S02	11210	-		2.5 SFAM01.1	PPB	11/1/2023 13:42	Std
	5 S03	23630	-		5 SFAM01.1	PPB	11/1/2023 13:52	Std
	7.5 S04	35983	-		7.5 SFAM01.1	PPB	11/1/2023 13:57	Std
	10 S05	46586	-		10 SFAM01.1	PPB	11/1/2023 14:04	Std
ICV047	ICV047	18882	4.0151 -		SFAM01.1	PPB	11/1/2023 14:10	SMPL
ICB047	ICB047	-199	-0.0383 -		SFAM01.1	PPB	11/1/2023 14:12	SMPL
CCV024	CCV024	23451	4.9858 -		SFAM01.1	PPB	11/1/2023 14:14	SMPL
CCB024	CCB024	-210	-0.0406 -		SFAM01.1	PPB	11/1/2023 14:17	SMPL
PB156812BL	PBS	59	0.0165 -		SFAM01.1	PPB	11/1/2023 14:19	SMPL
O4968-01	YC612	28281	6.0118 -		SFAM01.1	PPB	11/1/2023 14:21	SMPL
O4968-02	YC613	49173	10.45 -		SFAM01.1	PPB	11/1/2023 14:23	SMPL
O4968-03	YC614	41237	8.7641 -		SFAM01.1	PPB	11/1/2023 14:26	SMPL
O4968-04	YC614D	38042	8.0854 -		SFAM01.1	PPB	11/1/2023 14:28	SMPL
O4968-05	YC614S	46973	9.9826 -		SFAM01.1	PPB	11/1/2023 14:33	SMPL
O4968-06	YC615	38196	8.1181 -		SFAM01.1	PPB	11/1/2023 14:36	SMPL
O4968-07	YC616	18236	3.8779 -		SFAM01.1	PPB	11/1/2023 14:38	SMPL
O4968-08	YC617	26179	5.5653 -		SFAM01.1	PPB	11/1/2023 14:40	SMPL
O4968-09	YC618	9557	2.0342 -		SFAM01.1	PPB	11/1/2023 14:43	SMPL
O4968-10	YC619	7524	1.6023 -		SFAM01.1	PPB	11/1/2023 14:45	SMPL
O4968-11	YC6E0	3105	0.6636 -		SFAM01.1	PPB	11/1/2023 14:47	SMPL
O4968-12	YC6K6	7578	1.6138 -		SFAM01.1	PPB	11/1/2023 14:49	SMPL
O4968-13	YC686	3426	0.7318 -		SFAM01.1	PPB	11/1/2023 14:52	SMPL
O4968-14	YC687	3522	0.7522 -		SFAM01.1	PPB	11/1/2023 14:54	SMPL
O4968-15	YC688	3121	0.667 -		SFAM01.1	PPB	11/1/2023 14:56	SMPL
O4968-16	YC689	5466	1.1651 -		SFAM01.1	PPB	11/1/2023 14:58	SMPL
O4968-17	YC690	4984	1.0627 -		SFAM01.1	PPB	11/1/2023 15:01	SMPL
O4968-18	YC6K4	6063	1.292 -		SFAM01.1	PPB	11/1/2023 15:03	SMPL
O4968-19	YC6K5	4990	1.064 -		SFAM01.1	PPB	11/1/2023 15:05	SMPL
CCV025	CCV025	23916	5.0845 -		SFAM01.1	PPB	11/1/2023 15:07	SMPL
CCB025	CCB025	-62	-0.0092 -		SFAM01.1	PPB	11/1/2023 15:10	SMPL
O4968-20	YC6K7	1772	0.3804 -		SFAM01.1	PPB	11/1/2023 15:12	SMPL
O4968-21	YC6K8	2740	0.586 -		SFAM01.1	PPB	11/1/2023 15:14	SMPL
O4968-22	YC6K9	4684	0.999 -		SFAM01.1	PPB	11/1/2023 15:16	SMPL
PB156813BL	PBS813	86	0.0222 -		SFAM01.1	PPB	11/1/2023 15:19	SMPL
O4969-01	YC646	52348	11.1244 -		SFAM01.1	PPB	11/1/2023 15:21	SMPL
O4969-02	YC647	18693	3.975 -		SFAM01.1	PPB	11/1/2023 15:23	SMPL
O4969-03	YC647D	17478	3.7169 -		SFAM01.1	PPB	11/1/2023 15:26	SMPL
O4969-04	YC647S	31909	6.7825 -		SFAM01.1	PPB	11/1/2023 15:28	SMPL
O4969-05	YC661	50111	10.6492 -		SFAM01.1	PPB	11/1/2023 15:30	SMPL
O4969-06	YC662	30959	6.5807 -		SFAM01.1	PPB	11/1/2023 15:32	SMPL
O4969-07	YC663	22055	4.6892 -		SFAM01.1	PPB	11/1/2023 15:35	SMPL
O4969-08	YC6E1	66647	14.162 -		SFAM01.1	PPB	11/1/2023 15:37	SMPL
O4969-09	YC6E2	46445	9.8705 -		SFAM01.1	PPB	11/1/2023 15:39	SMPL
O4969-10	YC6E3	42991	9.1367 -		SFAM01.1	PPB	11/1/2023 15:41	SMPL

O4969-11	YC6E4	44849	9.5314 -	SFAM01.1	PPB	11/1/2023 15:44 SMPL
O4969-12	YC6E7	24928	5.2995 -	SFAM01.1	PPB	11/1/2023 15:46 SMPL
O4969-13	YC692	73195	15.5531 -	SFAM01.1	PPB	11/1/2023 15:48 SMPL
O4969-14	YC693	46926	9.9726 -	SFAM01.1	PPB	11/1/2023 15:51 SMPL
O4969-15	YC694	76346	16.2224 -	SFAM01.1	PPB	11/1/2023 15:53 SMPL
O4969-16	YC695	91587	19.4601 -	SFAM01.1	PPB	11/1/2023 15:55 SMPL
CCV026	CCV026	23541	5.0049 -	SFAM01.1	PPB	11/1/2023 15:57 SMPL
CCB026	CCB026	-144	-0.0266 -	SFAM01.1	PPB	11/1/2023 16:00 SMPL
O4969-17	YC696	67163	14.2717 -	SFAM01.1	PPB	11/1/2023 16:02 SMPL
O4969-18	YC697	21611	4.5949 -	SFAM01.1	PPB	11/1/2023 16:04 SMPL
O4969-19	YC698	39681	8.4336 -	SFAM01.1	PPB	11/1/2023 16:07 SMPL
O4969-20	YC699	102193	21.7132 -	SFAM01.1	PPB	11/1/2023 16:09 SMPL
O4969-21	YC6E5	14805	3.1491 -	SFAM01.1	PPB	11/1/2023 16:11 SMPL
O4969-22	YC6E6	9561	2.0351 -	SFAM01.1	PPB	11/1/2023 16:13 SMPL
PB156814BL	PBS814	-103	-0.0179 -	SFAM01.1	PPB	11/1/2023 16:16 SMPL
O4970-01	YC6A0	103802	22.055 -	SFAM01.1	PPB	11/1/2023 16:18 SMPL
O4970-02	YC6A1	23468	4.9894 -	SFAM01.1	PPB	11/1/2023 16:20 SMPL
O4970-03	YC6A2	87386	18.5677 -	SFAM01.1	PPB	11/1/2023 16:23 SMPL
O4970-04	YC6A3	161549	34.3224 -	SFAM01.1	PPB	11/1/2023 16:25 SMPL
O4970-05	YC6A4	215426	45.7677 -	SFAM01.1	PPB	11/1/2023 16:27 SMPL
O4970-06	YC6A5	238112	50.587 -	SFAM01.1	PPB	11/1/2023 16:29 SMPL
O4970-07	YC6A6	105958	22.513 -	SFAM01.1	PPB	11/1/2023 16:32 SMPL
O4970-08	YC6A7	94057	19.9848 -	SFAM01.1	PPB	11/1/2023 16:34 SMPL
O4970-09	YC6A8	69721	14.8151 -	SFAM01.1	PPB	11/1/2023 16:36 SMPL
O4970-10	YC6A8D	80446	17.0934 -	SFAM01.1	PPB	11/1/2023 16:39 SMPL
O4970-11	YC6A8S	91356	19.4111 -	SFAM01.1	PPB	11/1/2023 16:41 SMPL
O4970-12	YC6B1	111700	23.7328 -	SFAM01.1	PPB	11/1/2023 16:43 SMPL
O4970-13	YC6A9	16053	3.4142 -	SFAM01.1	PPB	11/1/2023 16:46 SMPL
CCV027	CCV027	23875	5.0758 -	SFAM01.1	PPB	11/1/2023 16:48 SMPL
CCB027	CCB027	-82	-0.0134 -	SFAM01.1	PPB	11/1/2023 16:50 SMPL
O4970-14	YC6B0	14273	3.036 -	SFAM01.1	PPB	11/1/2023 16:52 SMPL
O4970-15	YC6B2	64357	13.6756 -	SFAM01.1	PPB	11/1/2023 16:55 SMPL
O4970-16	YC6B3	120927	25.6929 -	SFAM01.1	PPB	11/1/2023 16:57 SMPL
O4970-17	YC6B4	125786	26.7251 -	SFAM01.1	PPB	11/1/2023 16:59 SMPL
O4970-18	YC6B5	28506	6.0596 -	SFAM01.1	PPB	11/1/2023 17:01 SMPL
O4970-19	YC6B6	30689	6.5234 -	SFAM01.1	PPB	11/1/2023 17:04 SMPL
O4970-20	YC6B7	98896	21.0128 -	SFAM01.1	PPB	11/1/2023 17:06 SMPL
O4970-21	YC6B8	8807	1.8749 -	SFAM01.1	PPB	11/1/2023 17:08 SMPL
O4970-22	YC6B9	111308	23.6495 -	SFAM01.1	PPB	11/1/2023 17:11 SMPL
PB156815BL	PBS815	-167	-0.0315 -	SFAM01.1	PPB	11/1/2023 17:13 SMPL
O4971-01	YC6C0	74857	15.9061 -	SFAM01.1	PPB	11/1/2023 17:15 SMPL
O4971-02	YC6C0D	72431	15.3908 -	SFAM01.1	PPB	11/1/2023 17:17 SMPL
O4971-03	YC6C0S	96100	20.4188 -	SFAM01.1	PPB	11/1/2023 17:20 SMPL
O4971-04	YC6C1	15537	3.3046 -	SFAM01.1	PPB	11/1/2023 17:22 SMPL
O4971-05	YC6C2	55003	11.6885 -	SFAM01.1	PPB	11/1/2023 17:24 SMPL
O4971-06	YC6C3	72262	15.3549 -	SFAM01.1	PPB	11/1/2023 17:27 SMPL
O4971-07	YC6C4	45358	9.6395 -	SFAM01.1	PPB	11/1/2023 17:29 SMPL

INSTRUMENT ID : CV1

O4971-08	YC6C5	72531	15.412 -	SFAM01.1	PPB	11/1/2023 17:31 SMPL
O4971-09	YC6C6	50543	10.741 -	SFAM01.1	PPB	11/1/2023 17:33 SMPL
O4971-10	YC6C7	57042	12.1216 -	SFAM01.1	PPB	11/1/2023 17:36 SMPL
CCV028	CCV028	24558	5.2209 -	SFAM01.1	PPB	11/1/2023 17:38 SMPL
CCB028	CCB028	-40	-0.0045 -	SFAM01.1	PPB	11/1/2023 17:40 SMPL
O4971-11	YC6C8	73805	15.6826 -	SFAM01.1	PPB	11/1/2023 17:43 SMPL
O4971-12	YC6C9	79496	16.8916 -	SFAM01.1	PPB	11/1/2023 17:45 SMPL
O4971-13	YC6D0	98823	20.9973 -	SFAM01.1	PPB	11/1/2023 17:47 SMPL
O4971-14	YC6D1	108018	22.9506 -	SFAM01.1	PPB	11/1/2023 17:49 SMPL
O4971-15	YC6D2	8192	1.7442 -	SFAM01.1	PPB	11/1/2023 17:52 SMPL
O4971-16	YC6D3	13189	2.8058 -	SFAM01.1	PPB	11/1/2023 17:54 SMPL
O4971-17	YC6D4	12404	2.639 -	SFAM01.1	PPB	11/1/2023 17:56 SMPL
O4971-18	YC6D5	6083	1.2962 -	SFAM01.1	PPB	11/1/2023 17:59 SMPL
O4971-19	YC6D6	6886	1.4668 -	SFAM01.1	PPB	11/1/2023 18:01 SMPL
O4971-20	YC6D7	7080	1.508 -	SFAM01.1	PPB	11/1/2023 18:03 SMPL
O4971-21	YC6D8	3443	0.7354 -	SFAM01.1	PPB	11/1/2023 18:05 SMPL
O4971-22	YC6D9	4181	0.8922 -	SFAM01.1	PPB	11/1/2023 18:08 SMPL
O4968-02DLX2		25039	5.3231 -	SFAM01.1	PPB	11/1/2023 18:10 SMPL
O4969-01DLX2		27900	5.9309 -	SFAM01.1	PPB	11/1/2023 18:12 SMPL
O4969-05DLX2		25814	5.4877 -	SFAM01.1	PPB	11/1/2023 18:14 SMPL
O4969-13DLX5		14280	3.0375 -	SFAM01.1	PPB	11/1/2023 18:19 SMPL
O4969-15DLX5		14712	3.1293 -	SFAM01.1	PPB	11/1/2023 18:21 SMPL
O4969-16DLX5		22115	4.7019 -	SFAM01.1	PPB	11/1/2023 18:24 SMPL
CCV029	CCV029	24575	5.2245 -	SFAM01.1	PPB	11/1/2023 18:26 SMPL
CCB029	CCB029	-60	-0.0088 -	SFAM01.1	PPB	11/1/2023 18:28 SMPL
O4969-17DLX5		12839	2.7314 -	SFAM01.1	PPB	11/1/2023 18:30 SMPL
O4969-20DLX5		19151	4.0723 -	SFAM01.1	PPB	11/1/2023 18:33 SMPL
O4970-01DLX5		24565	5.2224 -	SFAM01.1	PPB	11/1/2023 18:35 SMPL
CCV030	CCV030	24610	5.232 -	SFAM01.1	PPB	11/1/2023 18:40 SMPL
CCB030	CCB030	-72	-0.0113 -	SFAM01.1	PPB	11/1/2023 18:42 SMPL