



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB128097

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.9	4.01	10/31/2023	17:38
2	CAL2	1	Water	NA	NA	20.9	7.00	10/31/2023	17:39
3	CAL3	1	Water	NA	NA	20.8	10.02	10/31/2023	17:40
4	ICV	1	Water	NA	NA	2.6	7.01	10/31/2023	17:45
5	CCV1	1	Water	NA	NA	20.8	2.01	10/31/2023	17:47
6	O5172-01	1	Solid	20.02	20	23.6	8.96	10/31/2023	17:55
7	O5172-01DUP	1	Solid	20.03	20	23.7	8.97	10/31/2023	17:56
8	CCV2	1	Water	NA	NA	20.7	12.02	10/31/2023	18:00

WORKLIST(Hardcopy Internal Chain)

VB 08097

WorkList Name : ph s 05172

WorkList ID : 175150

Department : Wet-Chemistry

Date : 10-31-2023 17:23:01

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
05172-01 D	AU-05-103123	Solid	pH	Cool 4 deg C	PSEG05	I41	10/31/2023	9045D

Date/Time

10/31/23 17:30

Raw Sample Received by:

20 GWC

Raw Sample Relinquished by:

A-C 0501

Date/Time

10/31/23

Raw Sample Received by:

20 GWC

Raw Sample Relinquished by:

20 GWC