

LB12810E/


7471B

INSTRUMENT ID: -CV1

Linear

μ Abs.:

7092E



A= 0.0000e+000

B= 1.3549e-004

C= 5.1564e-002

Rho= 0.9973865

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	4.D
0.0	0.000	0.051	0.051	-5	0.000	-5					
0.2	0.200	0.241	0.041	1397	0.0 %	1397					21
2.5	2.500	2.147	-0.353	15465	0.0 %	15465					-14
5.0	5.000	5.266	0.266	38486	0.0 %	38486					5
7.5	7.500	7.834	0.334	57444	0.0 %	57444					4
10.0	10.000	9.661	-0.339	70925	0.0 %	70925					-3

LB128104 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 01 Nov 2023 10:18:08

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SV	-5	-	0.0000	7471B	PPB	01 Nov 2023 10:22:50	S	Std
0.2 - 1	SD-2	1397	-	0.2000	7471B	PPB	01 Nov 2023 10:25:08	S	Std
2.5 - 1	SD-5	15465	-	2.5000	7471B	PPB	01 Nov 2023 10:27:24	S	Std
5.0 - 1	SD	38486	-	5.0000	7471B	PPB	01 Nov 2023 10:29:41	S	Std
7.5 - 1	SD-T	57444	-	7.5000	7471B	PPB	01 Nov 2023 10:31:57	S	Std
10.0 - 1	SD	70925	-	10.0000	7471B	PPB	01 Nov 2023 10:36:54	S	Std
ICV48 - 1	ICV48	31504	4.3200	-	7471B	PPB	01 Nov 2023 10:40:42	U	SMPL
ICB48 - 1	ICB48	-373	0.0010	-	7471B	PPB	01 Nov 2023 10:42:56	U	SMPL
CCV24 - 1	CCV24	39558	5.4112	-	7471B	PPB	01 Nov 2023 10:45:14	U	SMPL
CCB24 - 1	CCB24	-455	-0.0101	-	7471B	PPB	01 Nov 2023 10:47:29	U	SMPL
CRA - 1	CRA	1285	0.2257	-	7471B	PPB	01 Nov 2023 10:49:46	U	SMPL
HighStd - 1	HighStd	71026	9.6747	-	7471B	PPB	01 Nov 2023 10:52:01	U	SMPL
ChkStd - 1	ChkStd	54898	7.4895	-	7471B	PPB	01 Nov 2023 10:54:17	U	SMPL
PB156833BL - 1	PBS	-790	-0.0555	-	7471B	PPB	01 Nov 2023 10:56:36	U	SMPL
PB156833BS - 1	LCSS	30274	4.1533	-	7471B	PPB	01 Nov 2023 10:58:55	U	SMPL
O5162-01 - 1	HD-01-10302023	7169	1.0229	-	7471B	PPB	01 Nov 2023 11:01:12	U	SMPL
O5162-03 - 1	HD-02-10302023	2923	0.4476	-	7471B	PPB	01 Nov 2023 11:03:28	U	SMPL
O5164-01 - 1	ETGI-352	2570	0.3998	-	7471B	PPB	01 Nov 2023 11:05:47	U	SMPL
O5167-01 - 1	WC-1	18980	2.6231	-	7471B	PPB	01 Nov 2023 11:08:05	U	SMPL
O5169-01 - 1	SU-02-10302023	2529	0.3942	-	7471B	PPB	01 Nov 2023 11:10:19	U	SMPL
CCV25 - 1	CCV25	38356	5.2483	-	7471B	PPB	01 Nov 2023 11:12:35	U	SMPL
CCB25 - 1	CCB25	-83	0.0403	-	7471B	PPB	01 Nov 2023 11:14:52	U	SMPL
O5169-01DUP - 1	SU-02-10302023DUP	2546	0.3965	-	7471B	PPB	01 Nov 2023 11:17:08	U	SMPL
O5169-01MS - 1	SU-02-10302023MS	37005	5.0653	-	7471B	PPB	01 Nov 2023 11:19:24	U	SMPL
O5169-01MSD - 1	SU-02-10302023MSD	36248	4.9627	-	7471B	PPB	01 Nov 2023 11:21:39	U	SMPL
O5169-01LX5 - 1		432	0.1101	-	7471B	PPB	01 Nov 2023 11:23:56	U	SMPL
O5169-01A - 1		34710	4.7543	-	7471B	PPB	01 Nov 2023 11:26:12	U	SMPL
CCV26 - 1	CCV26	38630	5.2854	-	7471B	PPB	01 Nov 2023 11:28:28	U	SMPL
CCB26 - 1	CCB26	-52	0.0445	-	7471B	PPB	01 Nov 2023 11:30:45	U	SMPL