

LB128117

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

11/1/2023 14:33 Reviewed by : 12 Instrument ID : Konelab

Test: Phenolics-

Sample Id	^{well} Result	Dil. 1 +	Response	Errors
ICV1	1.014	0.0	0.101	
ICB1	-0.001	0.0	0.021	
CCV1	1.015	0.0	0.101	
CCB1	0.007	0.0	0.022	
PB156843BL	0.008	0.0	0.022	
PB156843BS	1.016	0.0	0.101	
O4699-07	0.028	0.0	0.023	
O4699-08	0.055	0.0	0.026	
PB156844BL	0.005	0.0	0.022	
PB156844BS	1.010	0.0	0.100	
O4699-01	0.032	0.0	0.024	
O4699-02	0.053	0.0	0.025	
CCV2	1.018	0.0	0.101	
CCB2	0.003	0.0	0.021	

N	14
Mean	0.376
SD	0.4942
CV%	131.52

Aquakem v. 7.2AQ1

Results from time period:

Wed Nov 01 13:42:58 2023

Wed Nov 01 14:33:06 2023

Sample Id	Sam/Ctr/c	Test short	Test type	Result	Result unit	Result date and time	Stat
0.00PPM	A	Phenolics-	P	0.0094	mg/l	11/1/2023 13:42:58	
0.05PPM	A	Phenolics-	P	0.0565	mg/l	11/1/2023 13:42:59	
0.1PPM	A	Phenolics-	P	0.0966	mg/l	11/1/2023 13:43:00	
0.25PPM	A	Phenolics-	P	0.2497	mg/l	11/1/2023 13:43:01	
0.50PPM	A	Phenolics-	P	0.5018	mg/l	11/1/2023 13:43:02	
1.0PPM	A	Phenolics-	P	0.9727	mg/l	11/1/2023 13:43:03	
2.0PPM	A	Phenolics-	P	2.0132	mg/l	11/1/2023 13:43:04	
ICV1	S	Phenolics-	P	1.0138	mg/l	11/1/2023 14:17:38	
ICB1	S	Phenolics-	P	-0.0009	mg/l	11/1/2023 14:17:41	
CCV1	S	Phenolics-	P	1.0146	mg/l	11/1/2023 14:17:42	
CCB1	S	Phenolics-	P	0.0069	mg/l	11/1/2023 14:17:44	
PB156843BL	S	Phenolics-	P	0.0079	mg/l	11/1/2023 14:17:48	
PB156843BS	S	Phenolics-	P	1.0156	mg/l	11/1/2023 14:27:40	
O4699-07	S	Phenolics-	P	0.0277	mg/l	11/1/2023 14:27:43	
O4699-08	S	Phenolics-	P	0.0552	mg/l	11/1/2023 14:27:44	
PB156844BL	S	Phenolics-	P	0.005	mg/l	11/1/2023 14:27:46	
PB156844BS	S	Phenolics-	P	1.0097	mg/l	11/1/2023 14:27:48	
O4699-01	S	Phenolics-	P	0.0318	mg/l	11/1/2023 14:27:51	
O4699-02	S	Phenolics-	P	0.053	mg/l	11/1/2023 14:33:03	
CCV2	S	Phenolics-	P	1.0176	mg/l	11/1/2023 14:33:05	
CCB2	S	Phenolics-	P	0.0025	mg/l	11/1/2023 14:33:06	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12 Instrument ID : Konelab

11/1/2023 13:47


Test Phenolics-

Accepted 11/1/2023 13:47

Factor 12.75
Bias 0.021

Coeff. of det. 0.999662

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.00PPM	0.022	0.0094	0.0000	—
2	Phenol-2	0.026	0.0565	0.0500	13.0
3	Phenol-2	0.029	0.0966	0.1000	-3.4
4	Phenol-2	0.041	0.2497	0.2500	-0.1
5	Phenol-2	0.061	0.5018	0.5000	0.4
6	Phenol-2	0.098	0.9727	1.0000	-2.7
7	Phenol-2	0.179	2.0132	2.0000	0.7

12

11/1/23