

LB128118

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC

284 Sheffield Street, Mountainside, NJ 07092

11/1/2023 15:36

Reviewed by : 12

Instrument ID : Konelab

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.021	0.0	0.101	
ICB1	0.009	0.0	0.022	
CCV1	1.029	0.0	0.102	
CCB1	0.010	0.0	0.022	
PB156845BL	0.006	0.0	0.022	
PB156845BS	1.027	0.0	0.102	
O4699-12	0.032	0.0	0.024	
CCV2	1.024	0.0	0.102	
CCB2	0.009	0.0	0.022	

N	9
Mean	0.463
SD	0.5335
CV%	115.20

Aquakem v. 7.2AQ1

Results from time period:

Wed Nov 01 14:55:05 2023

Wed Nov 01 15:01:00 2023

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.00PPM	A	Phenolics-	P	0.0094	mg/l	11/1/2023 13:42:58	
0.05PPM	A	Phenolics-	P	0.0565	mg/l	11/1/2023 13:42:59	
0.1PPM	A	Phenolics-	P	0.0966	mg/l	11/1/2023 13:43:00	
0.25PPM	A	Phenolics-	P	0.2497	mg/l	11/1/2023 13:43:01	
0.50PPM	A	Phenolics-	P	0.5018	mg/l	11/1/2023 13:43:02	
1.0PPM	A	Phenolics-	P	0.9727	mg/l	11/1/2023 13:43:03	
2.0PPM	A	Phenolics-	P	2.0132	mg/l	11/1/2023 13:43:04	
ICV1	S	Phenolics-	P	1.0213	mg/l	11/1/2023 14:55:06	
ICB1	S	Phenolics-	P	0.0093	mg/l	11/1/2023 14:55:07	
CCV1	S	Phenolics-	P	1.0286	mg/l	11/1/2023 14:55:09	
CCB1	S	Phenolics-	P	0.0096	mg/l	11/1/2023 14:55:12	
PB156845BL	S	Phenolics-	P	0.0059	mg/l	11/1/2023 14:55:14	
PB156845BS	S	Phenolics-	P	1.0273	mg/l	11/1/2023 14:55:16	
O4699-12	S	Phenolics-	P	0.0324	mg/l	11/1/2023 15:00:56	
CCV2	S	Phenolics-	P	1.0243	mg/l	11/1/2023 15:00:58	
CCB2	S	Phenolics-	P	0.009	mg/l	11/1/2023 15:01:00	

CHEMTECH CONSULTING GROUP INC

284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12

Instrument ID : Konelab

11/1/2023 13:47

Test Phenolics-


Accepted 11/1/2023 13:47

Factor 12.75

Bias 0.021

Coeff. of det. 0.999662

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.00PPM	0.022	0.0094	0.0000	—
2	Phenol-2	0.026	0.0565	0.0500	13.0
3	Phenol-2	0.029	0.0966	0.1000	-34
4	Phenol-2	0.041	0.2497	0.2500	-01
5	Phenol-2	0.061	0.5018	0.5000	0.4
6	Phenol-2	0.098	0.9727	1.0000	-2.7
7	Phenol-2	0.179	2.0132	2.0000	0.7

12

11/1/23