

LB128123


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

50248



A= 0.0000e+000

B= 2.0279e-004

C= 2.2120e-002

Rho= 0.9989558

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.014	0.014	-42	0.000	-42	0				4
0.2	0.200	0.209	0.009	922	0.0 %	922					7
2.5	2.500	2.679	0.179	13100	0.0 %	13100					-3
5.0	5.000	4.856	-0.144	23834	0.0 %	23834					-4
7.5	7.500	7.231	-0.269	35547	0.0 %	35547					2
10.0	10.000	10.212	0.212	50248	0.0 %	50248					

LB128123 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 02 Nov 2023 08:52:32

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-42	-	0.0000	7470A	PPB	02 Nov 2023 09:11:41	S	Std
0.2 - 1	50.2	922	-	0.2000	7470A	PPB	02 Nov 2023 09:13:58	S	Std
2.5 - 1	52.5	13100	-	2.5000	7470A	PPB	02 Nov 2023 09:19:54	S	Std
5.0 - 1	55	23834	-	5.0000	7470A	PPB	02 Nov 2023 09:22:11	S	Std
7.5 - 1	57.5	35547	-	7.5000	7470A	PPB	02 Nov 2023 09:24:27	S	Std
10.0 - 1	510	50248	-	10.0000	7470A	PPB	02 Nov 2023 09:26:44	S	Std
ICV49 - 1	ICV49	18565	3.7870	-	7470A	PPB	02 Nov 2023 09:29:40	U	SMPL
ICB49 - 1	ICB49	-435	-0.0661	-	7470A	PPB	02 Nov 2023 09:31:56	U	SMPL
CCV31 - 1	CCV31	22257	4.5357	-	7470A	PPB	02 Nov 2023 09:34:11	U	SMPL
CCB31 - 1	CCB31	-408	-0.0606	-	7470A	PPB	02 Nov 2023 09:36:26	U	SMPL
CRA - 1	CRA	809	0.1862	-	7470A	PPB	02 Nov 2023 09:38:43	U	SMPL
HighStd - 1	HighStd	54115	10.9964	-	7470A	PPB	02 Nov 2023 09:40:58	U	SMPL
ChkStd - 1	ChkStd	35971	7.3169	-	7470A	PPB	02 Nov 2023 09:43:14	U	SMPL
PB156846BL - 1	PBW	-584	-0.0963	-	7470A	PPB	02 Nov 2023 09:45:31	U	SMPL
PB156846BS - 1	LCSW	20398	4.1587	-	7470A	PPB	02 Nov 2023 09:47:49	U	SMPL
O5164-02 - 1	ETGI-352	-354	-0.0497	-	7470A	PPB	02 Nov 2023 09:50:05	U	SMPL
O5164-02DUP - 1	ETGI-352DUP	-150	-0.0083	-	7470A	PPB	02 Nov 2023 09:52:23	U	SMPL
O5164-02MS - 1	ETGI-352MS	19720	4.0212	-	7470A	PPB	02 Nov 2023 09:54:39	U	SMPL
O5164-02MSD - 1	ETGI-352MSD	21336	4.3490	-	7470A	PPB	02 Nov 2023 09:56:56	U	SMPL
CCV32 - 1	CCV32	22269	4.5382	-	7470A	PPB	02 Nov 2023 10:03:19	U	SMPL
CCB32 - 1	CCB32	-396	-0.0582	-	7470A	PPB	02 Nov 2023 10:05:34	U	SMPL
PB156788TB - 1	PB156788TB	-170	-0.0124	-	7470A	PPB	02 Nov 2023 10:07:50	U	SMPL
O5164-02LX5 - 1		29	0.0280	-	7470A	PPB	02 Nov 2023 10:10:05	U	SMPL
O5164-02A - 1		20101	4.0985	-	7470A	PPB	02 Nov 2023 10:12:21	U	SMPL
CCV33 - 1	CCV33	26098	5.3147	-	7470A	PPB	02 Nov 2023 10:14:36	U	SMPL
CCB33 - 1	CCB33	-432	-0.0655	-	7470A	PPB	02 Nov 2023 10:16:53	U	SMPL