

LB128124


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

50730



A= 0.0000e+000

B= 1.9536e-004

C= -2.4079e-002

Rho= 0.9997154

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% (1)
0.0	0.000	-0.045	-0.045	-109	0.000	-109					3
0.2	0.200	0.206	0.006	1178	0.0 %	1178					3
2.5	2.500	2.436	-0.064	12591	0.0 %	12591					3
5.0	5.000	5.152	0.152	26496	0.0 %	26496					3
7.5	7.500	7.565	0.065	38847	0.0 %	38847					3
10.0	10.000	9.886	-0.114	50730	0.0 %	50730					1

LB128124 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 02 Nov 2023 10:25:42

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	-109	-	0.0000	7471B	PPB	02 Nov 2023 10:53:12	S	Std
0.2 - 1	S02	1178	-	0.2000	7471B	PPB	02 Nov 2023 10:55:28	S	Std
2.5 - 1	S25	12591	-	2.5000	7471B	PPB	02 Nov 2023 10:57:45	S	Std
5.0 - 1	S5	26496	-	5.0000	7471B	PPB	02 Nov 2023 11:00:03	S	Std
7.5 - 1	S75	38847	-	7.5000	7471B	PPB	02 Nov 2023 11:04:58	S	Std
10.0 - 1	S10	50730	-	10.0000	7471B	PPB	02 Nov 2023 11:07:15	S	Std
ICV50 - 1	ICV50	22321	4.3365	-	7471B	PPB	02 Nov 2023 11:13:05	U	SMPL
ICB50 - 1	ICB50	-358	-0.0940	-	7471B	PPB	02 Nov 2023 11:15:20	U	SMPL
CCV34 - 1	CCV34	26526	5.1580	-	7471B	PPB	02 Nov 2023 11:17:36	U	SMPL
CCB34 - 1	CCB34	-343	-0.0911	-	7471B	PPB	02 Nov 2023 11:19:52	U	SMPL
CRA - 1	CRA	1060	0.1830	-	7471B	PPB	02 Nov 2023 11:22:09	U	SMPL
HighStd - 1	HighStd	50900	9.9197	-	7471B	PPB	02 Nov 2023 11:24:24	U	SMPL
ChkStd - 1	ChkStd	34555	6.7265	-	7471B	PPB	02 Nov 2023 11:26:40	U	SMPL
PB156858BL - 1	PBS	-104	-0.0444	-	7471B	PPB	02 Nov 2023 11:31:39	U	SMPL
PB156858BS - 1	LCSS	20207	3.9235	-	7471B	PPB	02 Nov 2023 11:36:26	U	SMPL
O5107-01 - 1	1048	4252	0.8066	-	7471B	PPB	02 Nov 2023 11:43:38	U	SMPL
O5107-01DUP - 1	1048DUP	3712	0.7011	-	7471B	PPB	02 Nov 2023 11:45:54	U	SMPL
O5107-01MS - 1	1048MS	21747	4.2244	-	7471B	PPB	02 Nov 2023 11:48:12	U	SMPL
O5107-01MSD - 1	1048MSD	21295	4.1361	-	7471B	PPB	02 Nov 2023 11:50:29	U	SMPL
O5107-02 - 1	1049	9993	1.9281	-	7471B	PPB	02 Nov 2023 11:52:46	U	SMPL
CCV35 - 1	CCV35	27617	5.3711	-	7471B	PPB	02 Nov 2023 11:55:03	U	SMPL
CCB35 - 1	CCB35	-86	-0.0409	-	7471B	PPB	02 Nov 2023 11:57:18	U	SMPL
O5107-03 - 1	1073	3465	0.6528	-	7471B	PPB	02 Nov 2023 11:59:34	U	SMPL
O5107-04 - 1	1074	5127	0.9775	-	7471B	PPB	02 Nov 2023 12:01:49	U	SMPL
O5107-05 - 1	1075	15537	3.0112	-	7471B	PPB	02 Nov 2023 12:04:04	U	SMPL
O5107-06 - 1	1076	14403	2.7897	-	7471B	PPB	02 Nov 2023 12:06:19	U	SMPL
O5107-07 - 1	1077	3105	0.5825	-	7471B	PPB	02 Nov 2023 12:08:36	U	SMPL
O5107-08 - 1	1078	8416	1.6201	-	7471B	PPB	02 Nov 2023 12:10:53	U	SMPL
O5107-09 - 1	1079	1521	0.2731	-	7471B	PPB	02 Nov 2023 12:12:05	U	SMPL
O5107-10 - 1	1080	5678	1.0852	-	7471B	PPB	02 Nov 2023 12:14:53	U	SMPL
O5107-11 - 1	1081	1801	0.3278	-	7471B	PPB	02 Nov 2023 12:17:10	U	SMPL
O5107-23 - 1	1047	686	0.1099	-	7471B	PPB	02 Nov 2023 12:19:27	U	SMPL
CCV36 - 1	CCV36	23874	4.6399	-	7471B	PPB	02 Nov 2023 12:21:42	U	SMPL
CCB36 - 1	CCB36	-130	-0.0495	-	7471B	PPB	02 Nov 2023 12:23:58	U	SMPL
O5170-01 - 1	OK-02-103123	1120	0.1947	-	7471B	PPB	02 Nov 2023 12:26:14	U	SMPL
O5171-01 - 1	TR-05-103123	1278	0.2256	-	7471B	PPB	02 Nov 2023 12:28:30	U	SMPL
O5172-01 - 1	AU-05-103123	4510	0.8570	-	7471B	PPB	02 Nov 2023 12:30:45	U	SMPL
O5197-01 - 1	SCF	15663	3.0358	-	7471B	PPB	02 Nov 2023 12:33:00	U	SMPL
O5107-01LX5 - 1		818	0.1357	-	7471B	PPB	02 Nov 2023 12:35:17	U	SMPL
O5107-01A - 1		23259	4.5198	-	7471B	PPB	02 Nov 2023 12:37:33	U	SMPL
CCV37 - 1	CCV37	25770	5.0103	-	7471B	PPB	02 Nov 2023 12:39:49	U	SMPL
CCB37 - 1	CCB37	-72	-0.0381	-	7471B	PPB	02 Nov 2023 12:42:07	U	SMPL