



PERCENT SOLID

Supervisor: Iwona

Analyst: JIGNESH

Date: 11/3/2023

OVENTEMP IN Celsius(°C): 107

Time IN: 13:25

In Date: 11/02/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 07:45

Out Date: 11/03/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB128132

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O5174-02	COAM5	1	1.14	8.80	9.94	8.49	83.5	
O5174-03	COAM6	2	1.15	8.38	9.53	8.17	83.8	
O5174-04	COAM7	3	1.19	8.54	9.73	7.86	78.1	
O5174-05	COAM8	4	1.19	8.79	9.98	8.4	82.0	
O5174-06	COAM9	5	1.19	8.45	9.64	8.25	83.6	
O5174-08	VHBLK002	6	1.00	1.00	2.00	2.00	100.0	VHBLK
O5174-09	COAN1	7	1.19	8.60	9.79	8.99	90.7	
O5174-10	COAN2	8	1.19	8.50	9.69	8.62	87.4	
O5174-11	COAN3	9	1.15	8.82	9.97	8.41	82.3	
O5174-12	COAN4	10	1.19	8.73	9.92	8.06	78.7	
O5174-13	COAN5	11	1.19	8.58	9.77	8.1	80.5	
O5174-14	COAN6	12	1.19	8.59	9.78	8.98	90.7	
O5174-15	COAN7	13	1.17	8.60	9.77	8.81	88.8	
O5174-17	COAN1	14	1.18	8.78	9.96	9.46	94.3	
O5174-18	COAN2	15	1.14	8.80	9.94	8.98	89.1	
O5174-19	COAN6	16	1.13	8.76	9.89	9.13	91.3	
O5174-20	COAN6MS	17	1.13	8.76	9.89	9.13	91.3	
O5174-21	COAN6MSD	18	1.13	8.76	9.89	9.13	91.3	
O5174-22	COAN7	19	1.19	8.55	9.74	8.85	89.6	
O5174-23	COAP0	20	1.19	8.79	9.98	8.54	83.6	
O5174-24	COAP1	21	1.19	8.62	9.81	8.51	84.9	
O5174-25	COAP2	22	1.18	8.79	9.97	8.81	86.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$