



PERCENT SOLID

Supervisor: Iwona
Analyst: JIGNESH
Date: 11/3/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 13:50
In Date: 11/02/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:45
Out Date: 11/03/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB128135

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O5185-01	EZNL8	1	1.18	8.67	9.85	8.35	82.7	
O5185-02	EZNL8D	2	1.18	8.67	9.85	8.35	82.7	
O5185-03	EZNL8S	3	1.18	8.67	9.85	8.35	82.7	
O5185-04	EZNP8	4	1.15	8.53	9.68	8.87	90.5	
O5185-05	EZNP9	5	1.11	8.79	9.9	9.00	89.8	
O5185-06	EZNQ0	6	1.14	8.40	9.54	8.91	92.5	
O5185-07	EZNQ1	7	1.15	8.71	9.86	9.22	92.7	
O5185-08	EZNQ2	8	1.12	8.46	9.58	8.98	92.9	
O5185-09	EZNQ3	9	1.18	8.63	9.81	9.08	91.5	
O5185-10	EZNQ4	10	1.15	8.82	9.97	9.22	91.5	
O5185-11	EZNQ5	11	1.15	8.40	9.55	8.44	86.8	
O5185-12	EZNQ6	12	1.18	8.51	9.69	8.72	88.6	
O5185-13	EZNQ7	13	1.14	8.40	9.54	8.74	90.5	
O5185-14	EZNQ8	14	1.19	8.51	9.7	8.51	86.0	
O5185-15	EZNQ9	15	1.1	8.77	9.87	8.81	87.9	
O5185-16	EZNR0	16	1.16	8.70	9.86	8.95	89.5	
O5185-17	EZNR1	17	1.15	8.50	9.65	8.8	90.0	
O5185-18	EZNR2	18	1.18	8.46	9.64	8.73	89.2	
O5185-19	EZNR4	19	1.13	8.40	9.53	8.67	89.8	
O5185-20	EZNR5	20	1.15	8.64	9.79	8.47	84.7	
O5185-21	EZNR6	21	1.13	8.69	9.82	9.06	91.3	
O5185-22	EZNR7	22	1.18	8.32	9.5	8.86	92.3	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$