



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB128145

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	81.4	11/02/2023	12:00
2	O5217-01		1	12	100.00	>212.0	>212.0	11/02/2023	12:30
3	O5217-01DUP		1	12	100.00	>212.0	>212.0	11/02/2023	13:00
4	O5217-02		1	14	100.00	>212.0	>212.0	11/02/2023	13:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

6128145

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp-11-02

WorkList ID : 175198

Department : Wet-Chemistry

Date : 11-02-2023 08:21:00

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5217-01C	1193	Water	Flash Point	Cool 4 deg C	PSEG03	I31	11/01/2023	1010B
O5217-02 C	1194	Water	Flash Point	Cool 4 deg C	PSEG03	I31	11/01/2023	1010B

Date/Time 11/02/2023 11:50
Raw Sample Received by: PR1 CWC
Raw Sample Relinquished by: JH6001

Date/Time 11/02/2023 13:35
Raw Sample Received by: JH6001
Raw Sample Relinquished by: PR1 CWC