



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB128155

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.02	11/03/2023	08:20
2	CAL2	1	Water	NA	NA	20.4	7.00	11/03/2023	08:22
3	CAL3	1	Water	NA	NA	20.5	10.02	11/03/2023	08:25
4	ICV	1	Water	NA	NA	20.5	7.00	11/03/2023	08:30
5	CCV1	1	Water	NA	NA	20.5	2.01	11/03/2023	08:36
6	O5095-01	1	Water	NA	NA	20.6	6.84	11/03/2023	08:40
7	O5095-01DUP	1	Water	NA	NA	20.7	6.85	11/03/2023	08:41
8	O5217-01	1	Water	NA	NA	20.8	8.93	11/03/2023	08:45
9	O5227-01	1	Water	NA	NA	20.5	7.36	11/03/2023	08:50
10	O5227-03	1	Water	NA	NA	20.4	7.31	11/03/2023	09:00
11	O5230-03	1	Water	NA	NA	20.1	4.76	11/03/2023	09:05
12	O5230-05	1	Water	NA	NA	20.3	4.71	11/03/2023	09:10
13	O5230-06	1	Water	NA	NA	20.3	4.80	11/03/2023	09:15
14	O5230-07	1	Water	NA	NA	20.4	5.10	11/03/2023	09:20
15	O5230-08	1	Water	NA	NA	20.1	5.91	11/03/2023	09:25
16	CCV2	1	Water	NA	NA	20.5	12.03	11/03/2023	09:30

128155

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 05231

WorkList ID : 175229

Department : Wet-Chemistry

Date : 11-03-2023 08:02:10

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5095-01	R F07367	Water	pH	Cool 4 deg C	PSEG03	I41	11/02/2023	9040C
O5217-01	1193	Water	pH	Cool 4 deg C	PSEG03	I31	11/01/2023	9040C
O5227-01	M21005	Water	pH	Cool 4 deg C	PSEG03	I41	11/02/2023	9040C
O5227-03	C A6934	Water	pH	Cool 4 deg C	PSEG03	I41	11/02/2023	9040C
O5230-03	60088	Water	pH	Cool 4 deg C	PSEG03	I31	11/02/2023	9040C
O5230-05	60109	Water	pH	Cool 4 deg C	PSEG03	I31	11/02/2023	9040C
O5230-06	60110	Water	pH	Cool 4 deg C	PSEG03	I31	11/02/2023	9040C
O5230-07	60111	Water	pH	Cool 4 deg C	PSEG03	I31	11/02/2023	9040C
O5230-08	60112	Water	pH	Cool 4 deg C	PSEG03	I31	11/02/2023	9040C

Date/Time 11/03/23 08:10
Raw Sample Received by: 2060001
Raw Sample Relinquished by: 2060001

Date/Time 11/03/23 11:00
Raw Sample Received by: 2060001
Raw Sample Relinquished by: 2060001