



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Rubina

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: Iwona

Run Number: LB128158

Reagent/Standard	Lot/Log #
Residual chlorine LOD, 0.05PPM	WP105130
Residual chlorine ICV-LCS, 0.4PPM	WP105129
Chlorine Calibration std, 0.1ppm	WP105124
Chlorine Calibration std, 0.2ppm	WP105125
Chlorine Calibration std, 0.8ppm	WP105127
Chlorine Calibration std, 0.0ppm	WP105123
Chlorine Calibration std, 1.6ppm	WP105128
Residual Chlorine Calibration and CCV std, 0	WP105126
Total Chlorine Powder Pillows	W2759

Intercept: 0.0077

Slope: 1.0012

Regression: 0.999702

Seq	Lab ID	True Val (mg/l)	DF	Initial Reading	Final Reading	Difference	Result (mg/l)	%D	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	-0.01		11/02/2023	15:45
2	CAL2	0.1	1	0.000	0.120	0.120	0.11	12	11/02/2023	15:48
3	CAL3	0.2	1	0.000	0.190	0.190	0.18	-9	11/02/2023	15:51
4	CAL4	0.4	1	0.000	0.410	0.410	0.40	0.5	11/02/2023	15:54
5	CAL5	0.8	1	0.000	0.830	0.830	0.82	2.6	11/02/2023	15:57
6	CAL6	1.6	1	0.000	1.600	1.600	1.59	-0.6	11/02/2023	16:00



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Rubina

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: Iwona

Run Number: LB128158

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.4100	0.4100	0.4020	11/02/2023	16:03
2	ICB				1	0.0000	0.0000	0.0000	-0.0080	11/02/2023	16:06
3	CCV1			0.4	1	0.0000	0.3900	0.3900	0.3820	11/02/2023	16:09
4	CCB1				1	0.0000	0.0100	0.0100	0.0020	11/02/2023	16:12
5	LB128158BL	50	50		1	0.0000	0.0100	0.0100	0.0020	11/02/2023	16:15
6	LB128158BS	50	50	0.4	1	0.0000	0.4100	0.4100	0.4020	11/02/2023	16:18
7	O4699-09	50	50		1	0.0000	0.0500	0.0500	0.0420	11/02/2023	16:21
8	O5235-01	50	50		1	0.0000	0.0400	0.0400	0.0320	11/02/2023	16:24
9	O5235-01DUP	50	50		1	0.0000	0.0400	0.0400	0.0320	11/02/2023	16:30
10	O5235-01MS	50	50	0.4	1	0.0000	0.4000	0.4000	0.3920	11/02/2023	16:33
11	O5235-01MSD	50	50	0.4	1	0.0000	0.3900	0.3900	0.3820	11/02/2023	16:37
12	CCV2			0.4	1	0.0000	0.4200	0.4200	0.4120	11/02/2023	16:40
13	CCB2				1	0.0000	0.0000	0.0000	-0.0080	11/02/2023	16:43

WORKLIST(Hardcopy Internal Chain)

LB128158

WorkList Name : Residual Chl-110223

WorkList ID : 175220

Department : Wet-Chemistry

Date : 11-02-2023 14:57:27

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5235-01	WATER-TREATMENT-DISCHAI	Water	Residual Chlorine	Cool 4 deg C	VERI01	I41	11/02/2023	SM4500 Cl G

Date/Time 11/2/23 15:30
 Raw Sample Received by: 12 (VC) [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 11/2/23 16:50
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: 12 (VC) [Signature]