


LB128159

SFAM01.1 INSTRUMENT ID: CV1

Linear

μ Abs.:

49761



A= 0.0000e+000

B= 2.0040e-004 slope

C= -1.0990e-001 intercept

Rho= 0.9997075

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	(1,1)
0.00	0.000	-0.071	-0.071	194	0.000	194					
0.05	0.050					419	0				
0.20	0.200	0.213	0.013	1611	0.0 %	1611					67 MB
2.50	2.500	2.482	-0.018	12936	0.0 %	12936					-1 1/1/23
5.00	5.000	5.092	0.092	25956	0.0 %	25956					2
7.50	7.500	7.622	0.122	38582	0.0 %	38582					2
10.0	10.000	9.862	-0.138	49761	0.0 %	49761					-1

LB128159

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	194	-		0 SFAM01.1	PPB	11/3/2023 12:07	Std
	0.2 S01	1611	-		0.2 SFAM01.1	PPB	11/3/2023 12:10	Std
	2.5 S02	12936	-		2.5 SFAM01.1	PPB	11/3/2023 12:12	Std
	5 S03	25956	-		5 SFAM01.1	PPB	11/3/2023 12:14	Std
	7.5 S04	38582	-		7.5 SFAM01.1	PPB	11/3/2023 12:16	Std
	10 S05	49761	-		10 SFAM01.1	PPB	11/3/2023 12:19	Std
ICV052	ICV052	20742	4.0468 -		SFAM01.1	PPB	11/3/2023 12:21	SMPL
ICB052	ICB052	86	-0.0927 -		SFAM01.1	PPB	11/3/2023 12:26	SMPL
CCV042	CCV042	26119	5.1243 -		SFAM01.1	PPB	11/3/2023 12:28	SMPL
CCB042	CCB042	98	-0.0903 -		SFAM01.1	PPB	11/3/2023 12:33	SMPL
PB156886BL	PBW886	629	0.0161 -		SFAM01.1	PPB	11/3/2023 12:41	SMPL
O4921-10	MEZZB0	7	-0.1085 -		SFAM01.1	PPB	11/3/2023 12:43	SMPL
O4921-12	MEZZB1	23	-0.1053 -		SFAM01.1	PPB	11/3/2023 12:45	SMPL
O4972-10	YC6K3	130	-0.0839 -		SFAM01.1	PPB	11/3/2023 12:47	SMPL
O4967-10	MEZZ82	88	-0.0923 -		SFAM01.1	PPB	11/3/2023 12:50	SMPL
CCV043	CCV043	25053	4.9107 -		SFAM01.1	PPB	11/3/2023 12:52	SMPL
CCB043	CCB043	482	-0.0133 -		SFAM01.1	PPB	11/3/2023 12:58	SMPL