



PERCENT SOLID

Supervisor: Iwona

Analyst: JIGNESH

Date: 11/6/2023

OVENTEMP IN Celsius(°C): 107

Time IN: 13:15

In Date: 11/03/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 07:45

Out Date: 11/04/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB128160

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O5190-01	EZNZ7	1	1.15	8.49	9.64	8.74	89.4	
O5190-02	EZNZ8	2	1.15	8.42	9.57	8.8	90.9	
O5190-03	EZNZ9	3	1.19	8.43	9.62	8.88	91.2	
O5190-04	EZP00	4	1.19	8.46	9.65	8.81	90.1	
O5190-05	EZP01	5	1.19	8.47	9.66	8.75	89.3	
O5190-06	EZP02	6	1.19	8.56	9.75	8.82	89.1	
O5190-07	EZP03	7	1.16	8.51	9.67	8.8	89.8	
O5190-08	EZP03D	8	1.16	8.51	9.67	8.8	89.8	
O5190-09	EZP03S	9	1.16	8.51	9.67	8.8	89.8	
O5190-10	EZP04	10	1.17	8.60	9.77	8.89	89.8	
O5190-11	EZP05	11	1.16	8.49	9.65	8.74	89.3	
O5190-12	EZP06	12	1.16	8.69	9.85	8.99	90.1	
O5190-13	EZP07	13	1.19	8.58	9.77	8.89	89.7	
O5190-14	EZP08	14	1.15	8.38	9.53	8.69	90.0	
O5190-15	EZP09	15	1.19	8.52	9.71	8.86	90.0	
O5190-16	EZP10	16	1.19	8.46	9.65	8.84	90.4	
O5190-17	EZP11	17	1.19	8.63	9.82	8.46	84.2	
O5190-18	EZP12	18	1.19	8.63	9.82	8.68	86.8	
O5190-19	EZP13	19	1.15	8.51	9.66	8.73	89.1	
O5190-20	EZP14	20	1.18	8.40	9.58	8.6	88.3	
O5190-21	EZP15	21	1.12	8.45	9.57	8.73	90.1	
O5190-22	EZP16	22	1.13	8.76	9.89	8.87	88.4	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$