



PERCENT SOLID

Supervisor: Iwona
Analyst: JIGNESH
Date: 11/6/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 14:30
In Date: 11/03/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:05
Out Date: 11/04/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB128164

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O5193-01	EZP48	1	1.18	8.72	9.9	8.64	85.6	
O5193-02	EZP48D	2	1.18	8.72	9.9	8.64	85.6	
O5193-03	EZP48S	3	1.18	8.72	9.9	8.64	85.6	
O5193-04	EZP60	4	1.19	8.63	9.82	7.53	73.5	
O5193-05	EZP61	5	1.16	8.69	9.85	7.62	74.3	
O5193-06	EZP62	6	1.1	8.78	9.88	8.5	84.3	
O5193-07	EZP63	7	1.19	8.62	9.81	8.81	88.4	
O5193-08	EZP64	8	1.18	8.54	9.72	7.87	78.3	
O5193-09	EZP65	9	1.19	8.72	9.91	8.22	80.6	
O5193-10	EZP66	10	1.18	8.44	9.62	7.85	79.0	
O5193-11	EZP67	11	1.15	8.54	9.69	7.8	77.9	
O5193-12	EZP68	12	1.19	8.69	9.88	8.14	80.0	
O5193-13	EZP69	13	1.18	8.40	9.58	8.44	86.4	
O5193-14	EZP70	14	1.16	8.50	9.66	8.09	81.5	
O5193-15	EZP71	15	1.19	8.62	9.81	8.2	81.3	
O5193-16	EZP72	16	1.19	8.53	9.72	8.29	83.2	
O5193-17	EZP73	17	1.12	8.77	9.89	7.83	76.5	
O5193-18	EZP74	18	1.13	8.56	9.69	8.17	82.2	
O5193-19	EZP75	19	1.15	8.40	9.55	8.61	88.8	
O5193-20	EZP76	20	1.17	8.57	9.74	8.22	82.3	
O5193-21	EZP77	21	1.1	8.72	9.82	8.64	86.5	
O5193-22	EZP78	22	1.11	8.73	9.84	8.87	88.9	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$