



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB128165

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.4	11/03/2023	08:40
2	O5095-01		1	12	100.00	>212.0	>212.0	11/03/2023	09:10
3	O5227-01		1	14	100.00	>212.0	>212.0	11/03/2023	09:40
4	O5227-03		1	12	100.00	>212.0	>212.0	11/03/2023	10:10
5	O5230-03		1	12	100.00	>212.0	>212.0	11/03/2023	10:40
6	O5230-04		1	14	100.00	>212.0	>212.0	11/03/2023	11:10
7	O5230-04DUP		1	14	100.00	>212.0	>212.0	11/03/2023	11:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB128165

WorkList Name : fp-11-03

WorkList ID : 175233

Department : Wet-Chemistry

Date : 11-03-2023 08:16:01

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5095-01 L	F07367	Water	Flash Point	Cool 4 deg C	PSEG03	I41	11/02/2023	1010B
O5227-01 D	M21005	Water	Flash Point	Cool 4 deg C	PSEG03	I41	11/02/2023	1010B
O5227-03 D	A6934	Water	Flash Point	Cool 4 deg C	PSEG03	I41	11/02/2023	1010B
O5230-03 D	60088	Water	Flash Point	Cool 4 deg C	PSEG03	I31	11/02/2023	1010B
O5230-04 A	60109-60112	Water	Flash Point	Cool 4 deg C	PSEG03	I31	11/02/2023	1010B

Date/Time 11/03/2023 08:30
 Raw Sample Received by: RAC
 Raw Sample Relinquished by: RAC

Date/Time 11/03/2023 12:15
 Raw Sample Received by: RAC
 Raw Sample Relinquished by: RAC