



TOTAL VOLATILE SOLIDS SM2540 G

TEMP1 IN: 104 °C 11/03/2023 16:45 TEMP1 OUT: 103 °C 11/06/2023 07:30
TEMP2 IN: 104 °C 11/06/2023 08:00 TEMP2 OUT: 104 °C 11/06/2023 09:30
TEMP3 IN: 540 °C 11/06/2023 10:00 TEMP3 OUT: 540 °C 11/06/2023 11:30
TEMP4 IN: 560 °C 11/06/2023 12:00 TEMP4 OUT: 560 °C 11/06/2023 13:30

Run Number: LB128176

SUPERVISOR: Iwona

ANALYST: JIGNESH

BalanceID: WC SC-6

OvenID: WC OVEN-1

Dish #	Lab ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish + SampleWt Drying @103-@105°C (g)	Final Dish + SampleWt Drying @103-@105°C (g)	Dish + Samplewt Drying @550 (±50) °C (g)	Final Dish + Samplewt Drying @550 (±50) °C (g)	Weight Diff (g)	Result (mg/L)
1	LB128176BL	65.9863	65.9863	100	65.9864	65.9864	65.9864	65.9864	0.0000	0
2	O5095-01	89.3273	89.3273	100	90.0265	90.0265	90.0027	90.0027	0.0238	238
3	O5095-01DUP	89.9173	89.9173	100	90.0199	90.0199	89.9953	89.9953	0.0246	246

A = Sample Volume (ml)

B = Final Dish + Samplewt Drying @550 (±50) °C (g)

C = Final Dish + SampleWt Drying @103-@105°C (g)

D = Weight (g)

E = Final Empty Dish Weight (g)

F = Final Dish + SampleWt Drying @103-@105°C (g)

Weight D = C - B

Result % = $\frac{D}{A} \star 1000000$

VB 128176

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ts water o5095 WorkList ID : 175277 Department : Wet-Chemistry Date : 11-03-2023 15:34:46

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5095-01	7 F07367	Water	TS	Cool 4 deg C	PSEG03	I41	11/02/2023	SM2540 B

Date/Time 11/03/23 15:40
Raw Sample Received by: JO GAG
Raw Sample Relinquished by: CP SM

Date/Time 11/03/23 17:00
Raw Sample Received by: CP SM
Raw Sample Relinquished by: JO GAG