

LB12818C

7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

64855

A= 0.0000e+000

B= 1.5356e-004

C= -1.1464e-002

Rho= 0.9998923

Accept=Accepted

Max. Conc: 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/6D
0.0	0.000	0.003	0.003	91	0.000	91					1
0.2	0.200	0.188	-0.012	1297	0.0 %	1297					6
2.5	2.500	2.522	0.022	16501	0.0 %	16501					1
5.0	5.000	4.938	-0.062	32233	0.0 %	32233					1
7.5	7.500	7.602	0.102	49579	0.0 %	49579					1
10.0	10.000	9.947	-0.053	64855	0.0 %	64855					1

LB128186 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 06 Nov 2023 08:28:51

Sample ID	Extended ID	Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	90	91	-	0.00007471B	PPB		06 Nov 2023 08:33:46	S	Std
0.2 - 1	51.2	1297	-	0.20007471B	PPB		06 Nov 2023 08:36:03	S	Std
2.5 - 1	52.5	16501	-	2.50007471B	PPB		06 Nov 2023 08:41:28	S	Std
5.0 - 1	55	32233	-	5.00007471B	PPB		06 Nov 2023 08:43:45	S	Std
7.5 - 1	57.5	49579	-	7.50007471B	PPB		06 Nov 2023 08:46:01	S	Std
10.0 - 1	510	64855	-	10.00007471B	PPB		06 Nov 2023 08:48:18	S	Std
ICV56 - 1	ICV56	26839	4.1099	-7471B	PPB		06 Nov 2023 08:51:11	U	SMPL
ICB56 - 1	ICB56	-320	-0.0606	-7471B	PPB		06 Nov 2023 08:53:26	U	SMPL
CCV53 - 1	CCV53	31445	4.8171	-7471B	PPB		06 Nov 2023 08:55:43	U	SMPL
CCB53 - 1	CCB53	-326	-0.0615	-7471B	PPB		06 Nov 2023 08:57:58	U	SMPL
CRA - 1	CRA	1176	0.1691	-7471B	PPB		06 Nov 2023 09:10:16	U	SMPL
HighStd - 1	HighStd	64961	9.9638	-7471B	PPB		06 Nov 2023 09:12:31	U	SMPL
ChkStd - 1	ChkStd	45014	6.9008	-7471B	PPB		06 Nov 2023 09:14:46	U	SMPL
PB156909BL - 1	PBS	-430	-0.0775	-7471B	PPB		06 Nov 2023 09:17:04	U	SMPL
PB156909BS - 1	LCSS	25567	3.9145	-7471B	PPB		06 Nov 2023 09:19:22	U	SMPL
O5213-02 - 1	ETGI-293	4959	0.7500	-7471B	PPB		06 Nov 2023 09:21:39	U	SMPL
O5214-02 - 1	VNJ-211	2530	0.3770	-7471B	PPB		06 Nov 2023 09:23:56	U	SMPL
O5214-04 - 1	VNJ-221	1907	0.2814	-7471B	PPB		06 Nov 2023 09:26:14	U	SMPL
O5214-05 - 1	NL-OILY-DEBRIS	460	0.0592	-7471B	PPB		06 Nov 2023 09:28:32	U	SMPL
O5215-01 - 1	600566096	5445	0.8247	-7471B	PPB		06 Nov 2023 09:30:48	U	SMPL
CCV54 - 1	CCV54	30749	4.7103	-7471B	PPB		06 Nov 2023 09:33:03	U	SMPL
CCB54 - 1	CCB54	-146	-0.0339	-7471B	PPB		06 Nov 2023 09:35:18	U	SMPL
O5216-02 - 1	AT-OILY-DEBRIS	99	0.0037	-7471B	PPB		06 Nov 2023 09:37:35	U	SMPL
O5220-02 - 1	Z-5(0-5)	18197	2.7828	-7471B	PPB		06 Nov 2023 09:39:49	U	SMPL
O5220-03 - 1	Z-6(5-10)	155652	23.8900	-7471B	PPB		06 Nov 2023 09:42:05	U	SMPL
O5220-04 - 1	Z-7(0-5)	23491	3.5958	-7471B	PPB		06 Nov 2023 09:44:22	U	SMPL
O5220-04DUP - 1	Z-7(0-5)DUP	26104	3.9970	-7471B	PPB		06 Nov 2023 09:46:41	U	SMPL
O5220-05 - 1	O5220-04MS	50384	7.7254	-7471B	PPB		06 Nov 2023 09:49:00	U	SMPL
O5220-06 - 1	O5220-04MSD	48903	7.4980	-7471B	PPB		06 Nov 2023 09:56:58	U	SMPL
O5220-07 - 1	Z-2(125FT)(0-5)	1439	0.2095	-7471B	PPB		06 Nov 2023 09:59:14	U	SMPL
O5228-01 - 1	CHRT-26882	2531	0.3772	-7471B	PPB		06 Nov 2023 10:01:34	U	SMPL
O5229-01 - 1	HR-1-110223	2478	0.3691	-7471B	PPB		06 Nov 2023 10:03:49	U	SMPL
CCV55 - 1	CCV55	31251	4.7874	-7471B	PPB		06 Nov 2023 10:06:04	U	SMPL
CCB55 - 1	CCB55	-127	-0.0310	-7471B	PPB		06 Nov 2023 10:08:20	U	SMPL
O5234-01 - 1	SP-A	797	0.1109	-7471B	PPB		06 Nov 2023 10:10:37	U	SMPL
O5234-03 - 1	SP-B	447	0.0572	-7471B	PPB		06 Nov 2023 10:12:53	U	SMPL
O5234-06 - 1	ATH-OILY-DEBRIS	3433	0.5157	-7471B	PPB		06 Nov 2023 10:15:08	U	SMPL
O5220-04LX5 - 1		5381	0.8148	-7471B	PPB		06 Nov 2023 10:17:24	U	SMPL
O5220-04A - 1		48889	7.4958	-7471B	PPB		06 Nov 2023 10:19:40	U	SMPL
O5220-03DLX5 - 1		33016	5.0584	-7471B	PPB		06 Nov 2023 10:24:55	U	SMPL
CCV56 - 1	CCV56	31433	4.8153	-7471B	PPB		06 Nov 2023 10:29:50	U	SMPL
CCB56 - 1	CCB56	-186	-0.0400	-7471B	PPB		06 Nov 2023 10:32:06	U	SMPL