


LB128193

7470A

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 1.6090e-004

C= -1.6155e-001

Rho= 0.9998564

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Std
0.0	0.000	-0.017	-0.017	896	0.000	896	0				1
0.2	0.200	0.196	-0.004	2224	0.0 %	2224					2
2.5	2.500	2.523	0.023	16687	0.0 %	16687					1
5.0	5.000	4.951	-0.049	31776	0.0 %	31776					1
7.5	7.500	7.621	0.121	48370	0.0 %	48370					2
10.0	10.000	9.925	-0.075	62692	0.0 %	62692					1

LB128193 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 06 Nov 2023 11:07:15

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	896	-	0.0000	7470A	PPB	06 Nov 2023 11:16:24	S	Std
0.2 - 1	51.2	2224	-	0.2000	7470A	PPB	06 Nov 2023 11:18:40	S	Std
2.5 - 1	52.5	16687	-	2.5000	7470A	PPB	06 Nov 2023 11:20:57	S	Std
5.0 - 1	55	31776	-	5.0000	7470A	PPB	06 Nov 2023 11:23:14	S	Std
7.5 - 1	57.5	48370	-	7.5000	7470A	PPB	06 Nov 2023 11:28:18	S	Std
10.0 - 1	57.8	62692	-	10.0000	7470A	PPB	06 Nov 2023 11:33:10	S	Std
ICV57 - 1	ICV57	25592	3.9562	-	7470A	PPB	06 Nov 2023 11:40:58	U	SMPL
ICB57 - 1	ICB57	533	-0.0758	-	7470A	PPB	06 Nov 2023 11:43:13	U	SMPL
CCV57 - 1	CCV57	32736	5.1056	-	7470A	PPB	06 Nov 2023 11:50:32	U	SMPL
CCB57 - 1	CCB57	458	-0.0879	-	7470A	PPB	06 Nov 2023 11:52:47	U	SMPL
CRA - 1	CRA	2105	0.1771	-	7470A	PPB	06 Nov 2023 11:55:04	U	SMPL
HighStd - 1	HighStd	61606	9.7508	-	7470A	PPB	06 Nov 2023 11:57:19	U	SMPL
ChkStd - 1	ChkStd	44783	7.0440	-	7470A	PPB	06 Nov 2023 11:59:35	U	SMPL
PB156928BL - 1	PBW	1012	0.0013	-	7470A	PPB	06 Nov 2023 12:04:02	U	SMPL
PB156928BS - 1	LCSW	23202	3.5716	-	7470A	PPB	06 Nov 2023 12:06:18	U	SMPL
O5095-01 - 1	F07367	286	-0.1155	-	7470A	PPB	06 Nov 2023 12:08:34	U	SMPL
O5208-02 - 1	COMP	1003	-0.0002	-	7470A	PPB	06 Nov 2023 12:10:52	U	SMPL
O5217-01 - 1	1193	345	-0.1060	-	7470A	PPB	06 Nov 2023 12:13:09	U	SMPL
O5227-01 - 1	M21005	-437	-0.2319	-	7470A	PPB	06 Nov 2023 12:15:27	U	SMPL
CCV58 - 1	CCV58	34428	5.3779	-	7470A	PPB	06 Nov 2023 12:20:14	U	SMPL
CCB58 - 1	CCB58	501	-0.0809	-	7470A	PPB	06 Nov 2023 12:22:29	U	SMPL
O5227-03 - 1	A6934	2327	0.2129	-	7470A	PPB	06 Nov 2023 12:30:53	U	SMPL
O5227-03DUP - 1	A6934DUP	2204	0.1931	-	7470A	PPB	06 Nov 2023 12:33:08	U	SMPL
O5227-03MS - 1	A6934MS	17692	2.6851	-	7470A	PPB	06 Nov 2023 12:35:25	U	SMPL
O5227-03MSD - 1	A6934MSD	16872	2.5531	-	7470A	PPB	06 Nov 2023 12:37:42	U	SMPL
O5227-03LX5 - 1		1024	0.0032	-	7470A	PPB	06 Nov 2023 12:42:35	U	SMPL
O5227-03A - 1		23867	3.6786	-	7470A	PPB	06 Nov 2023 12:44:50	U	SMPL
CCV59 - 1	CCV59	31959	4.9806	-	7470A	PPB	06 Nov 2023 12:47:07	U	SMPL
CCB59 - 1	CCB59	945	-0.0095	-	7470A	PPB	06 Nov 2023 12:51:53	U	SMPL