



PERCENT SOLID

Supervisor: Iwona

Analyst: JIGNESH

Date: 11/7/2023

OVENTEMP IN Celsius(°C): 107

Time IN: 17:25

In Date: 11/06/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:15

Out Date: 11/07/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB128195

| Lab ID   | Client SampleID | Dish # | Dish Wt(g) (A) | Sample Wt(g) | Dish + Sample Wt(g) (B) | Dish+Dry Sample Wt(g) (C) | % Solid | Comments |
|----------|-----------------|--------|----------------|--------------|-------------------------|---------------------------|---------|----------|
| O5102-02 | 67S-SB02-0815   | 89     | 1.15           | 8.58         | 9.73                    | 7.6                       | 75.2    |          |
| O5102-03 | 67S-SB02-2628   | 90     | 1.14           | 8.64         | 9.78                    | 6.04                      | 56.7    |          |
| O5102-04 | 67S-SB04-1419   | 91     | 1.16           | 8.74         | 9.9                     | 7.15                      | 68.5    |          |
| O5102-05 | 67S-SB07-1418   | 92     | 1.15           | 8.81         | 9.96                    | 7.55                      | 72.6    |          |
| O5102-06 | 67S-SB08-1619   | 93     | 1.15           | 8.81         | 9.96                    | 9.00                      | 89.1    |          |
| O5102-07 | 67S-SB09-1314.5 | 94     | 1.19           | 8.79         | 9.98                    | 5.8                       | 52.4    |          |
| O5102-08 | 67S-SB11-1218   | 95     | 1.12           | 8.46         | 9.58                    | 8.11                      | 82.6    |          |
| O5102-09 | 67S-SB11-1218-D | 96     | 1.19           | 8.47         | 9.66                    | 8.02                      | 80.6    |          |
| O5102-10 | 67S-SB14-1216   | 97     | 1.11           | 8.76         | 9.87                    | 8.07                      | 79.5    |          |
| O5102-11 | 67S-SB16-1617   | 98     | 1.17           | 8.55         | 9.72                    | 7.46                      | 73.6    |          |
| O5102-13 | 67-IDW-01       | 99     | 1.13           | 8.52         | 9.65                    | 8.02                      | 80.9    |          |
| O5244-01 | 1A              | 1      | 1.14           | 8.63         | 9.77                    | 7.54                      | 74.2    |          |
| O5244-04 | 2A              | 2      | 1.18           | 8.48         | 9.66                    | 8.15                      | 82.2    |          |
| O5244-07 | 3A              | 3      | 1.16           | 8.50         | 9.66                    | 8.22                      | 83.1    |          |
| O5244-10 | 4A              | 4      | 1.19           | 8.78         | 9.97                    | 8.58                      | 84.2    |          |
| O5244-13 | 5A              | 5      | 1.11           | 8.47         | 9.58                    | 8.47                      | 86.9    |          |
| O5244-16 | 6A              | 6      | 1.19           | 8.71         | 9.9                     | 6.99                      | 66.6    |          |
| O5244-19 | 7A              | 7      | 1.19           | 8.79         | 9.98                    | 7.01                      | 66.2    |          |
| O5245-02 | 8A              | 8      | 1.16           | 8.48         | 9.64                    | 6.25                      | 60.0    |          |
| O5245-05 | 9A              | 9      | 1.19           | 8.55         | 9.74                    | 7.21                      | 70.4    |          |
| O5245-08 | 10A             | 10     | 1.13           | 8.67         | 9.8                     | 7.04                      | 68.2    |          |
| O5245-11 | 11A             | 11     | 1.16           | 8.68         | 9.84                    | 7.72                      | 75.6    |          |
| O5245-14 | 12A             | 12     | 1.15           | 8.83         | 9.98                    | 8.19                      | 79.7    |          |
| O5245-17 | 13A             | 13     | 1.14           | 8.68         | 9.82                    | 8.6                       | 85.9    |          |
| O5245-20 | 14A             | 14     | 1.19           | 8.79         | 9.98                    | 8.86                      | 87.3    |          |
| O5246-03 | 15A             | 15     | 1.18           | 8.64         | 9.82                    | 8.47                      | 84.4    |          |
| O5246-06 | 16A             | 16     | 1.13           | 8.84         | 9.97                    | 9.00                      | 89.0    |          |
| O5246-09 | 17A             | 17     | 1.16           | 8.80         | 9.96                    | 8.81                      | 86.9    |          |



PERCENT SOLID

Supervisor: Iwona

Analyst: JIGNESH

Date: 11/7/2023

OVENTEMP IN Celsius(°C): 107

Time IN: 17:25

In Date: 11/06/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:15

Out Date: 11/07/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB128195

| Lab ID   | Client SampleID | Dish # | Dish Wt(g) (A) | Sample Wt(g) | Dish + Sample Wt(g) (B) | Dish+Dry Sample Wt(g) (C) | % Solid | Comments |
|----------|-----------------|--------|----------------|--------------|-------------------------|---------------------------|---------|----------|
| O5246-12 | 18A             | 18     | 1.13           | 8.75         | 9.88                    | 8.64                      | 85.8    |          |
| O5246-15 | 19A             | 19     | 1.18           | 8.50         | 9.68                    | 8.56                      | 86.8    |          |
| O5246-18 | 20B             | 20     | 1.18           | 8.78         | 9.96                    | 7.48                      | 71.8    |          |
| O5247-01 | 21B             | 21     | 1.18           | 8.46         | 9.64                    | 8.2                       | 83.0    |          |
| O5247-04 | 22B             | 22     | 1.18           | 8.79         | 9.97                    | 8.7                       | 85.6    |          |
| O5247-07 | 23B             | 23     | 1.15           | 8.80         | 9.95                    | 8.51                      | 83.6    |          |
| O5247-10 | 24B             | 24     | 1.18           | 8.41         | 9.59                    | 8.37                      | 85.5    |          |
| O5247-13 | 25B             | 25     | 1.14           | 8.65         | 9.79                    | 8.84                      | 89.0    |          |
| O5247-16 | 26A             | 26     | 1.17           | 8.57         | 9.74                    | 8.42                      | 84.6    |          |
| O5247-19 | 27A             | 27     | 1.19           | 8.59         | 9.78                    | 7.36                      | 71.8    |          |
| O5248-02 | 28A             | 28     | 1.19           | 8.80         | 9.99                    | 7.44                      | 71.0    |          |
| O5248-03 | 29A             | 29     | 1.12           | 8.71         | 9.83                    | 8.21                      | 81.4    |          |
| O5248-04 | 30A             | 30     | 1.19           | 8.58         | 9.77                    | 7.99                      | 79.3    |          |
| O5248-05 | 31A             | 31     | 1.15           | 8.53         | 9.68                    | 7.61                      | 75.7    |          |
| O5248-06 | 32A             | 32     | 1.19           | 8.73         | 9.92                    | 7.88                      | 76.6    |          |
| O5248-07 | 33A             | 33     | 1.18           | 8.80         | 9.98                    | 7.29                      | 69.4    |          |
| O5248-08 | 34A             | 34     | 1.19           | 8.60         | 9.79                    | 8.26                      | 82.2    |          |
| O5248-09 | 35A             | 35     | 1.16           | 8.58         | 9.74                    | 7.74                      | 76.7    |          |
| O5248-10 | DUP-1           | 36     | 1.15           | 8.51         | 9.66                    | 7.49                      | 74.5    |          |
| O5248-13 | DUP-4           | 37     | 1.16           | 8.81         | 9.97                    | 7.83                      | 75.7    |          |
| O5248-14 | DUP-5           | 38     | 1.19           | 8.55         | 9.74                    | 8.32                      | 83.4    |          |
| O5248-16 | DUP-7           | 39     | 1.19           | 8.51         | 9.7                     | 7.1                       | 69.4    |          |
| O5251-01 | T-1             | 40     | 1.12           | 8.76         | 9.88                    | 8.27                      | 81.6    |          |
| O5251-02 | T-2             | 41     | 1.19           | 8.50         | 9.69                    | 8.42                      | 85.1    |          |
| O5251-03 | T-3             | 42     | 1.15           | 8.52         | 9.67                    | 8.14                      | 82.0    |          |
| O5251-04 | T-4             | 43     | 1.19           | 8.43         | 9.62                    | 8.39                      | 85.4    |          |
| O5251-05 | T-5             | 44     | 1.16           | 8.39         | 9.55                    | 8.39                      | 86.2    |          |
| O5251-06 | T-6             | 45     | 1.18           | 8.63         | 9.81                    | 8.22                      | 81.6    |          |



PERCENT SOLID

Supervisor: Iwona  
Analyst: JIGNESH  
Date: 11/7/2023

OVENTEMP IN Celsius(°C): 107  
Time IN: 17:25  
In Date: 11/06/2023  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:15  
Out Date: 11/07/2023  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB128195

| Lab ID   | Client SampleID   | Dish # | Dish Wt(g) (A) | Sample Wt(g) | Dish + Sample Wt(g) (B) | Dish+Dry Sample Wt(g) (C) | % Solid | Comments        |
|----------|-------------------|--------|----------------|--------------|-------------------------|---------------------------|---------|-----------------|
| O5251-07 | T-7               | 46     | 1.17           | 8.50         | 9.67                    | 8.55                      | 86.8    |                 |
| O5251-08 | T-8               | 47     | 1.19           | 8.69         | 9.88                    | 8.73                      | 86.8    |                 |
| O5251-09 | T-9               | 48     | 1.12           | 8.70         | 9.82                    | 8.76                      | 87.8    |                 |
| O5251-10 | T-10              | 49     | 1.19           | 8.73         | 9.92                    | 8.85                      | 87.7    |                 |
| O5251-11 | T-11              | 50     | 1.19           | 8.65         | 9.84                    | 8.68                      | 86.6    |                 |
| O5252-01 | WASTE             | 51     | 1.17           | 8.60         | 9.77                    | 8.96                      | 90.6    |                 |
| O5253-01 | L-1 (65FT) (5-10) | 52     | 1.18           | 8.53         | 9.71                    | 8.35                      | 84.1    |                 |
| O5253-02 | L-6 (0-5)         | 53     | 1.13           | 8.53         | 9.66                    | 8.76                      | 89.4    |                 |
| O5253-03 | L-3 (120FT) (0-5) | 54     | 1.15           | 8.57         | 9.72                    | 9.00                      | 91.6    |                 |
| O5253-04 | L-3 (195FT) (0-5) | 55     | 1.19           | 8.58         | 9.77                    | 9.02                      | 91.3    |                 |
| O5258-01 | 01-A-01-B-01-C    | 56     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CAULKING SAMPLE |
| O5258-02 | 02-A-02-B-02-C    | 57     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CAULKING SAMPLE |
| O5258-03 | 03-A-03-B-03-C    | 58     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CAULKING SAMPLE |
| O5258-04 | 04-A-04-B-04-C    | 59     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CAULKING SAMPLE |
| O5258-05 | 05-A-05-B-05-C    | 60     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CAULKING SAMPLE |
| O5258-06 | 06-A-06-B-06-C    | 61     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CAULKING SAMPLE |
| O5258-07 | 07-A-07-B-07-C    | 62     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CAULKING SAMPLE |
| O5258-08 | 08-A-08-B-08-C    | 63     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CAULKING SAMPLE |
| O5258-09 | 09-A-09-B-09-C    | 64     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CAULKING SAMPLE |
| O5258-10 | 10-A-10-B-10-C    | 65     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CAULKING SAMPLE |
| O5258-11 | 11-A-11-B-11-C    | 66     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CAULKING SAMPLE |
| O5258-12 | 12-A-12-B-12-C    | 67     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CAULKING SAMPLE |
| O5258-13 | 13-A-14-B-14-C    | 68     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CAULKING SAMPLE |
| O5258-14 | 14-A-14-B-14-C    | 69     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CAULKING SAMPLE |
| O5258-15 | 15-A-15-B-15-C    | 70     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CAULKING SAMPLE |
| O5258-16 | 16-A-16-B-16-C    | 71     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CAULKING SAMPLE |
| O5258-17 | 17-A-17-B-17-C    | 72     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CAULKING SAMPLE |
| O5258-18 | 18-A-18-B-18-C    | 73     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CAULKING SAMPLE |



PERCENT SOLID

Supervisor: Iwona  
Analyst: JIGNESH  
Date: 11/7/2023

OVENTEMP IN Celsius(°C): 107  
Time IN: 17:25  
In Date: 11/06/2023  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:15  
Out Date: 11/07/2023  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB128195

| Lab ID   | Client SampleID | Dish # | Dish Wt(g) (A) | Sample Wt(g) | Dish + Sample Wt(g) (B) | Dish+Dry Sample Wt(g) (C) | % Solid | Comments        |
|----------|-----------------|--------|----------------|--------------|-------------------------|---------------------------|---------|-----------------|
| O5258-19 | 19-A-19-B-19-C  | 74     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CAULKING SAMPLE |
| O5258-20 | 20-A-20-B-20-C  | 75     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CAULKING SAMPLE |
| O5258-21 | 21-A-21-B-21-C  | 76     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CAULKING SAMPLE |
| O5258-22 | 22-A-22-B-22-C  | 77     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CAULKING SAMPLE |
| O5258-23 | 23-A-23-B-23-C  | 78     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CAULKING SAMPLE |
| O5258-24 | 24-A-24-B-24-C  | 79     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CAULKING SAMPLE |
| O5258-25 | 25-A-25-B-25-C  | 80     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CAULKING SAMPLE |
| O5258-26 | 26-A-26-B-26-C  | 81     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CAULKING SAMPLE |
| O5258-27 | 27-A-27-B-27-C  | 82     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CAULKING SAMPLE |
| O5266-01 | 1203            | 83     | 1.15           | 8.42         | 9.57                    | 9.24                      | 96.1    |                 |
| O5267-01 | ETGI-320        | 86     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CONCRETE SAMPLE |
| O5270-01 | OR-3-110623     | 84     | 1.18           | 8.80         | 9.98                    | 9.34                      | 92.7    |                 |
| O5270-02 | OR-3-110623-E2  | 85     | 1.13           | 8.80         | 9.93                    | 9.01                      | 89.5    |                 |
| O5272-01 | 001             | 87     | 1.15           | 8.39         | 9.54                    | 7.94                      | 80.9    |                 |
| O5272-02 | 002             | 88     | 1.15           | 8.39         | 9.54                    | 7.94                      | 80.9    |                 |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$