

LB128198


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

54307



A= 0.0000e+000

B= 1.8422e-004 slope

C= -6.5302e-002 tolerance

Rho= 0.9999022

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	σ _o h
0.00	0.000	-0.044	-0.044	113	0.000	113					5
0.05	0.050					419	0				-1
0.20	0.200	0.199	-0.001	1435	0.0 %	1435					2
2.50	2.500	2.556	0.056	14227	0.0 %	14227					-1
5.00	5.000	4.971	-0.029	27336	0.0 %	27336					1
7.50	7.500	7.580	0.080	41499	0.0 %	41499					-1
10.0	10.000	9.939	-0.061	54307	0.0 %	54307					

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	113	-		0 SFAM01.1	PPB	11/6/2023 15:35	Std
	0.2 S01	1435	-		0.2 SFAM01.1	PPB	11/6/2023 15:40	Std
	2.5 S02	14227	-		2.5 SFAM01.1	PPB	11/6/2023 15:42	Std
	5 S03	27336	-		5 SFAM01.1	PPB	11/6/2023 15:45	Std
	7.5 S04	41499	-		7.5 SFAM01.1	PPB	11/6/2023 15:47	Std
	10 S05	54307	-		10 SFAM01.1	PPB	11/6/2023 15:57	Std
ICV058	ICV058	21390	3.8752 -		SFAM01.1	PPB	11/6/2023 16:00	SMPL
ICB058	ICB058	-99	-0.0835 -		SFAM01.1	PPB	11/6/2023 16:02	SMPL
CCV060	CCV060	27032	4.9146 -		SFAM01.1	PPB	11/6/2023 16:04	SMPL
CCB060	CCB060	-107	-0.085 -		SFAM01.1	PPB	11/6/2023 16:07	SMPL
PB156916BL	PBW916	-138	-0.0907 -		SFAM01.1	PPB	11/6/2023 16:09	SMPL
O4982-01	MA48Z5	-60	-0.0764 -		SFAM01.1	PPB	11/6/2023 16:11	SMPL
O4982-04	MA4A77	21538	3.9025 -		SFAM01.1	PPB	11/6/2023 16:13	SMPL
O4982-05	MA4977	-227	-0.1071 -		SFAM01.1	PPB	11/6/2023 16:16	SMPL
O4982-06	MA49E7	-61	-0.0765 -		SFAM01.1	PPB	11/6/2023 16:18	SMPL
O4982-08	MA49E5	-27	-0.0703 -		SFAM01.1	PPB	11/6/2023 16:20	SMPL
O4982-09	MA49D7	-42	-0.073 -		SFAM01.1	PPB	11/6/2023 16:22	SMPL
O4982-10	MA4955	-19	-0.0688 -		SFAM01.1	PPB	11/6/2023 16:25	SMPL
O4982-11	MA49M7	-27	-0.0703 -		SFAM01.1	PPB	11/6/2023 16:27	SMPL
O4982-12	MA49D3	-35	-0.0718 -		SFAM01.1	PPB	11/6/2023 16:29	SMPL
O4982-13	MA49D4	-39	-0.0725 -		SFAM01.1	PPB	11/6/2023 16:31	SMPL
O4982-14	MA4953	-35	-0.0718 -		SFAM01.1	PPB	11/6/2023 16:34	SMPL
O4982-15	MA49D1	-30	-0.0708 -		SFAM01.1	PPB	11/6/2023 16:36	SMPL
O4982-16	MA4935	-32	-0.0712 -		SFAM01.1	PPB	11/6/2023 16:38	SMPL
O4982-17	MA4949	22	-0.0612 -		SFAM01.1	PPB	11/6/2023 16:41	SMPL
O4982-18	MA49L3	-14	-0.0679 -		SFAM01.1	PPB	11/6/2023 16:43	SMPL
O4982-19	MA49C8	-14	-0.0679 -		SFAM01.1	PPB	11/6/2023 16:45	SMPL
O4982-20	MA4945	-15	-0.0681 -		SFAM01.1	PPB	11/6/2023 16:47	SMPL
O4982-21	MA4947	-9	-0.067 -		SFAM01.1	PPB	11/6/2023 16:50	SMPL
O4982-22	MA49K6	-22	-0.0694 -		SFAM01.1	PPB	11/6/2023 16:52	SMPL
CCV061	CCV061	25674	4.6645 -		SFAM01.1	PPB	11/6/2023 16:54	SMPL
CCB061	CCB061	-179	-0.0983 -		SFAM01.1	PPB	11/6/2023 16:56	SMPL
O4982-23	MA4937	-67	-0.0776 -		SFAM01.1	PPB	11/6/2023 16:59	SMPL
O4982-24	MA4937D	-32	-0.0712 -		SFAM01.1	PPB	11/6/2023 17:01	SMPL
O4982-25	MA4937S	5551	0.9573 -		SFAM01.1	PPB	11/6/2023 17:03	SMPL
PB156917BL	PBW917	-133	-0.0898 -		SFAM01.1	PPB	11/6/2023 17:05	SMPL
O4983-01	MA48Z6	-60	-0.0764 -		SFAM01.1	PPB	11/6/2023 17:08	SMPL
O4983-03	MA4974	-35	-0.0718 -		SFAM01.1	PPB	11/6/2023 17:10	SMPL
O4983-04	MA4976	2	-0.0649 -		SFAM01.1	PPB	11/6/2023 17:12	SMPL
O4983-05	MA4958	-5	-0.0662 -		SFAM01.1	PPB	11/6/2023 17:15	SMPL
O4983-06	MA49E4	-12	-0.0675 -		SFAM01.1	PPB	11/6/2023 17:17	SMPL
O4983-07	MA49D5	-11	-0.0673 -		SFAM01.1	PPB	11/6/2023 17:19	SMPL
O4983-08	MA49E6	-20	-0.069 -		SFAM01.1	PPB	11/6/2023 17:21	SMPL
O4983-09	MA49D6	-12	-0.0675 -		SFAM01.1	PPB	11/6/2023 17:24	SMPL
O4983-10	MA49F0	6	-0.0642 -		SFAM01.1	PPB	11/6/2023 17:26	SMPL
O4983-11	MA49E9	-10	-0.0671 -		SFAM01.1	PPB	11/6/2023 17:28	SMPL

INSTRUMENT ID : CV1

O4983-12	MA49L8	-13	-0.0677 -	SFAM01.1	PPB	11/6/2023 17:30	SMPL
O4983-13	MA49E2	-14	-0.0679 -	SFAM01.1	PPB	11/6/2023 17:33	SMPL
O4983-14	MA4936	-16	-0.0683 -	SFAM01.1	PPB	11/6/2023 17:35	SMPL
O4983-15	MA4950	0	-0.0653 -	SFAM01.1	PPB	11/6/2023 17:37	SMPL
O4983-16	MA49L0	-5	-0.0662 -	SFAM01.1	PPB	11/6/2023 17:39	SMPL
O4983-17	MA49C9	-2	-0.0657 -	SFAM01.1	PPB	11/6/2023 17:42	SMPL
CCV062	CCV062	26343	4.7877 -	SFAM01.1	PPB	11/6/2023 17:44	SMPL
CCB062	CCB062	11	-0.0633 -	SFAM01.1	PPB	11/6/2023 17:46	SMPL
O4983-18	MA4946	-67	-0.0776 -	SFAM01.1	PPB	11/6/2023 17:49	SMPL
O4983-19	MA4948	-44	-0.0734 -	SFAM01.1	PPB	11/6/2023 17:51	SMPL
O4983-20	MA49K7	-33	-0.0714 -	SFAM01.1	PPB	11/6/2023 17:53	SMPL
O4983-21	MA4938	-16	-0.0683 -	SFAM01.1	PPB	11/6/2023 17:55	SMPL
O4983-22	MA4938D	-21	-0.0692 -	SFAM01.1	PPB	11/6/2023 17:58	SMPL
O4983-23	MA4938S	6332	1.1012 -	SFAM01.1	PPB	11/6/2023 18:00	SMPL
O4983-24	MA4A67	44112	8.0612 -	SFAM01.1	PPB	11/6/2023 18:02	SMPL
PB156943BL	PBW943	-478	-0.1534 -	SFAM01.1	PPB	11/6/2023 18:04	SMPL
O4997-01	MA49L5	475	0.0222 -	SFAM01.1	PPB	11/6/2023 18:07	SMPL
O4997-05	MA49C6	407	0.0097 -	SFAM01.1	PPB	11/6/2023 18:11	SMPL
O4997-06	MA4941	506	0.0279 -	SFAM01.1	PPB	11/6/2023 18:14	SMPL
O4997-07	MA4943	522	0.0309 -	SFAM01.1	PPB	11/6/2023 18:16	SMPL
O4997-08	MA49J8	520	0.0305 -	SFAM01.1	PPB	11/6/2023 18:18	SMPL
O4997-09	MA49C3	542	0.0345 -	SFAM01.1	PPB	11/6/2023 18:20	SMPL
O4997-10	MA49C5	573	0.0403 -	SFAM01.1	PPB	11/6/2023 18:23	SMPL
O4997-11	MA49J7	485	0.024 -	SFAM01.1	PPB	11/6/2023 18:25	SMPL
O4997-12	MA4931	554	0.0368 -	SFAM01.1	PPB	11/6/2023 18:27	SMPL
O4997-13	MA49J4	572	0.0401 -	SFAM01.1	PPB	11/6/2023 18:29	SMPL
O4997-14	MA4927	564	0.0386 -	SFAM01.1	PPB	11/6/2023 18:32	SMPL
CCV063	CCV063	26702	4.8538 -	SFAM01.1	PPB	11/6/2023 18:34	SMPL
CCB063	CCB063	-60	-0.0764 -	SFAM01.1	PPB	11/6/2023 18:36	SMPL
O4997-15	MA49A1	518	0.0301 -	SFAM01.1	PPB	11/6/2023 18:39	SMPL
O4997-16	MA49A1D	532	0.0327 -	SFAM01.1	PPB	11/6/2023 18:41	SMPL
O4997-17	MA49A1S	3004	0.4881 -	SFAM01.1	PPB	11/6/2023 18:43	SMPL
O4997-18	MA49K1	565	0.0388 -	SFAM01.1	PPB	11/6/2023 18:45	SMPL
O4997-19	MA4995	535	0.0333 -	SFAM01.1	PPB	11/6/2023 18:48	SMPL
O4997-20	MA49A5	543	0.0347 -	SFAM01.1	PPB	11/6/2023 18:50	SMPL
O4997-21	MA4923	460	0.0194 -	SFAM01.1	PPB	11/6/2023 18:52	SMPL
O4997-22	MA49J0	434	0.0147 -	SFAM01.1	PPB	11/6/2023 18:54	SMPL
O4997-23	MA49C1	476	0.0224 -	SFAM01.1	PPB	11/6/2023 18:57	SMPL
O4997-24	MA4998	573	0.0403 -	SFAM01.1	PPB	11/6/2023 18:59	SMPL
PB156944BL	PBW944	302	-0.0097 -	SFAM01.1	PPB	11/6/2023 19:01	SMPL
O4998-01	MA49M6	477	0.0226 -	SFAM01.1	PPB	11/6/2023 19:04	SMPL
O4998-02	MA4A41	32647	5.9491 -	SFAM01.1	PPB	11/6/2023 19:06	SMPL
O4998-04	MA49D9	-247	-0.1108 -	SFAM01.1	PPB	11/6/2023 19:08	SMPL
O4998-05	MA4942	519	0.0303 -	SFAM01.1	PPB	11/6/2023 19:10	SMPL
O4998-06	MA4944	375	0.0038 -	SFAM01.1	PPB	11/6/2023 19:13	SMPL
O4998-07	MA49K4	346	-0.0016 -	SFAM01.1	PPB	11/6/2023 19:15	SMPL
O4998-08	MA49C4	453	0.0182 -	SFAM01.1	PPB	11/6/2023 19:17	SMPL

INSTRUMENT ID : CV1

O4998-09	MA49D8	458	0.0191 -	SFAM01.1	PPB	11/6/2023 19:19 SMPL
O4998-10	MA49K3	594	0.0441 -	SFAM01.1	PPB	11/6/2023 19:22 SMPL
CCV064	CCV064	26227	4.7663 -	SFAM01.1	PPB	11/6/2023 19:24 SMPL
CCB064	CCB064	-51	-0.0747 -	SFAM01.1	PPB	11/6/2023 19:26 SMPL
O4998-11	MA4932	548	0.0357 -	SFAM01.1	PPB	11/6/2023 19:29 SMPL
O4998-12	MA49K2	557	0.0373 -	SFAM01.1	PPB	11/6/2023 19:31 SMPL
O4998-13	MA4928	447	0.017 -	SFAM01.1	PPB	11/6/2023 19:33 SMPL
O4998-14	MA49J3	557	0.0373 -	SFAM01.1	PPB	11/6/2023 19:35 SMPL
O4998-15	MA49B6	583	0.0421 -	SFAM01.1	PPB	11/6/2023 19:38 SMPL
O4998-16	MA49C2	-6	-0.0664 -	SFAM01.1	PPB	11/6/2023 19:40 SMPL
O4998-17	MA4924	11	-0.0633 -	SFAM01.1	PPB	11/6/2023 19:42 SMPL
O4998-18	MA49J1	16	-0.0624 -	SFAM01.1	PPB	11/6/2023 19:45 SMPL
O4998-19	MA49A4	11	-0.0633 -	SFAM01.1	PPB	11/6/2023 19:47 SMPL
O4998-20	MA49B7	10	-0.0635 -	SFAM01.1	PPB	11/6/2023 19:49 SMPL
O4998-21	MA49C0	9	-0.0636 -	SFAM01.1	PPB	11/6/2023 19:51 SMPL
O4998-22	MA49C0D	3	-0.0647 -	SFAM01.1	PPB	11/6/2023 19:54 SMPL
O4998-23	MA49C0S	1598	0.2291 -	SFAM01.1	PPB	11/6/2023 19:56 SMPL
CCV065	CCV065	25925	4.7107 -	SFAM01.1	PPB	11/6/2023 19:58 SMPL
CCB065	CCB065	-53	-0.0751 -	SFAM01.1	PPB	11/6/2023 20:00 SMPL
O4997-04	MA4A78	11062	1.9726 -	SFAM01.1	PPB	11/6/2023 20:12 SMPL
CCV066	CCV066	26549	4.8257 -	SFAM01.1	PPB	11/6/2023 20:21 SMPL
CCB066	CCB066	-167	-0.0961 -	SFAM01.1	PPB	11/6/2023 20:23 SMPL