

7470A


LB128207

INSTRUMENT ID: CV1

Linear

 μ Abs.:

64890



A= 0.0000e+000

B= 1.5616e-004

C= -1.8859e-001

Rho= 0.9998476

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	4.12
0.0	0.000	-0.023	-0.023	1062	0.000	1062	0				19
0.2	0.200	0.238	0.038	2729	0.0 %	2729					0
2.5	2.500	2.493	-0.007	17172	0.0 %	17172					-1
5.0	5.000	4.927	-0.073	32760	0.0 %	32760					2
7.5	7.500	7.621	0.121	50008	0.0 %	50008					-1
10.0	10.000	9.945	-0.055	64890	0.0 %	64890					

LB128207 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 07 Nov 2023 11:12:57

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	1062	-	0.0000	7470A	PPB	07 Nov 2023 11:19:36	S	Std
0.2 - 1	50.2	2729	-	0.2000	7470A	PPB	07 Nov 2023 11:21:53	S	Std
2.5 - 1	52.5	17172	-	2.5000	7470A	PPB	07 Nov 2023 11:24:10	S	Std
5.0 - 1	55	32760	-	5.0000	7470A	PPB	07 Nov 2023 11:26:27	S	Std
7.5 - 1	57.5	50008	-	7.5000	7470A	PPB	07 Nov 2023 11:31:17	S	Std
10.0 - 1	59	64890	-	10.0000	7470A	PPB	07 Nov 2023 11:36:20	S	Std
ICV59 - 1	ICV59	26758	3.9899	-	7470A	PPB	07 Nov 2023 11:46:22	U	SMPL
ICB59 - 1	ICB59	913	-0.0460	-	7470A	PPB	07 Nov 2023 11:48:37	U	SMPL
CCV67 - 1	CCV67	30147	4.5191	-	7470A	PPB	07 Nov 2023 11:50:53	U	SMPL
CCB67 - 1	CCB67	899	-0.0482	-	7470A	PPB	07 Nov 2023 11:53:08	U	SMPL
CRA - 1	CRA	2584	0.2149	-	7470A	PPB	07 Nov 2023 11:55:24	U	SMPL
HighStd - 1	HighStd	64153	9.8294	-	7470A	PPB	07 Nov 2023 11:57:40	U	SMPL
ChkStd - 1	ChkStd	45703	6.9483	-	7470A	PPB	07 Nov 2023 11:59:55	U	SMPL
PB156947BL - 1	PBW	1126	-0.0128	-	7470A	PPB	07 Nov 2023 12:05:21	U	SMPL
PB156947BS - 1	LCSW	25121	3.7343	-	7470A	PPB	07 Nov 2023 12:10:15	U	SMPL
O5107-24 - 1	1048	624	-0.0912	-	7470A	PPB	07 Nov 2023 12:12:32	U	SMPL
O5107-25 - 1	1049	653	-0.0866	-	7470A	PPB	07 Nov 2023 12:14:50	U	SMPL
O5107-26 - 1	1073	5773	0.7129	-	7470A	PPB	07 Nov 2023 12:17:07	U	SMPL
O5107-27 - 1	1074	498	-0.1108	-	7470A	PPB	07 Nov 2023 12:19:25	U	SMPL
O5107-28 - 1	1075	1239	0.0049	-	7470A	PPB	07 Nov 2023 12:21:42	U	SMPL
CCV68 - 1	CCV68	31255	4.6921	-	7470A	PPB	07 Nov 2023 12:23:58	U	SMPL
CCB68 - 1	CCB68	902	-0.0477	-	7470A	PPB	07 Nov 2023 12:26:15	U	SMPL
O5107-29 - 1	1076	2241	0.1614	-	7470A	PPB	07 Nov 2023 12:28:31	U	SMPL
O5107-30 - 1	1077	1887	0.1061	-	7470A	PPB	07 Nov 2023 12:30:46	U	SMPL
O5107-31 - 1	1078	2249	0.1626	-	7470A	PPB	07 Nov 2023 12:33:03	U	SMPL
O5107-32 - 1	1079	1646	0.0684	-	7470A	PPB	07 Nov 2023 12:35:19	U	SMPL
O5107-33 - 1	1080	1094	-0.0178	-	7470A	PPB	07 Nov 2023 12:37:34	U	SMPL
O5107-34 - 1	1081	561	-0.1010	-	7470A	PPB	07 Nov 2023 12:39:51	U	SMPL
O5107-35 - 1	1047	491	-0.1119	-	7470A	PPB	07 Nov 2023 12:42:07	U	SMPL
O5218-01 - 1	149TH-ST-POLY-DRUM-17	415	-0.1238	-	7470A	PPB	07 Nov 2023 12:44:24	U	SMPL
O5218-02 - 1	149TH-ST-PPE-DRUM-18	223	-0.1538	-	7470A	PPB	07 Nov 2023 12:46:41	U	SMPL
O5230-01 - 1	60113	415	-0.1238	-	7470A	PPB	07 Nov 2023 12:48:57	U	SMPL
CCV69 - 1	CCV69	32938	4.9549	-	7470A	PPB	07 Nov 2023 12:51:12	U	SMPL
CCB69 - 1	CCB69	839	-0.0576	-	7470A	PPB	07 Nov 2023 12:56:05	U	SMPL
O5230-02 - 1	60108	66	-0.1783	-	7470A	PPB	07 Nov 2023 12:58:20	U	SMPL
O5237-03 - 1	11TH-ST-1	53	-0.1803	-	7470A	PPB	07 Nov 2023 13:00:35	U	SMPL
O5237-03DUP - 1	11TH-ST-1DUP	53	-0.1803	-	7470A	PPB	07 Nov 2023 13:02:51	U	SMPL
O5237-03MS - 1	11TH-ST-1MS	27116	4.0458	-	7470A	PPB	07 Nov 2023 13:05:07	U	SMPL
O5237-03MSD - 1	11TH-ST-1MSD	25695	3.8239	-	7470A	PPB	07 Nov 2023 13:07:23	U	SMPL
O5237-03LX5 - 1		24	-0.1848	-	7470A	PPB	07 Nov 2023 13:11:55	U	SMPL
O5237-03A - 1		27025	4.0316	-	7470A	PPB	07 Nov 2023 13:14:12	U	SMPL
PB156884TB - 1	PB156884TB	1095	-0.0176	-	7470A	PPB	07 Nov 2023 13:17:20	U	SMPL
CCV70 - 1	CCV70	34409	5.1847	-	7470A	PPB	07 Nov 2023 13:19:56	U	SMPL
CCB70 - 1	CCB70	1049	-0.0248	-	7470A	PPB	07 Nov 2023 13:22:11	U	SMPL