



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Rubina

Parameter: Turbidity

SUPERVISOR REVIEW BY: Iwona

Run Number: LB128213

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP105186
Turbidity Calibration std, 0NTU	WP105185
Turbidity Calibration std, 20NTU	WP105180
Turbidity Calibration std, 40NTU	WP105179
Turbidity Calibration std, 1NTU	WP105183
Turbidity Calibration std, 5NTU	WP105182
Turbidity LOQ std, 1.0NTU	WP105187
Turbidity Calibration - CCV std, 10 NTU	WP105181
10 NTU Standard 500 ml	W2971

Intercept: 0.2728

Slope: 1.0819

Regression: 0.9997

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.154	-0.11		11/07/2023	13:10
2	CAL2	1	1	1.280	0.93	-6.9	11/07/2023	13:14
3	CAL3	5	1	5.530	4.86	-2.8	11/07/2023	13:17
4	CAL4	10	1	11.800	10.66	6.5	11/07/2023	13:20
5	CAL5	20	1	21.500	19.62	-1.9	11/07/2023	13:23
6	CAL6	40	1	43.600	40.05	0.1	11/07/2023	13:27



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Rubina

Parameter: Turbidity

SUPERVISOR REVIEW BY: Iwona

Run Number: LB128213

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	11.200	10.100	11/07/2023	13:30
2	ICB		1	0.160	-0.104	11/07/2023	13:33
3	CCV1	10	1	10.800	9.730	11/07/2023	13:36
4	CCB1		1	0.164	-0.101	11/07/2023	13:40
5	LB128213BL		1	0.158	-0.106	11/07/2023	13:44
6	O4699-07		1	0.848	0.532	11/07/2023	13:47
7	O4699-08		1	1.350	0.996	11/07/2023	13:50
8	CCV2	10	1	10.800	9.730	11/07/2023	13:53
9	CCB2		1	0.164	-0.101	11/07/2023	13:56